A-A-54246
27 August 1990

COMMERCIAL ITEM DESCRIPTION

DISPENSER, FILTERING SOLVENT

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a solvent filtering dispenser.

Salient Characteristics:

Shall be a solvent filtering dispenser for dispensing small volumes of solvent, for intermittent application, by a squeeze bottle action.

Shall include a nominal 1.2 micrometer pore size membrane filter in the delivery line. The membrane filter shall be a nominal 25 millimeters in diameter and at least 3.85 square centimeters in filter area.

Shall be constructed for assembling in accordance with the configuration of figure 1 in ASTM Designation D2276.

Shall include a 1 liter capacity borosilicate glass flask, a neoprene bulb and stopper, a stainless steel filter holder, teflon gaskets and tubing, and glass elbow and dispensing nozzle.

Shall be supplied with a glass flask having a nominal overall diameter of 5-1/4 inches.

Shall have an overall height (fully assembled) of 12-1/2 inches.

Workmanship. The solvent filtering dispenser shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One dispenser, as specified, constitutes one unit.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and testing. Inspections and testing shall be conducted on the solvent filtering dispenser in accordance with commercial practice. Sampling for inspections and testing shall be in accordance with commercial practice.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:31Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Other publications. The following documents form a part of this specification to the extent that is specified herein. Unless otherwise indicated, the effect on date of invitation for bids or request for proposals shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM D 2276 Standard Test Method for Particulate Contaminant in Aviation Fuel.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:31Z
Check the source to verify that this is the current version before use.

A-A-54246

(Application for copies should be addressed to American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103)

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.)

NOTE: The following National Stock Number is covered by this document.

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6640-00-688-7882	DISPENSER, FILTERING SOLVENT

MILITARY INTERESTS:

Custodians:

Army - MD
Air Force - 99

PREPARING ACTIVITY:

DoD-MB

Agent:
DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project Number: 6640-1645

Location: ENABLE/DISPENCD/S31

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:31Z
Check the source to verify that this is the current version before use.