A-A-54252
21 September 1990

COMMERCIAL ITEM DESCRIPTION

SURGICAL PACK, DISPOSABLE

(Arthroscopic)

The General Services Administration has authorized the use of this commercial item description.

This Commercial Item Description covers a sterile, disposable, surgical drape pack for creating a sterile field for arthroscopic procedures

Salient Characteristics

The arthroscopic surgical pack shall be a sterile, disposable, surgical drape pack fabricated of nonwoven material, for creating a sterile field specifically for arthroscopic procedures The pack shall contain at least the following components in the listed approximate dimensions

a One tablecover/overwrap, 43 by 89 inches minimum
b One Mayo stand cover with pad, 23 by 55 inches.
c One scrub nurse gown with towel (overwrapped)
d One surgical gown and towel
e One extra protection surgical gown with towel
f One extremity prep pad with tape, 24 by 44 inches.
g One medium stockinette, 9 by 47 inches
h One adherent/impervious split sheet with a 20 inch adhesive split
i Ten feet non-conductive tubing (optional).
j Four tape strips, 4 by 23 inches (approximate)
k One suture bag
l One arthroscopic limb sheet, 80 by 116 inches

The tablecover/overwrap shall be fan-folded for ease in opening on the surgical back table The tablecover shall be impervious to fluids and moisture. The tablecover shall not

AMSC N/A FSC 6530
DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

tear and shall be resistant to pilling and linting

The gowns shall be large or extra large size One of the gowns shall be individually packaged and located in a memory-free overwrap with a hand towel, for use by the scrub nurse/technician One gown shall have extra protection on the front of the gown which shall be resistant to fluid and dry particle penetration The gowns shall also:

- Be soft, comfortable, pliable, quiet and memory-free
- Have high breathability facilitating comfort when wearing for long periods of time.
- Resist tearing, pilling, and linting
- Have stockinette cuffs which conform to wrists
- Be large enough to cover the back of the individual gowned
- Have a tie closure that shall not rip or separate when tying Shall be designed to minimize the possibility of contamination when tying

The Mayo tray and stand cover shall be a plastic material at least 2 50 mils thick and shall be impervious to fluids The cover shall be fan-folded with a 6 inch cuff on the distal end The pad shall be large enough to cover the surface of the Mayo tray

The pack shall contain an extremity prep pad with tape, approximately 24 by 44 inches for protecting the leg brace The pad shall be U-shaped The pad shall be absorbent on one side, and impervious to fluids on the reverse side

The pack shall have a large size fluid impervious stockinette, approximately 8 by 47 inches The stockinette shall be rolled or fan-folded to permit aseptic application

The pack shall contain a 60 by 70 inch (approximate) split sheet with a 20 inch pressure sensitive adhesive strip for securing the drape The sheet shall be impervious to fluids and resist tearing, pilling and linting The sheet shall be fan-folded to permit aseptic draping of the operative area

The pack shall have a length of non-conductive transparent, plastic tubing at least ten feet long The inside diameter of the tubing shall be approximately 1/4 inch (optional).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

The pack shall contain four (4) tape strips to secure the stockinette around the incision site if necessary. The strips shall be approximately 4 by 23 inches.

The pack shall contain an arthroscopy limb sheet approximately 80 by 116 inches The sheet shall contain an expandable, rubber, krylon hole in the middle to fit over and around the patient's thigh The area around the expandable hole, approximately 36 by 30 inches, shall be reinforced with a highly absorbent, fluid repellent fabric that eliminates strike-through while preventing pooling of fluids.

The pack shall contain a suture bag for discarded sutures and suture packets

All drapes, hand towels, gowns, and overwraps shall be fabricated of a soft material with minimal memory for good drapability. The material shall resist tearing, pilling, and linting.

The drape sheets shall have 'DIRECTIONS FOR PROPER PLACEMENT' clearly imprinted on each drape. The drapes shall be cuffed and fan-folded to permit aseptic draping of the operative area.

The pack shall be sterile and individually wrapped in a suitable tear resistant, heat sealed package which will allow aseptic handling and maintain sterility of the contents until opened or damaged.

Each pack shall also be wrapped in a heat sealed exterior dust cover.

All pack components shall be clearly listed on the exterior wrapper

Workmanship. The surgical pack shall be free from defects which detract from its appearance or impair its serviceability

Unit Package (PG) One package containing four individually packaged sterile surgical packs, as specified, constitutes one unit Each pack shall be wrapped in a suitable tear resistant, heat sealed package which shall allow aseptic handling and shall

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

be capable of maintaining sterility of contents unless the package is damaged or opened.

Quality Assurance Provisions

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1 0 (percent defective) and an inspection level of II The unit of product for examination purposes shall be one surgical pack

Sampling for tests Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1 The unit of product for test purposes shall be one surgical pack

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements Where feasible, the same sample shall be used for the determination of two or more test characteristics Tests shall include, but shall not be limited to, the following:

Sterility test The surgical packs shall be tested for sterility in accordance with the requirements of the Food and Drug Administration

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Regulatory requirements

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U.S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended and regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

Ordering data Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order

NOTE The following National Stock Number is covered by this document

6530-01-156-3402

MILITARY INTERESTS

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY

DoD-MB

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES

VA-OSS
PHS
FDA-MPQAS

Project Number 6530-2186

Location SUPACCID 15E

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in putting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:31Z
Check the source to verify that this is the current version before use.