MILITARY SPECIFICATION

BOX, MICROSCOPE SLIDE, Plastic, 5 Slide

This limited coordination Military Specification has been prepared by the Defense Personnel Support Center based upon currently available technical information. It is subject to modification. However, pending its promulgation as a coordinated military specification, it may be used in procurement.

1. SCOPE

1.1 Scope. This specification covers one type of plastic microscope slide box.

2. APPLICABLE DOCUMENTS

2.1 Specifications, standards and drawings. The following specification and standards, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

MMM-A-250	Adhesive, Water Resistant (For Sealing Waterproof Paper)
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure Sensitive Adhesive Paper, Water Resistant (For Carton Sealing)

Military

MIL-P-116	Preservation, Methods of.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

2.2 Other publications. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIAL (ASTM)

ASTM Designation D 2146 Propylene Plastic Molding and Extrusion Materials.

(Applications for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, Pa. 19103)

3. REQUIREMENTS

3.1 Material. Material shall be of the highest quality throughout and as specified herein.

3.1.1 Polypropylene. Polypropylene shall be in accordance with ASTM Designation D 2146, Type I 98108.

3.2 Construction. Box shall be fabricated of polypropylene and shall be constructed in accordance with figure 1. Slide holder ribs shall be so positioned and aligned that the box is capable of holding five 25 by 75 millimeter (mm) microscope slides (see 6.1). Door shall be secure when closed and shall be easily opened by application of finger pressue. The interior and exterior surface of the box shall be smooth with no protrusions except for areas for identification marking, slide holders, interior of door, door mating porition of the interior of the box. Surface of the box which is opposite the hinge shall be recessed to facilitate opening of the door. Door and door hinge shall not protrude beyond the outside surface of the box.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

3.3 Performance.

3.3.1 Hinge and closure durability. Hinge shall not break and door shall be capable of secure closure when tested in accordance with 4.4.1.

3.4 Identification marking. Box shall be permanently and legibly marked with the manufacturers name or registered trademark.

3.5 Workmanship. Workmanship shall be first class throughout. Boxes shall be free of deformation, breaks, cracks, dirt and other defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplier and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory applicances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificate of Quality. Certificates of Quality, supplied by the manufacturer of the polypropylene, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer, the FSN, the Item Identification, the name of the material, the lot number, the lot size, the sample size, the dates of testing, test methods, individual test results, and the specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

4.2 Sampling

4.2.1 For examination. Sampling shall be conducted in accordance with MIL-STD-105 and as indicated in Table I. Unit of product shall be the box. No more than 10 boxes shall be selected from any unit package.

Table I Sampling for Examination

Type defect	Inspection level	AQL (Percent defective)
For visual examination		
Major defects	II	2.5
Minor defects	II	4.0
For dimensional examination	S-2	2.5

4.2.2 For tests. Sampling for tests shall be in accordance with MIL-STD-105 and as indicated in Table II. Unit of product shall be the box.

Table II Sampling for Tests

Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Durability	3.3.1	4.4.1	S-2	1.0

4.3 Examination. The boxes shall be examined to determine compliance with all requirements contained this specification.

4.3.1 Classification of Defects. Examination shall be conducted in accordance with the classification of defects but shall not be limited to the defects indicated in Table III.

Table III Classification of defects

Categories	Defects
Major	
101	Box not free of deformation
102	Box not free of breakage
103	Box not free of cracks
104	Surfaces of box not free of protrusions (Protrusions required or permitted by Section 3 shall not be considered defects)
105	Box not free of embedded contamination

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

Table III Classification of defects (Con't)

Major	
106	Slides not easily inserted or removed (Slides of maximum dimensions shall be used, see 6.4)
107	Slides do not stay in slide holders. (Slides of minimum dimensions shall be used, see 6.4)
108	Door cannot be completely closed with slides inserted in box. (Slides of maximum dimensions shall be used, see 6.4)
109	Slide holders not aligned
110	Door not easily opened by application of finger pressure.
111	Door not secure when closed.
Minor	
201	Identification marking not complete, correct, permanent and legible.
202	Box not free of dirt and other contamination.

4.3.2 Dimensional examination. The boxes shall be examined for defects in dimensions. Any dimension not within the tolerance specified herein shall be classified as a defects.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Durability. The door shall be completely opened and closed 100 times. Hinge shall be examined for breakage. Visual evidence of breakage shall be considered failure to comply with requirements of 3.3.1. If box passes test for breakage of hinge, five 25 by 75 mm microscope slides shall be placed in the same box. Slides with minimum dimensions shall be used (see 6.4). One end of a string shall be secured to the end opposite the door. The other end of this string shall be secured to an area which is so positioned that the box can be dropped and brought to a stop without touching any object while falling or stopping. Length of string between attached points shall be 12 $\pm$ 1/8 inch. Door shall be closed and the box allowed to fall the length of the string with the door in the down position. If box touches any object while falling or stopping, results shall not be considered valid and test shall be repeated. Opening of the door or release of slides from the box shall be considered failure to comply with requirements of 3.3.1.

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:55Z
Check the source to verify that this is the currentversion before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C, as specified (see 6.2).

5.1.1 Unit of issue. One package containing 144 slide boxes, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV Packaging Quantities

package	Intermediate package	Exterior container
1 Unit	Not Required	10 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall incure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.3.2 Unit package. Slide boxes shall be packaged in a neat uniform manner within a box of appropriate size constructed in accordance with PPP-B-636, type CF, class weather-resistant. Closure shall be effected by taping the top and bottom flaps with 2 inch wide tape conforming to PPP-T-76. Closure of bottom flaps with adhesive conforming to MIL-A-250 or metal stitches is permitted.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

5.1.4 Level C. Unit shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packaging shall be level A, B, or C as specified see (6.2).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Unit package. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated ofwater-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:55Z
Check the source to verify that this is the currentversion before use.

6. NOTES

6.1 Intended use. Box covered by this specification is for use in storage and shipping pathological and cytological slides.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable level of packaging and packing required (see 5.1 and 5.2).

6.3 This specification covers the following item appearing in the Federal Supply Catalog.

Federal Stock Number	Item Identification
6640-241-0138	Box, Microscope, Slide, Plastic, 5 Slide, 144s.

6.4 Maximum dimensions of slides used with this box are, length 76.6 mm, width 25.8 mm and thickness 1.35 mm. Minimum dimensions of slides used with this box are, length 74.2 mm, width 24.2 mm and thickness 0.75 mm.

6.5 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:55Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements .

SPECIFICATION

MIL-B-36630(DSA-DM) FSN:

ORGANIZATION

CITY AND STATE | CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)* | DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
MAR 68

EDITION AUG 66, WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:55Z
Check the source to verify that this is the current version before use.