MIL-C-36643A
Amendment-1
15 March 1974

MILITARY SPECIFICATION

CANNULA, TRACHEOTOMY, MODIFIED, JACKSON

This Amendment forms a part of Military Specification MIL-C-36643A dated 12 July 1971, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 4:

3.2 Style, design and dimensions (Type 1). Delete ninth sentence and substitute: "Pilot shall fit smoothly into outer tube and project through the distal opening of the outer tube without completely falling through the opening, and shall not bind in the opening."

Page 5:

3.4 Contractor's retention standards for interchangeability. Delete fourth and fifth sentences, and substitute: "The entire contract quantity shall be tested (100 percent inspection) against the acceptable returned Contractor's retention standards for interchangeability when tested as specified in 4.4.3.1, except the fit with the 6 Contractor's retention standards shall be on a 100 percent basis and not a random basis. The components of the tube assemblies (end item) that are not completely interchangeable with the acceptable Contractor's retention standards shall be rejected, but not the entire production lot."

3.5 Interchangeability of production lot. Delete entirely and substitute:

"3.5 Interchangeability of complete assembly (end item). In addition to the interchangeability requirements specified above, a complete assembly consisting of the outer tube, inner tube and pilot components (end item) shall be completely interchangeable when tested as specified in 4.4.3.2. Should the pilot and inner tube be interchangeable with the outer tube, that unit shall be acceptable providing the conditions specified in 3.4 are met. Should the pilot and inner tube components not be completely interchangeable with a randomly selected outer tube, the three components shall be rejected, but not the entire production lot. Interchangeability of a complete assembly concerns testing the produced items to be delivered to the Government and does not involve testing with the Contractor's retention samples.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:57Z
Check the source to verify that this is the current version before use.

Page 7:

4.2.2 For tests. Delete Table II entirely and substitute:

"Table II. Sampling Criteria for Tests

Component	Characteristic	Requirement	Test procedure	Insp. level	AQL
Cannula Assembly	Hardness	3.1.1	4.4.2	*	---
	Dimensions	3.2 & 3.2.1	-----	S-2	0.65
	Interchangeability test (Contractor's retention samples)	3.4	4.4.3.1	100%	**
	Interchangeability of complete assembly	3.5	4.4.3.2	100%	***
	Boil test	3.7	4.4.1.1	S-2	0.65
	Copper Sulfate test	3.7	4.4.1.2	S-2	0.65
	Saline test	3.7	4.4.1.3	S-2	0.65
	Sterilization test	3.7.1	4.4.4	S-1	4.0
Pilot	Salt-spray test	3.7	4.4.1.4	S-2	0.65

* A minimum of two samples shall be tested.
** Components not interchangeable with Contractor's retention samples shall be rejected.
***Complete assemblies (end item) that are not completely interchangeable shall be rejected."

Page 10:

4.4.3 Interchangeability test. Delete entirely and substitute.

"4.4.3 Interchangeability tests. Interchangeability tests shall be performed in two stages. The first stage involves all of the Contractor's retention samples (CRT) against the production lot and the second stage involves testing the interchangeability of a given complete assembly of the production lot.

"4.4.3.1 Interchangeability test of production lot with the CRT. The production lot shall be tested with the CRT (that have been approved by the Government). The inner and outer tubes of the production lot shall be completely separated, including the pilots. The inner tubes of the production lot shall be checked with all of the CRT outer tubes. The inner tube shall be easily, freely and smoothly, capable of insertion into, and removal from, the CRT outer tubes without exertion of force and evidence of sticking, binding, or improper seating when the inner tube is secured to the CRT outer tubes by means of the swivel lock. The swivel lock shall readily secure the inner tube into proper

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:57Z
Check the source to verify that this is the current version before use.

4.4.3.1 Interchangeability test of production lot with the CRT. (cont)

position without the inner tube extending beyond the distal end of the CRT outer tubes when the assembly is held in the position as shown on the drawing. The swivel lock shall secure the inner tube anchor plate (Type I assemblies) or the inner tube ring (Type II assemblies) when the swivel back is rotated clockwise and counterclockwise. Should all of the above conditions be met, these inner tubes of the production lot shall have passed the first stage of the interchangeability testing.

The outer tubes of the production lot shall be checked with all of the CRT inner tubes and CRT pilots. The CRT inner tubes shall properly fit into the outer tube of the production lot in exactly the manner specified above for checking the inner tubes of the production lot against the CRT outer tubes. The CRT pilots shall properly fit into the outer tubes of the production lot. This shall be accomplished by holding the retaining device at the proximal end of the CRT pilots with the thumb. A portion of the olive shaped body shall extend beyond the distal edge of the CRT outer tubes, without falling completely through the outer tube opening, and not bind in the opening. Should these conditions be met, the outer tubes of the production lot shall have passed the first stage of the interchangeability testing.

The pilots of the production lot shall be checked with all of the CRT outer tubes in exactly the manner specified above for checking the CRT pilots with outer tubes of the production lot. The pilots of the production lot that are interchangeable with the CRT outer tubes shall have passed the first stage of the interchangeability testing.

"4.4.3.2 Interchangeability of complete assembly (end item). Only those components of the production lot that meet the requirements of 3.4, when tested as specified in 4.4.3.1, shall be conditionally accepted. Final acceptance, with respect to interchangeability requirements, of a given complete assembly (end item) shall occur after they have passed the second stage of interchangeability testing described herein. Inner tubes, outer tubes and pilots shall be randomly assembled. The inner tube shall be easily, freely and smoothly, capable of insertion into, and removal from, the outer tube without exertion of force and evidence of sticking, binding or improper seating when the inner tube is secured to the outer tube by means of the swivel lock. The swivel lock shall readily secure the inner tube into proper position without the inner tube extending beyond the distal end of the outer tube when the assembly is held in the position as shown on the drawing. The swivel lock shall secure the inner tube anchor plate (Type I assemblies) or the inner tube ring (Type II assemblies) when the swivel lock is rotated clockwise and counterclockwise. The pilot shall properly fit into the outer tube in the same manner as specified in 4.4.3.1."

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:57Z
Check the source to verify that this is the current version before use.

Page 11:

5.1.3.2 Intermediate package. Delete and substitute:

"5.1.3.2 Intermediate package. Intermediate package shall be a box of appropriate size constructed in accordance with PPP-B-636, type CF, class domestic. Closure shall be adequate to prevent spilling of contents under normal handling."

5.2.1.1 Exterior container. At end of paragraph add: "Fiberboard boxes shall conform to the special requirements as specified in Table II of PPP-B-636.

Insert the following paragraph:

"5.2.1.2 Case liner. Each level A wood box shall be lined with a water-proof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard boxes shall be waterproofed in accordance with 30.4 of PPP-B-636.

5.2.2.1 Exterior container. At end of paragraph add: "Fiberboard boxes shall conform to the special requirements as specified in Table I of PPP-B-636."

Page 12:

6.2 Ordering data. Delete "Federal Stock Number" and substitute "National stock number" for category "(c)".

6.3 Delete "Federal Stock Number" and substitute: "National Stock Number". For each stock number stated, insert "-00" after "6515". Example: 6515-00-518-5461 shall replace previous stated stock number 6515-518-5461.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project Number 6515-1808

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:57Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-C-36643A

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:57Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:57Z
Check the source to verify that this is the current version before use.