MIL-C-36672
AMENDMENT-1
30 April 1973

MILITARY SPECIFICATION

CANNULA, TRACHEOTOMY, UNIVERSITY OF KANSAS DESIGN, METAL

This Amendment forms a part of Military Specification MIL-C-36672, dated 1 June 1970, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 3:

2.2 Other publications. In line 3, delete "requests" and substitute "request".

3.1 Material. In last sentence, delete "after proper stress relieving,".

3.2 Style, design and dimensions. In fourth sentence, delete "securely spot-welded", and substitute: "permanently secured".

Page 4:

3.4 Finish. Delete entirely and substitute:

"3.4 Finish. All components shall be polished to a smooth bright finish. Definitions of acceptable surface characteristics shall be as described in ASTM F86."

3.4.1 Passivation. Delete entirely and substitute:

"3.4.1 Passivation. Assemblies that are of welded construction shall be properly passivated after completion of all final finishing operations. Passivation shall be accomplished by the methods described in ASTM F86 or equivalent. Components of assemblies that are of silver-solder construction shall be passivated before they are silver-soldered."

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:00Z
Check the source to verify that this is the current version before use.

Add new paragraph:

3.4.2 Welding and silver-soldering. Assemblies shall be completely of welded construction (such as heliarc-weld). Welding rod shall be fabricated from type 304, 316 or 317, Class 3 material in accordance with GG-I-526. As an alternate, the outer sleeve shall be securely spot-welded to the retaining plate, and the inner tube shall be of silver-solder construction."

3.6 Identification marking. At end of paragraph add the following: "In addition, assemblies of the foreign manufacturer shall bear the country of origin."

4.1 Responsibility for inspection. Delete second sentence and substitute: "Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government."

Page 5:

Add the following paragraph:

"4.1.3 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate."

4.2.2 For tests. In Table II, under Test Procedure column, delete "4.4.1" and substitute "4.4.2" relating to hardness. Under Inspection level column, delete "S-4" and substitute "S-2", three places; under AQL column pertaining to Boil Test, Copper Sulfate Test and Saline Test, delete "0.65" and substitute "1.5".

Page 6:

4.3 Examination. In Table III, delete Major defect 111 entirely, and substitute: "111 Finish not as specified in 3.4." Delete last word in parenthesis for Major defect 112, and substitute: "identification".

Delete Minor defect 201 and substitute:

"201 Identification markings not as specified in 3.6".



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:00Z
Check the source to verify that this is the current version before use.

Page 7:

4.4.2 Hardness test. Delete period at end of sentence and substitute: "or method 244 (Diamond Pyramid)".

4.4.3 Interchangeability test. In line 3, delete "shipment to the Government", and substitute, "Government contract". In lines 4 and 6, delete "shipment" and substitute "Government contract".

4.4.5 Retention samples. In third line, insert "Support" after "Personnel" and before "Center".

5.1.3.2 Preservative. Delete "Not required" and substitute: "Preservation shall not be applied."

5.2.1.1 Exterior container. End of paragraph, add: "Special requirements of applicable table shall apply for PPP-B-636 box."

5.2.2.1 Exterior container. End of paragraph, add: "Fiberboard box shall be closed by method II of PPP-B-636. Special requirements of applicable table shall apply for PPP-B-636 box."

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Suppoy Agency - DM

Project No. 6515-1726

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.



Source: https://assist.dla.mil -- Downloaded: 20261O-21T09:00Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:00Z
Check the source to verify that this is the current version before use.