MIL-D-36622
19 January 1970
SUPERSEDING
MIL-D-2207
29 May 1950

MILITARY SPECIFICATION

DISH, BIOLOGICAL PREPARATION
(Stender Design)

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers one type of biological preparation dish suitable for use in biological laboratory procedures.

1.2 Classification. Biological preparation dishes covered by this specification shall be of the following sizes, as specified (see 6.2):

Size 1. - 50 Millimeter Diameter
Size 2. - 60 Millimeter Diameter

2. APPLICABLE DOCUMENTS.

2.1 Specifications, standards and drawings. The following specifications and standards, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

MMM-A-250	Adhesive, Water Resistant (For Sealing Waterproofed Paper)
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Box, Setup.
PPP-T-76	Tape, Pressure Sensitive Adhesive Paper, Water Resistant (For Carton Sealing)

FSC 6640

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:55Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

(Copies of Military specifications, standards, and drawings required by suppliers in connection with specific procurement function should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

C-148 A or B Method-Polariscopic Examination of Glass Containers

C-225 B-W Resistance of Glass Containers to Chemical Attack.Test for.

"Applications for copies should be addressed to American Society for Testing and Materials, 1916 Race Street, Phila., Pa. 19103".

3. REQUIREMENTS

3.1 Material.

3.1.1 Glass. Biological preparation dishes shall be fabricated of a good quality colorless, transparent glass.

3.2 Solubility. No evidence of alkali oxides shall be present in the bottle extract when tested as specified in 4.4.1.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:55Z
Check the source to verify that this is the current version before use.

3.3 Strain. The ring strain in the dishes shall not exceed temper No. 1 and any longitudinal strain shall be faint and highly diffused when tested as specified in 4.4.2.

3.4 Construction. The biological preparation dish shall consist of two parts, i.e., a dish or container and a cover and shall be furnished as a unit. The cover shall have a recessed circular groove that is a minimum of 0.4 millimeters (mm) in depth, extending around its periphery which shall be ground to mate with the rim of the container.

3.5 Dimensions and tolerances. Dimensions used in this specification are expressed in mm. Unless otherwise specified, tolerances for all dimensions shall be ± 3.0 mm.

3.5.1 Containers and covers. Dimensions for containers and covers shall be in accordance with table 1.

3.6 Identification marking. Dishes shall be legibly and permanently marked with the name or registered trademark of the manufacturer.

Table I. - Dimensions of Containers and covers

Size	Thickness of cover	Height without cover	Height with cover	Outside Diameter of Body	Outside Diameter of cover
1	7 ± 1	27	31	50	58
2	7 ± 1	96	99	60	72

3.7 Workmanship. Workmanship shall be first class throughout. Dishes shall be free of sharp edges, checks, ships blisters, stones, seeds and other defects which detract from their appearance or may impair its serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:55Z
Check the source to verify that this is the current version before use.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.2 Sampling.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in table II. Unit of product for examination purposes shall be the dish, as specified in 3.4.

Table II. Sampling for Examination

	Inspection level	AQL (defects Per hundred units)
For visual examination		
Major defects	II	2.5
Minor defects	II	4.0
For dimensional examination	S-2	2.5

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and Table II, except that the acceptance number shall be zero for all sample sizes. Unit of product for testing purposes shall be the dish, as specified in 3.4.

4.3 Examination. Dishes shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted in accordance with Table III. The dishes shall be examined for defects but shall not be restricted to the classified possible defects.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects:

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

Table III.
Categories and defects

Major
101 Component part missing.
102 Container and cover do not properly mate.
103 Unit not free of sharp edges.
104 Unit not free of Blisters and Boils.
105 Unit not free of checks, chips and crizzles.
106 Unit not free of feathers.
107 Unit not free of milkiness.
108 Unit not free of offset and rocker bottom.
109 Unit not free of splits and tears.

Minor
201 Unit not free of stones.
202 Unit not free of seed and laps.
203 Unit not free of dirt, grease and other foreign particles.
204 Name or registered trademark of manufacturer not legibly or permanently marked.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements. Unit of product for test purposes shall be the dish (see 3.4).

Table IV.

Unit of product	Characteristic	Requirement	Test Procedure	Inspection level
Dish	Solubility	3.2	4.4.1	S-2
Dish	Strain	3.3	4.4.2	S-2

4.4.1 Solubility. Dishes shall be tested to determine their resistance to chemical attack in accordance with Method B-W, ASTM Designation C225-65.

4.4.2 Strain. Dishes shall be tested for strain in accordance with Method A or B, ASTM Designation C 148-65.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:55Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One dish, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in the following table shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table V. Packaging Quantities

Size	Unit package	Intermediate package	Exterior container
1	1 unit	12 units	144 units
2	1 unit	6 units	72 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be packaged in a box of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676, except commercial colors will be acceptable. Box shall be designed with suitable interior fittings or shall contain sufficient cushioning material to insure against shifting of, or damage to the contents. Closure shall be adequate to prevent accidental opening.

5.1.3.2 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2 inch wide tape conforming to PPP-T-76. Closing of bottom flap with adhesive conforming to MMM-A-250 or metal stitches is permitted.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. Biological preparation dishes covered by this specification are intended for use in preparing and staining histological specimens in microbiological procedures.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Specify size dish required (see 1.2).
(c) Selection of applicable levels of packing (see 5.1 and 5.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

6.3 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meed requirements.

6.4 This specification covers the following items shown in the Federal Supply Catalog:

FSN	Item
6640-422-5000	DISH, BIOLOGICAL PREPARATION
6640-422-4000	DISH, BIOLOGICAL PREPARATION

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS. This sheet is to be filled out by [illegible] tion in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements.

SPECIFICATION
MIL-D-36622 FSN:

ORGANIZATION

CITY AND STATE | CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)* | DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:55Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE [illegible]
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

Fold

DPSC FORM 2418
MAR 68

EDITION AUG 66, WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.