MIL-L-36628B
4 August 1978
SUPERSEDING
MIL-L-36628A
20 June 1973 and
AMENDMENT 1
8 January 1974

# MILITARY SPECIFICATION

# LARYNGOSCOPE SET, MACINTOSH

This specification is approved for use by all Departments and Agencies of the Department of Defense.

## 1. SCOPE

1.1 Scope. This specification covers a detachable laryngoscope of the MacIntosh design, suitable for use in medical procedures concerning the human larynx.

## 2. APPLICABLE DOCUMENTS

* 2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bid or request for proposal form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

| | |
|---|---|
| GG-I-526 | - Instruments, Dental and Surgical; General Specifications. |
| PPP-B-566 | - Boxes, Folding, Paperboard. |
| PPP-B-585 | - Boxes, Wood, Wirebound. |
| PPP-B-601 | - Boxes, Wood, Cleated-Plywood. |
| PPP-B-621 | - Boxes, Wood, Nailed and Lock-Corner. |
| PPP-B-636 | - Boxes, Shipping, Fiberboard. |
| PPP-B-676 | - Boxes, Setup. |

MILITARY

| | |
|---|---|
| MIL-P-116 | - Preservation-Packaging, Methods of. |
| MIL-L-10547 | - Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible. |

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, CODE-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

DRAWINGS

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

1249 - Lamp, Incandescent, 2.5 Volt, 0.28 Ampere, Bulb T-1-1/2, Frosted.
20182 - Lamp, Incandescent, 2.7 Volt, 0.31 Ampere, Bulb T-1, Frosted.
20473 - Laryngoscope, Infant-Child-Adult, MacIntosh.
24276 - Handle, Laryngoscope, Battery Type, Universal, Large.

(Copies of specifications, standards, drawings and publications required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bid or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM E 18 - Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials, Tests for.

ASTM E 92 - Vickers Hardness of Metallic Materials, Test for.

ASTM E 140 - Hardness Conversion Tables for Metals (Relationship Between Brinell Hardness, Vickers Hardness, Rockwell Hardness, Rockwell Superficial Hardness and Knoop Hardness).

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.)

(Technical society and technical association specifications and standards are generally available for reference from libraries. They are also distributed among technical groups and using Federal agencies.)

3. REQUIREMENTS

3.1 Material.

3.1.1 Laryngoscope blade, housing lock, lamp carrier and lamp socket. The laryngoscope blade, housing lock, lamp carrier and lamp socket shall be fabricated from class 3, type 302, 303 or 304 material, or class 8, type 260, 270 or 360



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

material, all in accordance with GG-I-526. As an alternate, the blade and lamp carrier shall be fabricated from chromium plated over nickel plated copper alloy with the following composition:

| | |
|---|---|
| Copper | - 62.0 to 65.0 percent |
| Lead | - 0.5 percent maximum |
| Iron | - 0.5 percent maximum |
| Zinc | - Remainder (Total other elements: 0.5 percent maximum) |

As another alternate, the housing lock portion and lamp socket portion shall be fabricated from chromium plated over nickel plated copper alloy with the following composition:

| | |
|---|---|
| Copper | - 56.0 percent minimum |
| Lead | - 3.5 percent maximum |
| Iron | - 0.5 percent maximum |
| Zinc | - Remainder (Total other elements: 0.5 percent maximum) |

The plating and corrosion requirements for the alternate copper alloys shall be in accordance with GG-I-526 for class 8, type 260 material.

3.1.2 Battery handle. The battery handle shall be fabricated from class 8, type 260, 270 or 360 material in accordance with GG-I-526. As an alternate, the sleeve or shaft of the battery handle shall be fabricated from class 3, type 302, 303 or 304 material in accordance with GG-I-526. As an alternate, the internal surfaces of the handle, excluding removable threaded end cap, shall be nickel plated.

As another alternate, the battery handle shall be fabricated from chromium plated over nickel plated copper alloy with the following composition:

| Component | | Composition |
|---|---|---|
| Handle head | Copper: | 56.0 percent minimum |
| | Lead: | 3.0 percent maximum |
| | Iron: | 0.5 percent maximum |
| | Zinc: | Remainder (Total other elements: 0.5 percent maximum) |
| Handle barrel and end cap (Internal surfaces of the handle, excluding removable end cap, shall be nickel plated) | Copper: | 62.0 to 65.0 percent |
| | Lead: | 0.5 percent maximum |
| | Iron: | 0.5 percent maximum |
| | Zinc: | Remainder (Total other elements: 0.5 percent maximum) |

The plating and corrosion requirements for the alternate copper alloys shall be in accordance with GG-I-526 for class 8, type 260 material.

* 3.1.3 Lamps. The lamp shell base, and contact sleeve shall be fabricated from nickel plated, or chromium plated over nickel plated copper alloy with the following composition:

| | |
|---|---|
| Copper: | 56.0 percent minimum |
| Lead: | 3.7 percent maximum |
| Iron: | 0.35 percent maximum |
| Zinc: | Remainder (Total other elements: 0.5 percent maximum) |



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

quirements for class 8 material shall be ... material ... corrosion r...

The insulators for the lamps shall be fabricated from hard rubber, phenolic plastic or other material suitable for the purpose intended. The plating and corrosion tests shall be as specified for class B, type 360 material in accordance with GG-I-526.

3.2 Hardness. Rockwell hardness of class 3 material shall be B 75 minimum, when tested as specified in 4.4.1.

* 3.3 Style, design and dimensions. The complete laryngoscope set shall consist of a corrugated box (see 5.1.2.4) containing all of the following components in the quantities specified:

| Component | Size | Quantity |
|---|---|---|
| BLADE, LARYNGOSCOPE, Infant, MacIntosh, 87 mm (When furnished separately, National Stock Number (NSN) 6515-00-913-4605) | 1 | 1 |
| BLADE, LARYNGOSCOPE, Child, MacIntosh, 108 mm (When furnished separately, NSN 6515-00-955-8836) | 2 | 1 |
| BLADE, LARYNGOSCOPE, Medium Adult, MacIntosh, 130 mm (When furnished separately, NSN 6515-00-955-8865) | 3 | 1 |
| BLADE, LARYNGOSCOPE, Large Adult, MacIntosh, 158 mm (When furnished separately, NSN 6515-00-913-4607) | 4 | 1 |
| HANDLE, LARYNGOSCOPE, Battery Type, Universal, Large (When furnished separately,NSN 6515-00-890-2190) | - | 1 |
| LAMP, INCANDESCENT, 2.7 Volt, 0.31 Ampere, Bulb T-1, Frosted (When furnished separately, NSN 6240-00-552-9672) | Type I | 4 1/ (Includes extra lamps) |
| LAMP, INCANDESCENT, 2.5 Volt, 0.28 Ampere, Bulb T-1-1/2, Frosted (When furnished separately, NSN 6240-00-797-0420) | Type II | 4 1/ (Includes extra lamps) |

1/ Each laryngoscope set shall be furnished with one lamp assembled to each blade and one extra lamp for each blade. When laryngoscope blades are ordered as separate items of supply, no extra lamps shall be furnished.

* 3.3.1 Laryngoscope blade assemblies. The laryngoscope blade assemblies shall be fabricated from material specified in 3.1.1 and shall be furnished in the sizes and quantities specified in 3.3. The blade assemblies shall be of the MacIntosh design, and of the detachable and hook-on type suitable for use in medical procedures concerning the human larynx. The blades shall be capable of insertion onto the battery handle and designed to light at the distal end. The infant (size 1) and child (size 2) blades shall be furnished with Type I lamps; the medium adult (size 3) and large adult (size 4) blades shall be furnished with Type II lamps. Lamps shall be as specified in 3.3.3. The blade assemblies shall be in strict accordance with DPSC Drawing 20473.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

3.3.1.1 Internal threads of lamp socket. Internal threads of the lamp sockets for the size 1 and 2 blades shall allow the "GO" thread plug gage to be inserted and turned freely and shall not allow entry of the "NO GO" thread plug gage when tested as specified in 4.4.2. The same requirements shall apply for the lamp sockets of the size 3 and size 4 blades when tested as specified in 4.4.2.

* 3.3.2 Battery handle. Shall be a large size universal type laryngoscope battery handle with removable spring contact cap, capable of accommodating two "D" size batteries. Shall be suitable for use as a component or replacement handle for MacIntosh laryngoscope blade assemblies specified in 3.3.1. Material shall be as specified in 3.1.2. The battery handle shall also be suitable to accept laryngoscope blades of the Miller and Wisconsin designs. Style, design and dimensions of the handle shall be in strict accordance with DPSC Drawing 24276, for Type I or Type II design.

3.3.3 Lamps. Each laryngoscope blade specified in 3.3.1 shall contain one assembled frosted lamp. The lamps shall be fabricated from material specified in 3.1.3. When the complete laryngoscope set is furnished, two extra lamps, one each of Type I and Type II shall be supplied as specified in 3.3 and table I.

TABLE I. Lamp characteristics.

| Type | Volts | Amperes | Bulb type | Used with laryngoscope size |
|---|---|---|---|---|
| I | 2.7 | 0.31 | T-1 | 1 and 2 |
| II | 2.5 | 0.28 | T-1-1/2 | 3 and 4 |

Each lamp shall have a minimum 20 hour average life span and shall be designed to fit into the lamp sockets of the laryngoscope blades. The measured current through the lamp filament shall be within ± 10% of the ampere values specified in table I, when tested as specified in 4.4.4. The lamps shall be capable of sterilization without imparing their function or appearance when tested as specified in 4.4.5. The Type I lamps shall be in accordance with DPSC Drawing 20182, and the Type II lamps in accordance with DPSC Drawing 1249.

3.3.3.1 External threads of lamps. External threads of the required Type I lamps shall allow the "GO" thread ring gage to pass over the full length of the lamp threads and turn freely and shall not allow the "NO GO" thread ring gage to pass over the external lamp threads when tested as specified in 4.4.3. The same requirements shall apply for the Type II lamps when tested as specified in 4.4.3.

* 3.4 Finish. Finish of the laryngoscope blade assemblies shall be mirror and satin as designated on DPSC Drawing 20473. The battery handle shall have a bright finish and shall be knurled as shown on DPSC Drawing 24276. Finish on all edges and surfaces shall be uniform, free from burrs, sharp edges, pits, crevices, cracks, rough areas, grind marks, scale and corrosion.

* 3.5 Identification marking. Laryngoscope blade assemblies and battery handle shall be permanently and legibly marked with the contractor's name or registered trademark and country of origin when manufactured in countries outside of the United States. Blade assemblies shall be permanently and legibly marked with the blade name "MACINTOSH" or "MAC" and applicable blade size. Laryngoscope blades manufactured from class 3 material shall also be marked "stainless".



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

3.6 Workmanship. Workmanship shall be first class throughout. The laryngoscope blade assemblies, lamps and battery handles shall be free from defects which detract from their appearance or impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

* 4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

* 4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the blades, handles, and lamps may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

* 4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table II. The unit of product for examination purposes shall be one complete laryngoscope set, or one battery handle, or one laryngoscope blade, as applicable (see 3.3 and 6.2).

TABLE II. Sampling for examination.

| Type defect | Inspection level | AQL (percent defective) |
|---|---|---|
| Major | II | 1.0 |
| Minor | II | 2.5 |



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

* 4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table III. The unit of product for test purposes shall be one complete laryngoscope set, or one battery handle, or one laryngoscope blade, as applicable (see 3.3 and 6.2). Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

TABLE III. Sampling for tests.

| Component | Characteristic | Requirement | Test procedure | Inspection level | AQL (percent defective) |
|---|---|---|---|---|---|
| Laryngoscope blade assemblies | Dimensions | 3.3.1 | --- | S-1 | 2.5 |
| | Internal thread test | 3.3.1.1 | 4.4.2 | S-2 | 2.5 |
| | Hardness (For class 3 material) | 3.2 | 4.4.1 | S-2 | 2.5 |
| | Plating test (For class 8 material) | 3.1.1 and GG-I-526 | GG-I-526 | S-2 | 2.5 |
| | Salt spray test (For class 8 material) | 3.1.1 and GG-I-526 | GG-I-526 | S-2 | 2.5 |
| | Boil test (For all material) | 3.1.1 and GG-I-526 | GG-I-526 | S-2 | 2.5 |
| Battery handle | Dimensions | 3.3.2 | --- | S-1 | 2.5 |
| | Hardness (For class 3 material) | 3.2 | 4.4.1 | S-2 | 2.5 |
| | Plating test | 3.1.2 and GG-I-526 | GG-I-526 | S-2 | 2.5 |
| | Salt spray test (For class 8 material) | GG-I-526 | GG-I-526 | S-2 | 2.5 |
| | Boil test (For all material) | GG-I-526 | GG-I-526 | S-2 | 2.5 |
| Lamps | Plating test | 3.1.3 and GG-I-526 | GG-I-526 | S-1 | 2.5 |
| | Resistance to sterilization | 3.3.3 | 4.4.5 | S-2 | 2.5 |
| | Electrical properties | 3.3.3 | 4.4.4 | S-2 | 2.5 |
| | External thread test | 3.3.3.1 | 4.4.3 | S-2 | 2.5 |



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

* 4.2.3 Raw materials. Four ounces of each lot of each metal used in the manufacture of the laryngoscope sets, battery handle, or laryngoscope blade, as applicable, shall be selected for chemical analysis (see 3.3 and 6.2).

* 4.3 Examination.

* 4.3.1 Examination of laryngoscope. The laryngoscope sets shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted as specified in tables II and IV. Examination, however, shall not be restricted to defects listed in table IV. Any laryngoscopes in the sample containing one or more defects shall be rejected, and, if the number of defective items in any sample exceeds the acceptance number for that sample, the lot represented by that sample shall be rejected.

TABLE IV. Classification of defects.

| Categories | Defects |
|---|---|
| Major | |
| 101 | Lamps do not light when laryngoscope blade is inserted onto battery handle (when two "D" size batteries are inserted into battery handle). |
| 102 | Blades not of detachable, hook-on design with respect to battery handle. |
| 103 | Blades do not positively seat (hook-on) when inserted into battery handle. |
| 104 | Blades not of finish specified. |
| 105 | Blades not free from sharp edges, crevices, burrs, grind marks, or cracks. |
| 106 | Internal threads of lamp socket not complete or not well defined. |
| 107 | External threads of lamp not complete or not well defined. |
| 108 | Extra lamps of each type not furnished with complete laryngoscope set. |
| 109 | Battery handle finish not bright. |
| 110 | Battery handle cannot accommodate two standard "D" size batteries. |
| 111 | Glass envelope of lamp not properly secured to lamp base. |
| 112 | Glass envelope of lamp cracked. |
| 113 | Plated parts not free from flaking. |
| 114 | Blades and battery handle not permanently and legibly marked as specified in 3.5. |
| Minor | |
| 201 | Battery handle inner surfaces not plated. |
| 202 | Glass envelope of lamp not free from bubbles, dirt or other imperfections. |

* 4.3.2 Packaging inspection. The inspection of the preservation-packaging and interior package markings shall be in accordance with the group A and B quality conformance inspection requirements of MIL-P-116. The inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

* 4.4.1 Hardness test. Rockwell hardness of class 3 material shall be determined in accordance with ASTM E 18 or E 92. Conversion of Vickers hardness values to Rockwell values shall be in accordance with ASTM E 140.

* 4.4.2 Internal thread test of lamp sockets. The contractor shall provide the necessary thread plug gages to the Government Quality Assurance Representative (QAR) for checking the internal threads of the lamp sockets in accordance with the required thread dimensions for the appropriate lamp size shown on drawing. The following dimensional parameters shall be inscribed on the gages:

For size 1 and size 2 Laryngoscope Blades:

Thread Size: 1/8-72NS-3
'GO' PITCH DIAMETER - 0.1160
'NO GO' PITCH DIAMETER - 0.1174

For size 3 and size 4 Laryngoscope Blades:

Thread Size: 8-32 UNC-2B
'GO' PITCH DIAMETER - 0.1437
'NO GO' PITCH DIAMETER - 0.1475

To perform thread tests of appropriate lamp sockets, remove lamp from lamp socket and place plug gage in lamp socket. Remove "GO" gage and place "NO GO" plug gage in lamp socket. Fit of gages into lamp sockets shall be noted.

4.4.3 External thread tests of lamps. The contractor shall provide the necessary thread ring gages to the QAR for checking the external threads of the lamps in accordance with the required thread dimensions for the appropriate NSN shown on drawing. The following dimensional parameters shall be inscribed on the ring gages:

For Type I Lamps:

Thread Size: 1/8-72NS-3A
Major Diameter: 0.1214 - 0.1250
'GO' PITCH DIAMETER - 0.1160
'NO GO' PITCH DIAMETER - 0.1146

For Type II Lamps:

Thread Size: 8-32 UNC-2A
Major Diameter: 0.1571 - 0.1631
'GO' PITCH DIAMETER - 0.1428
'NO GO' PITCH DIAMETER - 0.1399

To perform thread tests for appropriate lamps, remove lamp from lamp socket and place "GO" ring gage over external threads on lamp. Remove "GO" gage and place "NO GO" ring gage over external threads on lamp. Fits of gages onto lamp threads shall be noted.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

4.4.4 Operation test for lamps. A sample incandescent lamp of each type shall be installed into an appropriate receptacle and shall be illuminated by the application of voltage which shall be within ± 10% of the voltage specified in table I (3.3.3). In addition, the circuit current shall be measured with the use of a suitable instrument which shall not load the circuit. The lamp shall be lighted at rated voltage for a period of not less than 20 hours. This shall be accomplished in increments of not less than 4 hours each, until 20 hours have been accumulated.

4.4.5 Sterilization test for lamp. A sample lamp shall be completely immersed in a 95% ethyl alcohol solution for a minimum period of 6 hours. After soaking, the lamp shall be dried and inspected for any internal absorption of the solution.

## 5. PACKAGING

5.1 Preservation-packing. Preservation-packing shall be level A or C, as specified (see 6.1).

5.1.1 Unit of issue. One laryngoscope set or one blade or one handle, as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Cleaning. Laryngoscope sets or blades and handles shall be cleaned in accordance with MIL-P-116, process C-1.

5.1.2.2 Drying. Laryngoscope sets or blades and handles shall be dried in accordance with MIL-P-116.

5.1.2.3 Preservation application. Preservatives shall not be used.

5.1.2.4 Unit packaging.

5.1.2.4.1 Laryngoscope set. Components comprising each set along with extra lamps shall be wrapped in neutral tissue or other suitable non-corrosive material and individually packaged in a box of appropriate size and design. Each complete set shall be further packaged in accordance with MIL-P-116, method IC-2. Interior container shall be a corrugated box (see 3.3).

5.1.2.4.2 Blades or handles. Each blade or each handle shall be wrapped in neutral tissue or other suitable non-corrosive material and packaged in accordance with MIL-P-116, method IC-2.

5.1.2.5 Intermediate package. Twelve unit packages containing blades or twelve unit packages containing handles shall be packaged in a box conforming to PPP-B-566, PPP-B-636, class domestic or PPP-B-676. Closure shall be as specified in the appendix of the box specification. An intermediate package shall not be required for complete sets.

5.1.3 Level C. Each laryngoscope set shall be clean, dry and packaged in a manner that will afford adequate protection against corrosion, deterioration and physical damage during shipment from the supply source to the first receiving activity.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.1).

5.2.1 Level A. Twenty-four unit packages containing complete sets or twelve intermediate packages (144 units) containing blades or eight intermediate packages (96 units) containing handles shall be packed in a shipping container designed for a type 2 load and conforming to PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. Twenty-four unit packages containing complete sets or twelve intermediate packages (144 units) containing blades or eight intermediate packages (96 units) containing handles shall be packed in a shipping container designed for a type 2 load and conforming to PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1 or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The packaged sets, blades or handles, shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the carrier's rules and regulations.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Shipping container. Shipping container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents shall specify the following:

a. Title, number and date of this specification.
b. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
c. NSN of complete laryngoscope set, or laryngoscope blade, or battery handle, as applicable (see 3.3 and 6.2).



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

6.2 This specification covers the following items appearing in the Federal Supply Catalog:

| National Stock Number | Item Identification |
|---|---|
| 6515-00-616-5052 | LARYNGOSCOPE, Infant-Child-Adult, MacIntosh. |
| 6515-00-890-2190 | HANDLE, LARYNGOSCOPE, Battery Type, Universal, Large. |
| 6515-00-913-4605 | BLADE, LARYNGOSCOPE, Infant, MacIntosh, 87 mm. |
| 6515-00-913-4607 | BLADE, LARYNGOSCOPE, Large Adult, MacIntosh, 158 mm. |
| 6515-00-955-8836 | BLADE, LARYNGOSCOPE, Child, MacIntosh 108 mm. |
| 6515-00-955-8865 | BLADE, LARYNGOSCOPE, Medium Adult, MacIntosh, 130 mm. |

6.3 Changes from previous issue. The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Offerors and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activity:
DOD-MB

Preparing activity:
DLA-DM

Project No. 6515-2115



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

## STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS: This form is provided to solicit beneficial comments which may improve this document and enhance its use. DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity. A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed.

NOTE: This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

DOCUMENT IDENTIFIER *(Number)* AND TITLE

MIL-L-36628B LARYNGOSCOPE SET, MACINTOSH

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1. ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID, RESTRICTIVE, LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW.

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDED WORDING CHANGE

C. REASON FOR RECOMMENDED CHANGE(S)

2. REMARKS

SUBMITTED BY *(Printed or typed name and address — Optional)*

TELEPHONE NO.

DATE

DD FORM 1426 1 OCT 76

EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

POSTAGE AND FEES PAID

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

DEFENSE PERSONNEL SUPPORT CENTER
DIRECTORATE OF MEDICAL MATERIEL, CODE-ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.