MIL-S-36629A
13 May 1977
~~MIL-S-36629~~
19 February 1970

MILITARY SPECIFICATION

SCISSORS, BANDAGE, ANGULAR, 7-1/4 INCH

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a heavy duty bandage scissors, suitable for cutting heavy or multiple layers of cloth and zippered flight clothing.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

FEDERAL

GG-I-526	-	Instruments, Dental and Surgical; General Specifications.
PPP-B-566	-	Boxes, Folding, Paperboard.
PPP-B-585	-	Boxes, Wood, Wirebound.
PPP-B-601	-	Boxes, Wood, Cleated-Plywood.
PPP-B-621	-	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	-	Boxes, Shipping, Fiberboard.
PPP-B-676	-	Boxes, Set-Up.
V-P-106	-	Fastener, Slide, Interlocking.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Hq, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:55Z
Check the source to verify that this is the current version before use.

Military

MIL-P-116 - Preservation, Methods of
MIL-L-10547 - Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexbile.
MIL-W-530 - Webbing, Textile, Cotton, General Purpose, Natural or in Colors.

STANDARDS

Military

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)
No. 23175 SCISSORS, BANDAGE, ANGULAR, 7-1/4 Inch.

(Copies of specifications, standards, drawings, and publications required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS

D 638 or D 651	Tensile Strength.
D 747	Modulus of Elasticity.
D 785	Rockwell Hardness.
D 792	Specific Gravity.

(Applications for copies should be addressed to The American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103).

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:55Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Material.

3.1.1 Blades and shanks. The scissors, blades and shanks shall be fabricated from Class 4, type 420 material in accordance with GG-I-526, except that the carbon content shall be 0.18 percent minimum. After proper stress relieving, the Rockwell hardness of the blades and shanks shall be C45 to C50. The rivet shall be fabricated from austenitic or martensitic corrosion-resisting steel. Rivets fabricated from austenitic corrosion-resisting steel shall have a maximum carbon content of 0.08 percent; a minimum nickel content of 10.0 percent; and a minimum chromium content of 16.0 percent. All metallic material shall be properly passivated and shall pass the boil and copper sulfate tests specified in GG-I-526.

3.1.2 Finger rings. The scissor finger rings shall be fabricated from black, molded, polypropylene. The material shall be safe (nontoxic) and shall not separate from the shanks of the blades, soften, or be adversely affected by exposure to temperatures up to 290° Fahrenheit (143° Celsius). The following physical requirements shall apply for the polypropylene when tested as specified in 4.4.1:

Crystallinity Melting Point	160° Celsius.
Tensile Strength	4450 pounds per square inch (30.5 Megapascals)
Tear Strength or Tenacity	4600 pounds per square inch (31.5 Megapascals)
Modulus of Elasticity	178,000 pounds per square inch (1225 Megapascals)
Specific Gravity	0.90
Rockwell Hardness	R 85

3.2 Style, design and dimensions. Shall be an angular bandage scissors with molded, black, polypropylene finger rings. Scissors shall be capable of cutting heavy or multiple layers of clothing and all zippers used in military flight clothing. Each half of the scissors shall be formed with a continuous longitudinal stiffening rib smoothly blending into a circular rib at the pivoting point. The scissors half, containing the thumb finger, shall have a smoothly finished probe tip and a finely serrated cutting blade. The remaining scissors half shall have a rounded tip. The rings of the scissors shall be designed to permit use of the scissors when wearing heavy gloves. The scissors shall be in accordance with DPSC Drawing No. 23175.

3.3 Finish. Finish of the blades and shanks shall be as indicated on the drawing. All non-cutting edges shall be free from burrs and sharp edges. Molded finger rings shall be free from sharp edges and burrs.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

3.4 Sterilization. The entire scissors shall be capable of being autoclaved without showing any damage to metal components or to the molded, polypropylene finger rings when tested as specified in 4.4.2. The molded, polypropylene shall not separate or loosen from the steel shanks after the specified test is performed. In addition, there shall be no evidence of blooming or discoloration of the plastic as a result of this test.

3.5 Scissors Performance. The scissors, when tested as specified in 4.4.3, shall completely cut through the slide, interlocking, fastener (zipper) and through the cotton webbing (belting). The cutting requirement shall apply for the entire length of the specified cuts, including the scissor tips. All three cuts, through each material, made with each scissors tested shall meet the requirements of this specification.

3.6 Identification marking. Scissors manufactured in U.S.A. shall be permanently marked with the name or registered trademark of the contractor and "stainless." Scissors of foreign (other than U.S.) manufacture shall be marked with the name of the country of origin, together with the name or registered trademark of the contractor and "stainless."

3.7 Scissors weight. The finished scissors, less packing and packaging, shall not weigh more than 2.2 ounces (68.5 grams).

3.8 Workmanship. Workmanship shall be first class throughout. Scissors shall be free from defects which detract from their appearance or impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure suppliers and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:55Z
Check the source to verify that this is the currentversion before use.

4.1.2 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.3 Certificates of quality. Certificates of quality supplied by the manufacturer of the steel and polypropylene may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number (NSN), the item identification, the name of the component and material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as indicated in table I. The unit of product for examination purposes shall be one scissors.

TABLE I. Sampling criteria for examination.

Type defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in table II. The unit of product for test purposes shall be one scissors. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

TABLE II. Sampling criteria for tests.

Component	Characteristic	Requirement	Procedure	Inspection level	AQL (percent defective)
Scissors	Hardness	3.1.1	GG-I-526	S-2	1.5
	Copper Sulfate	3.1.1	GG-I-526	S-2	1.5
	Boil Test	3.1.1	GG-I-526	S-2	1.5
	Physical properties (Polypropylene)	3.1.2	4.4.1	S-2	1.5
	Sterilization test	3.4	4.4.2	S-2	1.5
	Performance test	3.5	4.4.3	S-3	1.0

4.2.3 Raw materials. A sufficient quantity of each lot of each metal used in the construction of the scissors shall be selected for chemical analysis.

4.3 Examination. The scissors shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted as specified in tables I and III. Examination, however, shall not be restricted to defects listed in table III. Any scissors in the sample containing one or more defects shall be rejected and if the number of defective items in any sample exceeds the acceptance number for the sample, the lot represented by the sample shall be rejected.

TABLE III. Classification of defects

Categories	Defects
Critical	None defined.
Major:	
101	Molded plastic handles not firmly secured to shanks.
102	Non-cutting edges not free from sharp edges or burrs.
103	Finish not as specified.
104	Plastic rings not free from flash, burrs, or sharp edges.
105	Shanks not completely enclosed in plastic finger rings.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects (continued)

Categories	Defects
Major:	
106	Surfaces of shanks and finger rings not free from pits, cracks, slivers, scale or corrosion.
107	Shanks not free from excessive play at the fastening point in the closed position.
Minor:	
201	Scissors not permanently identified as specified in 3.6.

4.3.1 Packaging inspection. The inspection of the preservation-packaging and interior package marking shall be in accordance with the group A and B quality conformance inspection requirements of MIL-P-116. The inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Physical properties. The following physical properties, specified in 3.1.1, shall be tested in accordance with the following ASTM Test Methods.

Physical Properties	ASTM Test Method
Tensile strength	D 638, D 651
Modulus of elasticity	D 747
Specific gravity	D 792
Rockwell hardness	D 785

Certification of the above physical properties by the manufacturer of the polypropylene shall be acceptable, provided that the contractor has previously supplied acceptable scissors to the Government and that there have been no major changes in the manufacturing operations, the material composition or supplier of raw materials.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:55Z
Check the source to verify that this is the current version before use.

4.4.2 Sterilization test. Sample scissors shall be placed into a suitable sterilizer (container) and steam sterilized at a temperature of 272 degrees Fahrenheit (136°C) for a period of 10 minutes. The scissors shall then be removed from the sterilizer and allowed to cool to room temperature. This test shall be performed a total of 10 times.

4.4.3 Scissors performance test. The test materials shall be in accordance with type I, MHS size, (any style) of Federal Specification V-F-106 Fastener, Slide, Interlocking and type IV Webbing, 1-1/4 inches nominal width, of Military Specification MIL-W-530 Webbing, Textile, Cotton, General Purpose Scissors to be tested shall be cleaned prior to test performance. Three separate, consecutive tests shall be performed for each scissors on each test material. The same scissors shall be used for cutting fastener and webbing material. Each test shall consist of making one cut across the chain and through the scoops of the fastener (zipper), with the fastener in the closed position, and one cut through the webbing (belting). The cut through the webbing shall be across the width of the webbing. Each cut shall be made using the scissors in the normal manner (with the examiner's fingers in the finger rings of the scissors).

5. PACKAGING

5.1 Preservation-packaging. Preservation-packaging shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Cleaning. Scissors shall be cleaned in accordance with MIL-P-116, process C-1.

5.1.1.2 Drying. Scissors shall be dried in accordance with MIL-P-116.

5.1.1.3 Preservative application. Preservatives shall not be used.

5.1.1.4 Unit packaging. Each pair of scissors shall be packaged in accordance with MIL-P-116, method IC-1.

5.1.1.5 Intermediate package. Twelve unit packages shall be packaged in a box conforming to PPP-B-566, PPP-B-636 class domestic or PPP-B-676. Closure shall be as specified in the appendix of the applicable box specification.

5.1.2 Level C. Scissor shall be clean, dry and individually packaged in a manner that will afford adequate protection against corrosion, deterioration and physical damage during shipment from the supply source to the first receiving activity.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.2).

5.2.1 Level A. Twenty-four intermediate packages (288 units) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.2 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.2 Level B. Twenty-four intermediate packages shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The packaged scissors shall be packed in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the applicable carrier's rules and regulations.

5.3 Marking.

5.3.1 Interior packages. Each unit and intermediate package shall be marked as specified in MIL-STD-129. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior containers shall be marked as specified in MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

a. Title, number, and date of this specification.
b. Selection of applicable levels of packing and packaging (see 5.1 and 5.2).
c. National stock number (see 6.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

6.2 This specification covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6515-00-935-7138	SCISSORS, BANDAGE, ANGULAR, 7-1/4 Inch

6.3 Specification changes. Asterisks to indicate changes in this specification have been omitted due to the extensiveness of the changes.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Logistic Agency - DM

Project No. 6515-2066

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:55Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

O. FICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
hiladelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:55Z
Check the source to verify that this is the current version before use.

Plastic envelope.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*: _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:55Z
Check the source to verify that this is the current version before use.

cal and chemical

2.5 Performance — when measured as spec... technical means of proof