MIL-S-36635A
1 August 1975
SUPERSEDING
MIL-S-36635
26 January 1970

MILITARY SPECIFICATION

SAW, FINGER RING

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a manually operable saw suitable for cutting metal finger rings.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
~~PPP-B-601~~	~~Boxes, Wood, Cleated-Plywood.~~
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Boxes, Shipping, Fiberboard.
PPP-B-676	Boxes, Set-up.

Military

MIL-P-116	Preservation, Methods of.
MIL-L-10547	Liners, Case and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:56Z
Check the source to verify that this is the current version before use.

DRAWING

Defense Personnel Support Center (DPSC)
No. 1820 SAW, FINGER RING AND BLADE

(Copies of specifications, standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. The blade of the saw shall be fabricated from unplated high carbon steel. The bearing washer and hub shall be fabricated from austenitic or martensitic carbon-resisting steel. The handle, tip and lever shall be fabricated from austenitic corrosion-resisting steel, martensitic corrosion-resisting steel, or chromium plated over nickel plated copper alloy. The thumbpiece shall be fabricated from corrosion-resisting steel or from a combination of molded reinforced nylon and austenitic corrosion-resisting steel. As an alternate, the handle shall be fabricated from molded reinforced nylon. All other fastening devices such as screws, pins or rivets shall be fabricated from austenitic or martensitic corrosion-resisting steel.

3.2 Style, design and dimensions. The finger ring saw shall be designed to manually cut metal finger rings from the human hand without damage to patient's fingers. The assembly shall consist of a blade that is accurately fitted to a bearing hub and bearing washer that is secured to the handle proper by means of a thumbpiece. The hub shall be designed to form a tight to interference fit (maximum 0.002 inch) with the blade. The thumbpiece shall be readily removed from the handle and bearing hub for easy replacement of the blade. The lever shall be removable, secured to the handle and designed to slip under a finger ring for the cutting of finger rings by manually turning the thumbpiece. The lever slot shall be in alignment with the blade, without the blade butting, jamming or hitting the lever when the lever is moved toward the blade. The washer, blade and hub components shall be readily assembled to each other in any order (i.e. both blade surfaces and both washer surfaces shall be capable of being placed onto the hub). For saws of style 1 design, the blade shall be prevented from cutting through the lever by having the lever tip coming to a stop at the saw hub and washer. For saws of style 2 design, the lever shall be capable of restraining the blade for blade replacement, by manually turning the thumbpiece in a counter-clockwise direction. The blade shall possess well defined teeth, consistent depth and be in accordance with good manufacturing practice. The finger ring saw assembly, including the blade (see 6.2) shall be in accordance with DPSC Drawing No. 1820, for style 1 or style 2 design.

3.3 Hardness. Rockwell hardness of the blade shall be 73 to 79 on the 30-N scale. Rockwell hardness of the hub and washer shall be C36 minimum when fabricated from martensitic corrosion-resisting steel after proper stress relieving, and B 75 minimum when fabricated from austenitic corrosion-resisting steel. The hardness tests shall be performed as specified in [illegible]

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:56Z
Check the source to verify that this is the current version before use.

3.4 Performance test. When tested as specified in 4.4.1 the saw shall cut completely through the test specimen within two complete revolutions without breaking the blade teeth.

3.5 Plating and corrosion requirements. The plating and corrosion requirements for components fabricated from copper alloy shall be in accordance with Federal Specification GG-I-526 for class 8, type 260 material, except that the corrosion test shall be the boil test only. Except for the blade, components fabricated from carbon steel shall comply with the plating and corrosion tests for class 1 material of GG-I-526, except that the corrosion test shall be the boil test only. Components fabricated from austenitic or martensitic corrosion-resisting steel shall comply with the corrosion tests specified for class 3 and class 4 material, respectively, of GG-I-526.

3.6 Finish. The finish of the finger ring saw assembly shall be smooth and uniform, free from cracks, burrs, sharp edges (except for blade teeth), pits, scale and corrosion.

3.7 Passivation. All components fabricated from austenitic or martensitic corrosion-resisting steel shall be properly passivated.

3.8 Identification marking. The finger ring saw assembly shall be permanently and legibly marked with the name or registered trademark of the contractor. In addition to the above, assemblies of foreign manufacture (other than U.S.), shall also be marked with the country of origin.

3.9 Workmanship. Workmanship shall be first class throughout. Finger ring saw assemblies and blades shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:56Z
Check the source to verify that this is the currentversion before use.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality supplied by the manufacturer of the corrosion-resistant steel alloys, carbon steel alloys, copper alloys and reinforced nylon may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, the individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as indicated in Table I. The unit of product for examination purposes shall be one complete finger ring saw assembly or blade (see (6.2).

TABLE I Sampling Criteria For Examination

Type defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II. The unit of product for test purposes shall be one complete finger ring saw assembly or blade (see 6.2). Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

TABLE II. Sampling Criteria for Tests

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL
Saw assembly	Dimensions	3.2	----------	S-2	2.5
	Hardness	3.3	GG-I-526	S-2	1.5
	Performance test	3.4	4.4.1	S-1	4.0

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:56Z
Check the source to verify that this is the current version before use.

TABLE II. Sampling Criteria for Tests (cont'd)

Component	Characteristic	Requirements	Test Procedure	Inspection level	AQL
Saw assembly	Plating test	3.4	GG-I-526	S-1	2.5
	Boil test	3.4	GG-I-526	S-2	1.5
	Copper sulfate test	3.4	GG-I-526	S-2	1.5
Blade (when furnished separately)	Dimensions	3.2	-------	S-2	2.5
	Hardness	3.3	GG-I-526	S-2	1.5

4.2.3 Raw materials. A sufficient quantity of each lot of each metal used in the manufacture of the saw shall be selected for chemical analysis.

4.3 Examination.

4.3.1 Examination of finger ring saw assemblies and blades (see 6.2). The finger ring saw assemblies and blades shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in Tables I and III. Examination, however, shall not be restricted to defects listed in Table III. Any saw assemblies or blades in the sample containing one or more defects shall be rejected and if the number of defective items in any sample exceeds the acceptance number for that sample, the lot represented by that sample shall be rejected.

TABLE III. Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Saws not of the design specified.
102	Saw not free from sharp edges (except for blade), burrs, pits, cracks, scale and corrosion.
103	Saw assembly not complete.
104	Saw not capable of manual rotation.
105	Blade faces not smooth.
106	Blade does not align properly with lever slot.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:56Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of Defects (cont'd)

Categories	Defects
107	Slot and teeth of blade not free from burrs and not well defined.
108	Threads of thumbpiece and hub not well defined.
109	Washer, blade and hub components not capable of easy assembly as specified in 3.2.
110	Blade teeth broken, missing or dull.
Minor	
201	Saw assemblies not identified as specified in 3.7

4.3.2 Examination of preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

TABLE IV. Examination for Preparation for Delivery

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging and packing materials	Any component missing or damaged.
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Performance test. A sufficient straight length of 1/16 inch minimum diameter, brass wire or rod of quarter-hard [illegible] specimen. The test specimen shall be free from cracks, pits or other evidence of unsuitability. The test specimen shall be securely placed between the lever tip and blade. The saw thumbpiece shall be turned using nominal technique (without excessive pressure) in a clockwise manner. To minimize slippage of the test specimen, the test specimen may be appropriately restrained.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:56Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be Level A or C as specified (see 6.1).

5.1.2 Level A.

5.1.2.1 Unit preparation. Each finger ring saw or blade shall be cleaned and dried in accordance with the applicable method as specified in MIL-P-116. A preservative shall not be utilized on the items, and each item shall be wrapped in neutral tissue or other suitable anticorrosive material.

5.1.2.2 Unit package. Each finger ring saw shall be packaged in accordance with MIL-P-116, method 1C2. Each blade shall be packaged in accordance with MIL-P-116, method 1C1.

5.1.2.3 Intermediate package. Twelve (12) finger rings saws or twelve (12) blades shall be packaged in a box of appropriate size constructed in accordance with PPP-B-566, PPP-B-636, or PPP-B-676, type CF, class domestic. Closure shall be adequate to prevent accidental opening.

5.1.3 Level C. The finger ring saws or blades shall be packaged in standard commercial containers of the size and kind commonly used which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be Level A, B or C as specified (see 6.1).

5.2.1 Level A.

5.2.1.1 Exterior container. Two hundred eighty eight (288) unit packages (24 intermediate packages) of finger ring saws; or seventeen hundred twenty eight (1728) unit packages (144 intermediate packages) of blades shall be packed in an exterior container designed for type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.2 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed in accordance with 30.4 of PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:56Z
Check the source to verify that this is the current version before use.

5.2.2 Level B.

5.2.2.1 Exterior container. Two hundred eighty eight (288) unit packages (24 intermediate packages) of finger ring saws or seventeen hundred twenty eight (1728) unit packages (144 intermediate packages) of blades shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636 type CF, class domestic. Closure shall be as specified in the appendix of the applicable box specifications.

Closure of fiberboard boxes shall conform to Method I or II as specified PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit and intermediate packages shall be marked in accordance with MIL-STD-129. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

a. Title, number and date of this specification.
b. Selection of applicable levels of packaging and packing (see 5.1 & 5.2).
c. Marking requirements.
d. National Stock Number.

6.2 This specification is intended to cover the following items listed in the Federal Supply Catalog:

National Stock Number (NSN)	Item Identification
6515-00-363-4100	SAW, FINGER RING
6515-00-363-4150	BLADE, FINGER RING SAW

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:56Z
Check the source to verify that this is the current version before use.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1932
CS

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:56Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:56Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No 22 R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers, vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-S-36635A SAW, FINGER RING

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1 HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
A GIVE PARAGRAPH NUMBER AND WORDING

B RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2 COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3 IS THE DOCUMENT RESTRICTIVE?
YES NO (If "Yes", in what way?)

4 REMARKS

SUBMITTED BY (Printed or typed name and address - Optional)

TELEPHONE NO

DATE

DD FORM 1426

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:56Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:56Z
Check the source to verify that this is the current version before use.