MIL-S-36649
11 March 1970

MILITARY SPECIFICATION

SPECULUM, EYE, GUYTON-PARK, NONMAGNETIC, 3-1/2 INCH

This specification is mandatory for use by all Departments and Agencies of the Department of Defense

1. SCOPE

1.1 Scope This specification covers a nonmagnetic eye speculum, of the Guyton-Park design, for use in eye surgery.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
MMM-A-250	Adhesive, Water-Resistant, (for sealing fiberboard boxes).
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, (for carton sealing).

Military

MIL-P-116	Preservation, Methods of.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:57Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

1992 SPECULUM, EYE, Guyton-Park Nonmagnetic, 3-1/2 inch.

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 General Requirements. The eye speculum shall be in accordance with the applicable requirements of GG-I-526, and as specified herein.

3.2 Material. The entire eye speculum shall be fabricated from Class 3, Type 316 material. The spring shall be spring temper wire.

3.3 Hardness. Rockwell hardness of the shanks, blades, and canthus member shall be B75 to B85.

3.4 Style, design and dimensions. Shall be a nonmagnetic, adjustable eye speculum of the Guyton-Park design. Shall be suitable for use with FSN 6515-299-8008 MAGNET, EYE, ELECTRIC involving surgical procedures of the human eye. Speculum blades shall be capable of swiveling 360 degrees. Speculum blades shall be capable of being locked in any position of travel of adjusting mechanism. Speculum shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 1992.

3.5 Finish. Finish shall be mirror or satin.

3.5.1 Passivation. All components of the speculum shall be properly passivated prior to final assembly.

3.6 Nonmagnetic characteristics. The complete assembled eye speculum shall show no magnetic response when tested as specified in 4.4.1.

3.7 Identification marking. In addition to GG-I-526 requirements, instruments of foreign manufacture shall also be marked "stainless".

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:57Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Shall be in accordance with GG-I-526.

4.1.1 Inspection. Shall be in accordance with GG-I-526.

4.1.2 Certificates of quality. Certificates of quality, supplied by the manufacturer of the corrosion resisting steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in Table I. The unit of product for examination purposes shall be one eye speculum.

Table I. Sampling Criteria for Examination

Type of defect	Inspection level	AQL (percent defective
Major	II	1.0
Minor	II	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II. Unit of product for test purposes shall be one speculum.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:57Z
Check the source to verify that this is the currentversion before use.

Table II. Sampling Criteria for Tests

Component	Characteristic	Reqt.	Test procedure	Insp level	AQL
Speculum	Composition analysis	3.2 and GG-I-526	GG-I-526	**	*
	Hardness	3.3 and GG-I-526	GG-I-526	S-2	*
	Dimensions	3.4	---	S-1	2.5
	Boil Test	GG-I-526	GG-I-526	S-2	*
	Nonmagnetic Test	3.6	4.4.1	S-2	*

* Acceptance number shall be zero for all sample sizes.
** Four ounces of each lot of metal used in construction of the speculums shall be selected for chemical analysis.

4.2.3 Sample unit usage. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The speculums shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in Tables I and III. Inspection, however, shall not be restricted to defects listed in Table III. Any speculums in the sample containing one or more defects shall be rejected and if the number of defective items in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

Table III. Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Speculum not completely assembled.
102	Speculum not free from sharp edges, rough areas, burrs, cracks, tool marks and pits.
103	Finish not as specified.
104	Blades of speculum not capable of swiveling 360°.
105	Action of adjusting nut too loose or too tight.
Minor	
201	Identification marking not present, correct, complete or legible.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:57Z
Check the source to verify that this is the currentversion before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Nonmagnetic test. The eye speculums shall be placed at rest, laying flat, upon a smooth nonferrous surface. An Alnico horseshoe-shaped magnet, weighing between 0.131 and 0.145 pounds shall then be used to slowly approach, and make contact with, the surfaces of the eye speculum. The magnet shall be fabricated from Alnico No. 5 alloy, and shall be proven capable of supporting 43 cubic inches of steel (3-1/2 inches x 3-1/2 inches x 3-1/2 inches) weighing 12 ± 1/2 pounds.

5. PREPARATION FOR DELIVERY.

5.1 Packaging.

5.1.1 Unit of issue. One speculum, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not required	48 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.3.2 Preservative. Not required.

5.1.3.3 Intimate wrap. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive material.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:57Z
Check the source to verify that this is the current version before use.

ith dimensions within toe tolerances

5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method 1A-15.

5.1.3.5 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2 inch wide tape conforming to PPP-T-76. Closing of bottom flaps with adhesive conforming to MMM-A-250 or metal stitches is permitted.

5.1.4 Level C. Unit shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:57Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Intended use. The eye speculums are to be used in eye surgery.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Applicable levels of packaging and packing (see 5.1 and 5.2).
(c) Federal stock number.

6.3 This specification is intended to cover Federal Stock Number 6515-369-4100 SPECULUM, EYE, Guyton-Park, Nonmagnetic, 3-1/2 inch, appearing in the Federal Supply Catalog.

6.4 Supersession data. This specification includes requirements of Purchase Description No. 5, dated 1 November 1968.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1291

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:57Z
Check the source to verify that this is the currentversion before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:57Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements.

SPECIFICATION

MIL-S-36649 FSN:

ORGANIZATION

CITY AND STATE | CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)* | DATE

DD FORM 1 JAN 68 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:57Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:57Z
Check the source to verify that this is the current version before use.