MIL-T-36655
16 April 1970

MILITARY SPECIFICATION

TUBE, DUODENAL, SURGICAL, PLASTIC, RADIOPAQUE, DISPOSABLE, MILLER-ABBOTT

This specification is mandatory for use by all the Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers one type of a sterile, pyrogen-free, non-toxic, disposable plastic, duodenal tube, adult size of the Miller-Abbott design for aspiration of the small bowel.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

MMM-A-250	Adhesive, Water Resistant (for Sealing Fiberboard Boxes).
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, (for Carton Sealing).

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:58Z
Check the source to verify that this is the current version before use.

STANDARDS

Federal

FED-STD-406	Plastics: Methods of Testing.
FED-STD-601	Rubber: Sampling and Testing.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

23195	Tube, Duodenal, Miller-Abbott, 16 French, Adult Size

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply:

AMERICAN NATIONAL STANDARDS INSTITUTE (A.N.S.I.)

A.N.S.I. B70.1	American Standard Dimensions of Glass and Metal Luer Tapers for Medical Applications.

(Application for copies should be addressed to the American National Standards Institute, 1430 Broadway, New York, New York 10018)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:58Z
Check the source to verify that this is the current version before use.

THE UNITED STATES PHARMACOPEIAL CONVENTION, INC.
The Pharmacopeia of the United States of America (U.S.P.)

(Applications for copies shall be addressed to the Mack Publishing Company, Easton, Pennsylvania 18042).

FEDERAL, FOOD DRUG AND COSMETIC ACT AND REGULATIONS PROMULGATED, THEREUNDER

(Applications for copies should be addressed to Food and Drug Administration, Washington, D. C. 20204).

3. REQUIREMENTS

3.1.1 Tubing. All tubing shall be fabricated from medical-grade, pliant plastic of the best quality and suitable in all respects for the purpose intended. The plastic tubing shall be sterile, safe (non-toxic), pyrogen-free, and shall meet the requirements of the U.S.P. for class VI plastics with the exceptions specified in 4.4.2. The inflation tube member shall be radiopaque throughout its length. The aspirating tube shall be clear, transparent material.

3.1.2 Inflation female Luer connector. The inflation female Luer connector shall be fabricated from a suitable plastic that shall meet the requirements of the current edition of the U.S.P. for class II plastics, when tested as specified in 4.4.1.

3.1.3 Balloon. The balloon shall be fabricated from natural white latex rubber of the best quality, suitable for the purpose intended.

3.1.4 Aspirating tube connector and clamp. The aspirating tube connector and clamp shall be fabricated from plastic, suitable for the purpose intended.

3.1.5 Metal insert. The metal insert shall be fabricated from nickel plated or chromium plated brass, or from austenitic or martensitic corrosion-resistant steel.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:58Z
Check the source to verify that this is the current version before use.

3.2 Style, design and dimensions. Shall be a sterile, pyrogen-free, non-toxic, plastic, disposable, Miller-Abbott duodenal bi-lumenal tube. The tube shall be adult size, 8 feet (plus or minus 3 inches) in overall length, and suitable for small bowel aspiration. Shall consist essentially of a bilumenal tube with an externally inflatable balloon tip at its most distal end. The aspirating tube shall be combined with an integral smaller inflation tube. The inflation tube shall be formed over the outside diameter of the aspirating tube, and shall terminate at the balloon end within a metal insert. The aspirating tube shall be formed over the metal insert and shall seal the outside diameter of the inflation tube, leaving the inner diameter of the inflation tube unrestricted for inflation purposes. The aspirating tube shall have markings as shown on drawing with ink suitable for use in ingestion procedures. The aspirating tube shall have an integral aspirating open end connector at its most proximal end. The inflation tube shall have an integral female Luer connector at its most proximal end that shall make a leakproof fit with a standard male Luer taper as specified herein. A clamp shall be provided to seal the balloon lumen at the proximal end of the inflation tube. The bilumenal tube shall be designed so that the balloon is invaginated over the tip of the tube to make easy, comfortable, transnasal passage. Style, design and dimensions of the complete tube assembly shall be in accordance with DPSC Drawing 23195.

3.2.1 Diameter and depth of female Luer cavity. The female Luer cavity of the inflation tube connector shall be tested for diameter and depth of Luer cavity as specified in 4.4.7. Luer connectors which do not fit plug gage within the "GO" - "NO GO" limits specified in 4.4.7, shall be rejected. In addition, when the Luer connectors are fitted onto the plug gage as specified in 4.4.7, there shall be no movement of the connectors on the gage when the connectors are pressed lightly from side to side in all directions.

3.3 Sterility, safety, and freedom from pyrogens. All equipment shall be sterile, safe, and free from pyrogens when tested as specified in 4.4.4, 4.4.5, and 4.4.6, respectively.

3.4 Animal implantation test. The tubing shall show no evidence of toxicity or adverse tissue reaction upon gross and microscopic examination when tested as specified in 4.4.3.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:58Z
Check the source to verify that this is the current version before use.

3.5 Physical requirements of finished plastic tubing. The following minimum physical requirements shall apply for the finished plastic tubing when tested as specified in 4.4.10:

Identation hardness:	70 ± 5 durometer
Minimum breaking strength:	20 pounds
Minimum elongation:	200 percent

3.6 Accelerated aging and breaking strength of balloon. The natural latex rubber balloon shall not become brittle or tacky when tested as specified in 4.4.11 and shall not decrease in breaking strength and ultimate elongation by more than 25 percent when tested as specified in 4.4.12. Straight specimens shall be used for these tests.

3.7 Security of assembly. When tested as specified in 4.4.13, the inflation female Luer connector and the aspirating tube connector shall not break, distort, loosen, or pull out of their respective tubing members, nor shall each tubing exhibit elongation that results in tubing failure.

3.8 Fluid tightness of assembly. When tested as specified in 4.4.8, the tube assembly shall show no leakage. In addition, the Luer connection of the inflation tube connector with the taper test gage shall not show leakage exceeding the formation of one clinging drop when tested as specified in 4.4.8.

3.9 Balloon strength. When tested as specified in 4.4.9 the balloon shall not show evidence of rupture or leakage.

3.10 X-ray opacity. The inflation tube shall appear prominent and conspicuous when covered by at least 5/8 inch of aluminum when tested as specified in 4.4.14.

3.11 Certificates of sterility, freedom from pyrogens, safety, biological tests and animal implantation tests. A certificate listing all lot or control numbers shall state that representative samples from each lot or control have been tested and found to be sterile, free from pyrogens, nontoxic (safe), and have passed the biological tests and the animal implantation test. This certificate shall include the specific tests performed and the tests results. The record of the animal implantation test and its results shall be certified by the pathologist who performed or supervised performance of the animal implantation test. For destination inspection, certificates shall be submitted by the contractor directly to Defense Personnel Support Center, Directorate of Medical Materiel, Technical Operations Division, 2800 South 20th Street, Philadelphia, Pa. 19101, attention: Quality Control Branch, and one copy of each shall be forwarded with each shipment to the consignee. For source inspection, certificates shall be furnished to the procuring activity via the government inspector.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:58Z
Check the source to verify that this is the current version before use.

Approved for public release; distribution

3.12 Manufacturer's record. The manufacturer shall place a code or lot number on the paper label, or on the printed carton, which can be verified against the manufacturer's records of all sterility, freedom from pyrogens, safety, biological, and animal implantation tests performed on the tubes. These records shall be made available for inspection by any duly authorized Government representative. No shipment shall be made until all tests have been completed, verified, and recorded in the permanent records of the manufacturer.

3.13 Compatibility. The materials of which the tube assemblies are manufactured shall not cause any deleterious effects upon the body when used for the purpose intended.

3.14 Instructions. A detailed instruction sheet for proper use of the duodenal tubes shall be provided with each duodenal tube.

3.15 Workmanship. Workmanship shall be first class throughout. Tubes shall be free from defects which detract from their appearance or impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. Certificates of quality supplied by the manufacturer of the plastics, rubber, and corrosion-resistant metals may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, the individual test results, and the specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:58Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as indicated in Table I. Unit of product shall be one tube assembly.

Table I. Sampling Criteria for Examination

Type Defect	Inspection Level*	AQL (percent defective)
Major	S-4	1.0
Minor	S-2	2.5

* If acceptance number is not exceeded, lot shall be accepted. If rejection number is equalled or exceeded, then the inspection at level II may be performed to determine the final rejection or acceptance.

4.2.2 For tests.

4.2.2.1 For sterility test. Sampling for the sterility test shall be in accordance with the Page Proof to the United States Pharmacopeia, Eighteenth Revision (U.S.P. XVIII). Upon publication of the U.S.P. XVIII, sampling for sterility testing shall be as indicated therein for parenteral devices.

4.2.2.2 For pyrogen and safety tests. Sampling for pyrogen and safety tests shall be as indicated in the current edition of the U.S.P. for Transfusion and Infusion Assemblies.

4.2.2.3 For other tests. Sampling for other tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics. The unit of product for test purposes shall be one tube assembly.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:58Z
Check the source to verify that this is the current version before use.

Table II. Sampling Criteria for Tests

Component	Characteristic	Requirement	Test Procedure	Insp. Level	AQL
Tube Assembly	Dimensions	3.2	- - -	S-1	*
	Security of assembly	3.7	4.4.13	S-2	*
	Fluid tightness	3.8	4.4.8	S-2	*
Tubing	Material-Class VI	3.1.1	4.4.2	**	*
	Animal implantation	3.4	4.4.3	**	*
	Physical properties	3.5	4.4.10	S-2	*
Inflation Tube	X-ray opacity	3.10	4.4.14	S-1	*
Inflation Female Luer Connector	Material-Class II	3.1.2	4.4.1	**	*
	Diameter and depth of Luer cavity	3.2.1	4.4.7	S-2	*
Balloon	Accelerated aging	3.6	4.4.11	S-2	*
	Breaking strength	3.6	4.4.12	S-2	*
	Balloon strength	3.9	4.4.9	S-2	*

* Acceptance number shall be zero for all sample sizes.

** Sufficient samples to perform the biological tests shall be selected from each lot.

4.3 Examination. The tube assemblies shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted as specified in Tables I and III. Inspection, however, shall not be restricted to defects listed in Table III. Any tube assembly in the sample containing one or more defects shall be rejected and if the number of defective items in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

Source: https://assist.dla.mil -- Downloaded: 20261O-21T08:58Z
Check the source to verify that this is the current version before use.

Table III. Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Aspirating tube does not have markings (in centimeters) as shown on drawing.
102	Outside surfaces of inflation and aspirating tubes not smooth and free from holes, pits or surface imperfections.
103	Balloon not free from holes or tears.
104	Inflation tube not integral to, or formed to the aspirating tube.
105	Tubing clamp missing.
106	Holes in aspirating tube, at the balloon end, not free from burrs, slivers and flash.
Minor	
201	Instruction sheet for proper use of tube not furnished.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Biological tests for class II plastics. The inflation female Luer connector shall be tested in accordance with the requirements of the U.S.P. for class II plastics.

4.4.2 Biological tests for class VI plastics. All tubing shall be tested in accordance with requirements of the U.S.P. for class VI plastics.

4.4.3 Animal implantation test. Samples of the finished tubes shall be tested in accordance with the current edition of the U.S.P. for Implantation Test, Plastic Containers, for class VI plastics, except after "72 hours" add "and not to exceed 144 hours"; and delete 'Examine macroscopically the area of the tissue surrounding the center portion of each implant strip. Use a magnifying lens if necessary,' and substitute: "The muscle tissue surrounding the implants shall be examined, both grossly and microscopically, by a pathologist certified by the American Board of Pathology or the American College of Veterinary Pathologists'."

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:58Z
Check the source to verify that this is the current version before use.

4.4.4 Sterility test. Samples of the finished tubes shall be tested for sterility in accordance with the procedures for parenteral devices as described in the Page Proof of the U.S.P. XVIII. Upon publication of the U.S.P. XVIII, sterility testing shall be as indicated therein for parenteral devices.

4.4.5 Test for freedom from pyrogens. Samples of the finished tubes shall be tested for freedom from pyrogens in accordance with the U.S.P. for Transfusion and Infusion Assemblies.

4.4.6 Safety (toxicity) test. Samples of the finished tubes shall be tested for safety in accordance with the U.S.P. for Transfusion and Infusion Assemblies.

4.4.7 Diameter and depth of female Luer cavity. The Luer cavity of the inflation tube connector shall be tested with the male Luer taper gage of figure 5 of A.N.S.I. Standard Z/0.1 (this gage shall be as shown in the A.N.S.I. Standard, except that it shall be constructed with a suitable fluid path to permit performance of the fluid tightness test specified in 4.4.8). The Luer cavity shall be slipped onto this gage with very little pressure. The fit of Luer cavity on this gage shall be tested by means of the feeler gage in accordance with figures 5 and 6 of A.N.S.I. Z70.1.

4.4.8 Fluid tightness of assembly. With the tube assembly in a vertical position, the inflation tube Luer cavity connector shall be attached by means of a plug gage, as shown in figure 5 of A.N.S.I Standard Z70.1 (except that the gage shall have been provided with a fluid path), to a water pressure source having a pressure gage. The water shall be permitted to run through the inflation tube so as to fill the cavity space and the entire inflation tube member. A minimum pressure of 10 pounds per square inch gage (psig) shall be applied for 30 seconds. The tube assembly shall then be observed for leakage.

4.4.9 Balloon strength. With the tube assembly in a vertical position, a minimum volume of 30 cubic centimeters of an appropriate liquid shall be applied to the inflation tube member and allowed to settle into the rubber balloon. The balloon shall be observed for rupture and leakage.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:58Z
Check the source to verify that this is the current version before use.

4.4.10 Physical properties. The following physical properties, specified in 3.5, shall be tested in accordance with the following method numbers of Federal Test Method Standard 406:

Property	Method No. of Fed. Test Method Std. No. 406
Indentation hardness	1082
Minimum breaking strength	1011
Minimum elongation	1011

4.4.11 Accelerated aging test. The rubber balloon member of the tube assembly shall be subjected to method 7221 of Federal Test Method Standard No. 601.

4.4.12 Breaking strength. The breaking strength and ultimate elongation shall be conducted in accordance with methods 4111 and 4121 of Federal Test Method Standard No. 601. In using method 4111 to determine the breaking strength, the value defined as "breaking force" shall be used as the value for breaking strength.

4.4.13 Security of assembly. Suitable segments of the inflation and aspirating tubes shall be clamped and the following steady forces shall be applied for a period of not less than 1 minute to the inflation tube connector and the aspirating tube connector.

	Applied Force (Minimum)
Inflation tube:	1.5 pounds
Aspirating tube	5.0 pounds

4.4.14 X-ray opacity test. The inflation tube members shall be placed on an unexposed X-ray plate and covered, with an aluminum ladder ranging in thickness from at least 1/8 inch to 3/4 inch in steps of 1/8 inch. X-ray exposure shall be 10 MAS, 70 KVP at 30 inches, with automatic processing, using appropriate speed screens in a convenient size cassette, such that the steps of at least 1/4 inch, 3/8 inch, 1/2 inch, and 5/8 inch are clearly delineated from each other.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:58Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. Each tube, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging quantities

Unit package	Intermediate package	Exterior container
1 unit	12 units	96 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each tube shall be coiled of such a diameter as to prevent the tubing from being kinked or taking a permanent set and packaged in a combination polyethylene and paperback tear type envelope of appropriate size. The tear type envelope shall be further packaged in a 3 mil transparent, heat sealed, polyethylene envelope of appropriate size. The tear type envelope shall be capable of maintaining the sterility of the contents unless the envelope is damaged or opened. The outer polyethylene envelope shall be so constructed as to be waterproof.

5.1.3.2 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2 inch wide tape conforming to PPP-T-76. Closing of bottom flap with adhesive conforming to MMM-A-250 or metal stitches is permitted.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:58Z
Check the source to verify that this is the current version before use.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Unit package. Unit package markings shall be printed on the interior container in such a manner as to be ready identified through the outer container. Marking of the outer container is not required. Markings shall be in accordance with MIL-STD-129, the requirements of the Federal Food, Drug and Cosmetic Act and in addition shall include the sterility lot control number, name and address of the manufacturer, the date and method of sterilization (G-Gas, S-Steam, F-Filtrations, R-Radiation) shall be shown below the last line of the identification markings. Example: "STER-1/69-G". In addition, marking shall include the legend "STERILE" and "STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED".

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:58Z
Check the source to verify that this is the current version before use.

5.3.2 Intermediate package. Each intermediate package shall be marked in accordance with Military Standard MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Sterility lot (control) number, contract or purchase order number and name of contractor shall be shown. Date and method of sterilization as specified in 5.3.1 shall be shown.

5.3.3 Exterior containers. Exterior container shall be marked in accordance with MIL-STD-129. Sterility lot (control) number and date and method of sterilization as specified in 5.3.1 shall be shown.

NOTE: Sterility lot control number shall be traceable to each sterilized charge.

6. NOTES

6.1 Intended use. These Miller-Abbott plastic duodenal tubes are to be used for aspiration of the small bowel in adults.

6.2 Ordering data. Procurement documents should specify the following:

a. Title, number and date of this specification.
b. Applicable levels of packaging and packing.
c. Federal stock number.

6.3 This specification covers the following item appearing in the Federal Supply Catalog:

Federal Stock Number	Item Identification
6515-400-5468	TUBE, DUODENAL, SURGICAL Plastic, Radiopaque, Disposable, Miller-Abbott, 16 Fr

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1010

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:58Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements.

SPECIFICATION

MIL-T-36655 FSN:

ORGANIZATION

CITY AND STATE | CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)* | DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:58Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 1910'

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
MAR 68

EDITION AUG 66, WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:58Z
Check the source to verify that this is the current version before use.