MIL-T-0036673A (DM)
16 June 1986
USED IN LIEU OF
MIL-T-36673
3 June 1970

MILITARY SPECIFICATION

TUBE, EXTENSION, INJECTION SET

This military specification has been prepared by the Defense Personnel Support Center based upon currently available technical information but is has not been approved for promulation as a coordinationed revision of MIL-T-36673. It is subject to modification. However, pending its promulation as a coordinated military specification, it may be used in acquisition.

1. SCOPE

1.1 Scope. This specification covers three types of sterile, disposable, plastic, intravenous extension sets, suitable for the administration of intravenous medications and for introduction of diagnostic dyes into the circulatory system.

1.2 Classification. The extension tubes shall be of the following types as specified (see 6.2):

Type I - Thirty-inch, one clamp

Type II - Ten feet, one clamp

Type III - Thirty-two inch, two Y injection sites, two clamps

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-ATT, 2800 S. 20th Street, Philadelphia, PA 19101, by using self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:00Z
Check the source to verify that this is the currentversion before use.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications and standards. The following specifications and standards form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DoDISS) and supplement thereto, cited in the solicitation.

SPECIFICATIONS

FEDERAL

MMM-A-250	- Adhesive, Water-Resistant (for Sealing Fiberboard Boxes).
PPP-B-566	- Boxes, Folding, Paperboard.
PPP-B-601	- Boxes, Wood, Cleated-plywood.
PPP-B-621	- Boxes, Wood, Nailed and Lock-corner.
PPP-B-636	- Boxes, Shipping, Fiberboard.
PPP-B-676	- Boxes, Set-up.
PPP-R-76	- Tape, Pressure-Sensitive Adhesive Paper (for Carton Sealing).

STANDARDS

MILITARY

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.

(Copies of specifications and standards required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of the documents which are DoD adopted shall be those listed in the issue of the DoDISS specified in the solicitation. Unless otherwise specified, the issues of documents not listed in the DoDISS shall be the issue of the nongovernment documents which is current on the date of the solicitation.

AMERICAN NATIONAL STANDARDS INSTITUTE (ANSI)

ANSI/HIMA MD70.1 - LUER TAPER FITTINGS

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:00Z
Check the source to verify that this is the current version before use.

(Application for copies should be addressed to the Publication Sales Dept., STD, American Institute of Physics, 335 E. 45th Street, NY, NY 10017).

THE UNITED STATES PHARMACOPEIAL CONVENTION, INC.
The Pharmacopeia of the United States of America (USP).

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042).

(Nongovernment standards and other publications are normally available from the organizations which prepare or which distribute the documents. These documents also may be available in or through libraries or other informational services).

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence. Nothing in this specification, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

3.1 Material.

3.1.1 Rubber and plastic components. All rubber and plastic components shall be of the best quality and suitable in all respects for the intended purpose. All rubber and plastic components shall be sterile, free from pyrogens, compatible, and safe when tested as specified in section 4. Those rubber and plastic components coming in contact with solutions being administered, and end protectors, shall contain not more than 5 parts per million (p.p.m.) total of lead, barium and cadmium when tested as specified in 4.4.1, or 4.4.1.1. The plastics shall meet the requirements of the Pharmacopeia of the United States (USP) for Class II plastics and for transfusion and infusion assemblies when tested as specified in 4.4.2.

3.1.2 Plastic tubing. The transparent PVC tubing shall be flexible and pliable, but be sufficiently rigid to prevent collapse during normal use. The tubing shall be free from all liquid additive substances which oxidize, exude or sweat-out onto the surface of the tubing. The tubing shall have the physical properties as shown below when tested as directed in 4.4.10.

Type	Ultimate Strength	% Elongation	100% Modulus
I, III	2200 psi minimum	250 minimum	1,000 to 1,600 psi
II	2200 psi minimum	250 minimum	800 to 1,600 psi

3.1.3 Adapters. Adapters shall be fabricated from a rigid, tough, dimensionally stable plastic.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:00Z
Check the source to verify that this is the currentversion before use.

3.1.4 Flow regulator. The flow regulators (control clamps), whether of the roller or slide type, shall be fabricated from a rigid, tough plastic.

3.2 Style, design and dimensions. The units shall be disposable plastic tubing extension sets suitable for the administration of intravenous medications and for the introduction of diagnostic dyes into the circulatory system. The extension sets shall have permanently attached female Luer adapters at one end and male slip-fit Luer adapters at the other end. The adapters shall each be fitted with a tight but readily removable end protector which shall maintain the sterility of the fluid path and adapters after sterilization. The Types I and III extension sets shall have a slotted slide clamp or roller clamp present to control the flow rate from complete shut-off to full flow. The type II extension set shall have two Y injection sites present 6 to 7 inches from the Luer adapters. A slotted slide clamp shall be present between the female Luer adapter and the Y injection site and a slide or roller clamp between the two Y injection sites.

3.2.1 Tubing. Dimensions for the tubing shall be as follows:

Type	Overall Length	Internal Diameter	External Diameter
I	30 + 5, -0	0.100 + .010, - .005	0.138 + .010, - .005
II	120 + 10, -5	0.120 + .010, - .005	0.170 + .007, - .005
III	32 + 5, -0	0.100 + .010, - .005	0.138 + .010, - .005

3.2.2 Adapters. The adapters shall be permanently attached at the ends of the plastic tubing in a leakproof manner.

3.2.2.1 Male adapter. The male slip-fit adapter shall be approximately 1-1/2 inches in overall length. The taper shall meet the requirements of ANSI/HIMA Standard MD70.1 as directed in 4.4.6 and shall form a tight, leak-proof fit with a Luer taper hub.

3.2.2.2 Female adapter. Female adapter shall be 3/4 inch to 1-3/4 inches in overall length. The taper shall meet the requirements of ANSI/HIMA Standard MD70.1 as directed in 4.4.7 and shall form a tight, leak-proof fit with a male Luer taper.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:00Z
Check the source to verify that this is the currentversion before use.

3.2.4 Flow regulator. Flow regulator shall be a roller or slotted slide-type clamp which will not accidentally be removable from the connecting tube and from which the roller of the roller clamp cannot be accidentally removed. Slide clamps shall be rigid, smooth and free of burs. The flow regulator shall be positioned on the tubing in an open position. The regulator shall provide a constant setting and quick shutoff from the operating position. It shall be movable, and shall be capable of one-handed operation. To assure constant delivery of fluid, it shall be impossible for the clamp setting to be changed drastically by an accidental jerk on the tubing. When tested as specified in 4.4.9, the clamp shall be capable of accomplishing a complete quick shutoff, with no falling drop apparent in 5 minutes, and with a constant head of liquid, shall maintain a constant flow rate at any point between complete shutoff and maximum flow. In addition, while the clamp is in the same position on the tubing, the clamp shall be operated from its fully open position to complete shutoff ten times: the tubing shall not be cut, pierced, or damaged in any manner which would cause leakage or restriction of the flow of solution.

3.2.4.1 Security of roller in clamp body. When the flow regulator is of the roller type, the roller shall not be removed from the roller clamp under a 3.65 kilogram load when tested as specified in 4.4.9.1.

3.2.5 Injection sites. The injection site shall be a rigid, transparent "Y" with a puncturable self-sealing, rubber, diaphragm. The diaphragm shall be permanently attached and the branch of the Y to which it is attached shall face the female adapter. The diaphragm shall be so made and secured that after puncture and withdrawal of the needle in accordance with 4.4.8 there shall be no separation or leakage.

3.3 Compatibility. The materials of which the sets are manufactured shall not react with, or be acted upon by, the liquids with which they will be in contact during normal use and operation, nor shall there by any materials incorporated in the set which will cause any harmful or deleterious effect when the set is used for the purpose intended.

3.4 Leakage and separation of set. The adapters shall be permanently attached to the plastic tubing in such manner that, with the opening blocked, it shall not leak or separate when subjected to an internal hydrostatic pressure of 45 pounds per square inch (psig) for 5 minutes. The sets shall be tested for leakage and separation of set as specified in 4.4.1.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:00Z
Check the source to verify that this is the current version before use.

3.5 Flow rate. Flow rate for Type I and III sets shall be 100 ml of water in no more than 120 seconds when tested in accordance with 4.4.9. The flow rate for the type II set shall be 1 liter in less than 16 minutes when tested in accordance with 4.4.9.

3.6 Security of assembly. The adapters shall not separate or loosen from the plastic tubing. The union shall withstand a straight pull of 7 pounds when tested as specified in 4.4.2.

3.7 Manufacturer's records of tests. The manufacturer shall place a code or lot number on the paper label, or on the printed carton, which can be verified against the manufacturer's records of all tests made on nontoxic, sterile, pyrogen-free sets. These records shall be made available for inspection by any duly authorized Government representative. No shipment shall be made until all tests have been completed, verified, and recorded in the permanent records of the manufacturer.

3.8 Sterility, safety, and freedom from pyrogens. The extension tube shall be pyrogen-free when tested as specified in 4.4.3 and sterile when tested as specified in 4.4.5. The sets shall be nontoxic (safe) when tested for safety as specified in 4.4.4.

3.9 Certificates of sterility, freedom from pyrogens, and safety tests. A certificate listing all lot or control numbers shall state that representative samples from each lot or control have been tested and found to be sterile, free from pyrogens, and safe. This certificate shall include the specific tests performed and the test results. For source inspection, certificates shall be furnished to the procuring activity via the government inspector. For destination inspection, certificates shall be submitted by the contractor directly to Defense Personnel Support Center, Directorate of Medical Materiel, Technical Operations Division, 2800 South 20th Street, Philadelphia, Pa. 19101, attention: Quality Assurance Branch, and one copy of each shall be forwarded with each shipment to the consignee.

3.10 Instructions. Instructions for proper use of the sets shall be provided with each unit package.

3.11 Workmanship. Workmanship shall be first class throughout. The sets shall be free from defects which detract from their appearance or impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:00Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in Table I.

TABLE I. Sampling for Examination

Type of defect	Inspection level	AQL (percent defective)*
Major	I	1.0
Minor	I	1.5

* If acceptance number is not exceeded, lot shall be acceptable. If rejection number is equalled or exceeded, inspection at level II may be performed to determine final acceptance or rejection.

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in table II. Where possible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:00Z
Check the source to verify that this is the current version before use.

TABLE II. Sampling for Tests

Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Heavy metals	3.1.1	4.4.11, 4.4.11.1	S-1	*
Tubing	3.1.2, 3.2.1	4.4.10	S-1	*
Male adapter	3.2.2.1	4.4.6	S-2	*
Female adapter	3.2.2.2	4.4.7	S-2	*
Flow regulator	3.2.4	4.4.9	S-2	*
Security of roller in clamp	3.2.4.1	4.4.9.1	S-3	*
Injection Sites	3.2.4	4.4.8	S-2	*
Leakage and separation of set)	3.4	4.4.1	S-3	*
Flow rate	3.5	4.4.8	S-2	*
Security of assembly	3.6	4.4.2	S-3	*
Freedom from pyrogens	3.8	4.4.3	S-2	*
Safety (non-toxicity)	3.8	4.4.4	S-2	*
Sterility	3.8	4.4.5	4.4.5	*

*Acceptance number shall be zero for all sample sizes.

4.3 Examination. The connecting tube shall be examined to determine compliance with specification requirements. Examination shall be conducted in accordance with MIL-STD-105 and the following classification of defects. Examination shall include but not be limited to the defects listed in Table III.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:00Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects

Categories	Defects
Major	
101	Tubing kinked.
102	Crack in an adapter.
103	Flow regulator is not easily controlled with one hand.
104	Poor design, fittings not provided with finger grip area.
105	Presence of foreign materials in extension tube.
106	Set incomplete.
107	Burrs on adapters.
Minor	
201	Burrs on slide clamp.
202	Instructions missing, incomplete or illegible.
203	Tubing not clear and free of discoloration.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Leakage and separation of set. Samples of sterile connecting tubes shall be subjected to an internal hydrostatic pressure, gradually applied, of 45 pounds per square inch gage (psig) and immersed in water for five minutes. The pressure at 4 psig, on the way up to 45 psig, shall be held for one minute. A stream of air bubbles constitutes failure. An air bubble which forms but does not break in the one minute is acceptable. This test shall be performed with a gage shown in Figure 8 of ANSI/HIMA MD70.1 (and having a suitable fluid path throughout the length of the gage) fitted to the Luer tapered tip of the male adapter.

4.4.2 Security of assembly test. The tubing of the set shall be severed at the midpoint. The security of assembly of each adapter shall then be tested by gradually applying a seven-pound weight to the severed end of the tubing while the adapter is held in a clamping device so that the tubing hangs vertically downward. This load shall be maintained for 1 minute without evidence of separation of the adapters from the tubing.

4.4.3 Freedom from pyrogens. Extension tubes shall be tested for freedom from pyrogens in accordance with the USP for Transfusion and Infusion Assemblies.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:00Z
Check the source to verify that this is the current version before use.

4.4.4 Safety (non-toxicity). Extension tubes shall be tested for safety in accordance with the USP for Transfusion and Infusion Assemblies.

4.4.5 Sterility test. The extension tube shall be tested for sterility in accordance with the method specified in the USP using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates it from product intended for distribution. The incubation medium for the paper spore strip (carrier) shall be soybean-casein digest. The incubation media for the inoculated product, or inoculated simulated product shall be fluid thioglycollate or alternative thioglycollate at 30^0 to 35^0 C for one half of the specimens and soybean-casein digest at 20^0 to 25^0 C for the remaining half of the specimens. The minimum incubation time shall be ten days. Sampling shall be as follows:

Medium	Paper Spore Strip	Inoc. Prod. or Inoc. Simulated Prod.
Soybean-casein digest	S-3*	S-3*
Fluid thioglycollate or Alternative thioglycollate		S-3*

*Use a minimum of 20 specimens. Acceptance number is zero.

4.4.5.1 Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of sterilization	Organism
Steam	B. stearothermophilus
Dry heat	B. stearothermophilus
	B. subtilis
Ethylene oxide	B. subtilis
Gamma radiation	B. pumilus

The number of viable spores shall be sufficient to provide assurance that the product has been sterilized.

4.4.5.2 Inoculated carrier. Inoculated carriers (paper spore strips) if used, shall be in accordance with MIL-S-36586A, dated 18 August 1976, Sterilization Test Strip Set, Bacterial Spore. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.5. Sterility test and 4.4.5.1. Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in 4.4.5.1.
NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP and paragraphs 4.4.5 and 4.4.5.1.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:00Z
Check the source to verify that this is the currentversion before use.

4.4.5.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

4.4.5.4 Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP. The inspection level shall be as stated in paragraph 4.4.5 above.

4.4.6 Taper of male Luer adapter. The male Luer adapter shall be tested with the ring gage shown in Figure 4 of ANSI/HIMA MD70.1 Standard. Tips that do not fit the gage shall be considered not acceptable.

4.4.7 Diameter and depth of cavity of female Luer adapter. The cavity of the female Luer adapter shall be tested on the male Luer taper gage in accordance with the plug gage as shown in figure 5 of ANSI/HIMA MD70.1 Standard.

4.4.8 Reseal test. The rubber diaphragm shall reseal without a falling drop within 10 seconds after puncture with a nonlubricated 18 gage needle when subjected to an internal pressure (hydrostatic pressure) of 4 psig for 30 seconds.

4.4.9 Flow regulator effectiveness. The types I and III extension tubes shall be connected to a regular administration set with 54 inch tubing set up with fluid (H20) flow from a solution bottle at a height of 5 feet. Flow regulator effectiveness on tubing shall be observed by varying flow rates from fully opened to fully closed. The type II extension set shall be tested for fluid flow by connecting it to a vented, one-liter bottle of water which shall be held three feet above the male adapter and the time necessary to empty the bottle determined.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:00Z
Check the source to verify that this is the current version before use.

4.4.9.1. Security of roller in clamp body. The roller clamp shall be removed from the set together with a piece of tubing approximately one inch longer than the clamp body. The tubing shall be present in the roller clamp in its normal position with a 1/2 inch length extending from each end of the roller clamp. The clamp shall be held vertically, with the large end down, over a supporting gap so that the roller if forced out of the body is free to fall. A polished steel rod is applied to the top of the roller, so that the roller is able to rotate in the course of the test, and the vertical downward force required to separate the roller from the clamp body is measured. As an alternate, any position of the clamp is acceptable providing the clamp body is not distorted and the load measurement is accurate.

4.4.10 Tubing physical properties. The physical property requirements of 3.1.2 shall be obtained with the following or equivalent equipment: Instron TM; CT load cell; G-61-36 pneumatic smooth face grips or G-61-5C elastomeric grips.

4.4.11 Heavy metals extractable test. A 2-gram composite sample shall be obtained from the plastic components. The material shall be cut into small slivers, so that no thickness exceeds 3 mm., and eluted for 72 hours at 50° C, 24 hours at 70° C, 60 minutes at 121° C in one menstruum of 0.9% sodium chloride. Each gram of the material shall be covered with 5 milliliters (ml) of the menstruum. The menstrua are collected, agitated after elution, and then submitted to spectographic atomic absorption, spectrophotometeric or specific wet analysis for lead, barium, and cadmium. The metals are reported in parts per million. Thus 1.0 ppm in elute is equivalent to extracting 1.0 mg from 200 g of test material dispersed in 1,000 ml of elution menstruum.

4.4.11.1 Alternate heavy metals extractable test. Accurately weight a 5-gram sample for each of the metals (barium, cadmium or lead) into a 1-oz. Vycor crucible. Reduce to ash in a muffle furnace at 650° C for eight hours. Let cool to room temperature and add 15 ml of 1:1 hydrochloric acid. Bring to a boil on a hot plate, allow to boil for 2 to 3 minutes, remove, let cool, and filter through Whatman #2 filter paper into a 50-ml volumetric flask. Rinse the residue on the filter with deionised water and dilute to 50 ml. Prepare standard solutions of 1,2 and 5 ppm for each of the metals, namely barium, cadmium, and lead. Determine the absorbances of these solutions at appropriate wave lengths using a Hollow Cathode Ba, Cd, or Pb lamp for each of the respective metals and a 10X scale expansion on a spectrophotometer, such as the Perkin-Elmer Model 303, Atomic Absorption Spectrophotometer. Determine the absorbance of the sample solutions using the same conditions as with the standards.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:00Z
Check the source to verify that this is the current version before use.

Calculation:

$$\frac{A \times F \times \text{dilution}}{W} = \text{ppm barium, cadmium, or lead}$$

A = Absorbance

F = Factor = $\frac{\text{(Concentration standard in ppm Ba, Cd, or Pb)*}}{\text{absorbance}}$

W = Sample weight in grams

The three standards are average to give ppp/unit A.

4.12 Inspection of Packaging. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Individual container. Each extension tube shall be coiled in such a manner as to prevent kinking. The coiled tube shall be further packaged sterile in sealed peel-open type container capable of maintaining sterility of contents.

5.1.1.2 Unit package. Twenty individually packaged extension tubes shall be packaged in a snug fitting box conforming to PPP-B-566 or PPP-B-676. Closure of the box shall be as specified in the applicable specification.

5.1.2 Level C. The level C preservation for extension tubes shall conform to the MIL-STD-794 requirements for this level.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Six unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure, strapping and waterproofing shall be as specified in the box specification.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:00Z
Check the source to verify that this is the current version before use.

5.2.2 Level B. Six unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636 class domestic. Closure shall be as specified in the box specification.

5.2.3 Level C. The level C packing for the extension tubes shall conform to the MIL-STD-794 requirements for this level.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in a exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination equal 40 cubic feet or more. Loads shall be unitized on pallets. Quantities less than 40 cubic feet need not be unitized. Unitized loads shall be uniform in size and quantities to the greatest extent practicable.

5.2.5.1 Levels A and B. Extension tubes packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets conforming to MIL-STD-147.

5.2.5.2 Level C. Extension tubes packed as specified in 5.2.3 shall be unitized on pallets as specified in MIL-STD-794.

5.3 Marking.

5.3.1 Levels A, B and C.

5.3.1.1 Individual container. Each individual container shall be marked in accordance with the requirements of the Federal Food, Drug and Cosmetic Act and in addition the marking shall include the item identification, lot number and legend: "STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED" or similar statement.

5.3.1.2 Unit package, exterior container and unitized load. Each unit package, exterior container and unitized load shall be marked as specified in MIL-STD-129. In addition, marking shall include the lot number.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

Source: https://assist.dla.mil -- Downloaded: 2026l0-21T09:00Z
Check the source to verify that this is the current version before use.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.4.

6. NOTES

6.1 Intended use. The injection set is intended for use in the administration and introduction of medication and diagnostic dyes into the circulatory system.

6.2 Ordering data.

6.2.1 Acquisition requirements. Acquisition documents should specify the following:

(a) Title, number, and date of this specification.
(b) Classification (See 1.2)
(c) Applicable levels of packaging and packing (see 5.1 and 5.2).
(d) National Stock Number.

6.3 National Stock Number (NSN) coverage. This specification includes the following items appearing in the Federal Supply Catalog:

NSN	ITEM IDENTIFICATION
6515-00-782-6516	TUBE, EXTENSION, INJECTION SET, Plastic, Disposable, Clamp, 30-Inch, 20s (Type I)
6515-00-926-2153	TUBE, EXTENSION, INJECTION SET, Plastic, Disposable, Clamp, 10-Ft. (Type II)
6515-00-959-8116	TUBE, EXTENSION, INJECTION SET, Plastic, Disposable, Two Clamps, Two Y Sites, 32-Inch, 20s (Type III)

6.4 Subject term (key word) listing.

Extension
Injection set
Tube

6.5 Changes from previous issue. Asterisks are not used in this revision to identify changes with respect to the previous issue due to the extensiveness of the changes.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:00Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Review Activities:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Project No. 6515-3415

U.S. GOVERNMENT PRINTING OFFICE: 1986 - 605-036/44378

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:00Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
MIL-T-0036673A(DM)	TUBE, EXTENSION, INJECTION SET

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:00Z
Check the source to verify that this is the current version before use.