METRIC

A-A-54725
10 August 1992

COMMERCIAL ITEM DESCRIPTION

TEST TUBE, HEMATOCRIT DETERMINATION, WINTROBE, DISPOSABLE

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-T-36153A which is cancelled.

This Commercial Item Description covers a disposable Wintrobe test tube suitable for use in military medical treatment facilities.

Salient characteristics:

Material. The test tube shall be fabricated from Type I, Class A or B or Type II glass in accordance with ASTM E438.

Construction. Tube shall be transparent and shall have a uniform bore. Inside surface of the bottom shall be flat, and outside surface of bottom shall be round.

Dimensions. Dimensions shall be as shown:

Outside diameter (mm)	5.00	±	.25
Inside diameter (mm)	3.40	±	.50
Length (mm)	113.5	±	3.5

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Scale. Tube shall be marked with a scale which is at least 10 cm long. Scale shall start at the flat inner bottom and shall be marked with at least 100 evenly spaced lines. Every tenth line shall be numbered consecutively with Arabic numbers; numbers shall go in two directions. One set of numbers shall start 10 lines above the bottom and shall go upward from 1 to 10. The other set of numbers shall start 100 lines above the bottom and shall go down from 0 to 9. Numbered lines shall be longer than other lines and numbering of bottom line is optional. Additional lines may be marked above the scale. These lines shall conform to the characteristics of the graduation lines on the scale. Tolerance on overall length of scale shall be ± 1.5 mm. Overall length of scale shall be computed by adding theoretical distance for additional lines to 100 mm.

Graduation marking. Graduation lines shall be of uniform width (maximum of .26 mm). Graduation lines shall be perpendicular to the vertical axis of the tube and shall be parallel to each other. Graduation marking shall be green or white, functionally legible, and sufficiently durable to withstand normal procedures. When the tube is immersed into water for 30 seconds and a soft paper tissue used to wipe graduated portion of the wet tube lightly with 10 complete up and down strokes, there shall be no change in marking.

Performance. The tube shall be tested for resistance to centrifugal force.

Workmanship. The test tube shall be free from sharp edges, cracks and breaks, chipping, protrusions or any other defects which detracts from its appearance or impair its serviceability.

Unit. Package (PG). One package containing five hundred tubes, as specified, constitutes one unit. Each tube shall be packaged in a manner to permit withdrawal of one tube at a time, leaving the remaining tubes secure and protected from dust and dirt.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-2. The unit of product for examination purposes shall be one test tube.

Sampling for tests. Sampling for tests (dimensional) shall be conducted in accordance with MIL-STD-105 with an AQL of 2.5 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one test tube.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Resistance to centrifugal force. Tube shall be filled with water to the top graduation and placed in a centrifuge. Speed and dimensions of centrifuge head shall be such that the inside bottom of the tube is subjected to a relative centrifugal force of not less than 2500 gravities. Formula for calculating relative centrifugal force (RCF) is as follows:

$RCF = .00001118 \times r \times n^2$.
RCF = relative centrigual force (gravities).
r = rotating radius to bottom of tube (centimeters).
n = rotating speed (revolutions per minute).

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor Certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	Item Identification
6640-00-982-5979	TEST TUBE, HEMATOCRIT DETERMINATION, WINTROBE, DISPOSABLE.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6640-1758

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER	2 DOCUMENT DATE (YYMMDD)
	A-A-54725	10 AUGUST 1992

3 DOCUMENT TITLE
TEST TUBE, HEMATOCRIT DETERMINATION, WINTROBE, DISPOSABLE

4 NATURE OF CHANGE (*Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.*)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

NAME	b TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON
c ADDRESS (Include Zip Code)	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 *Previous editions are obsolete* 198/29

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.