INCH-POUND

A-A-54732
30 September 1992

COMMERCIAL ITEM DESCRIPTION

FORCEPS, DRESSING (Dental)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specification GG-F-599B which is cancelled.

This commercial item description covers four types of dental dressing forceps.

Salient characteristics:

Classification. The forceps shall be of the following types:

Type I	-	Meriam
Type II	-	No. 17
Type III	-	Amalgam Reinforcement Pin
Type IV	-	Endodontic

Material. Material of the forceps shall be fabricated from Class 4 martensitic corrosion resistant steel or Class 6 ferritic corrosion steel Type 430. As an alternate, Type III forceps shall be fabricated from Class 3 austenitic corrosion resistant steel Type 302, 303 or 304. The material shall be in accordance with ASTM Designation F899.

Hardness. Forceps manufactured from martensitic corrosion resistant steel shall have a Rockwell Hardness C Scale (HRC) of 40 to 48. Forceps manufactured from ferritic corrosion resistant steel shall have a Rockwell Hardness C Scale of 18 to 24. Forceps manufactured from austenitic corrosion resistant steel shall have a Rockwell Hardness B Scale (HRB) of 85 minimum. The forceps shall be tested for hardness in accordance with ASTM Designations E18 or E92 and E140.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Style, design and dimensions. The Type I meriam forceps shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No.20216. The Type II No. 17 forceps shall be in accordance with DPSC Drawing No. 20216. The Type III amalgam reinforcement pin forceps shall be in accordance with DPSC Drawing No. 24116. The Type IV endodontic forceps shall be in accordance with DPSC Drawing No. 23322. The forceps tips shall mate precisely with no overlap in either the axial or lateral view. All teeth shall mate uniformly and completely.

Finish. All forceps shall have a bright or satin finish. The forceps shall be uniform and free from cracks, pits, rough areas, crevices and grind or tool marks.

Passivation. The forceps shall be properly passivated.

Deflection. The operative ends of the forceps shall coincide when tested as specified herein.

Corrosion resistance. The forceps shall show no evidence of corrosion when tested in accordance with ASTM Designation F1089.

Identification marking and labeling. The forceps shall be permanently and legibly marked with the contractor's name or registered trademark and the word "stainless". Foreign manufactured forceps shall bear the name of the country of origin.

Workmanship. The forceps shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One forceps, of the type specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspection are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality supplied by the manufacturer of the materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one forceps. Sampling shall be inspection level S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Deflection test. The forceps shall be placed on its side (outside face of one member) on a level table and a downward force of 4 to 12 ounces applied at 3-5/8 to 3-3/4 inches from the proximal (joined) end.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers and their respective Item Identifications are covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6520-00-299-9671	FORCEPS, DRESSING, Meriam, 6 Inch
6520-00-542-7000	FORCEPS, DRESSING, No. 17, 6-1/4 Inch
6520-00-935-1092	FORCEPS, DRESSING, Restorative Retention Pin, 6-5/16 Inch
6520-01-070-1498	FORCEPS, ENDODONTIC, 6 Inch

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6520-2513
Location: c:\word\doc\levin\forcepcd

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54732	2. DOCUMENT DATE (YYMMDD): 30 September 1992

3. DOCUMENT TITLE: FORCEPS, DRESSING (Dental)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)
Defense Personnel Support Center ATTN: DPSC-MSE	(1) Commercial: 215-737-2870 (2) AUTOVON: 444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.