INCH-POUND

A-A-54736
23 November 1992

COMMERCIAL ITEM DESCRIPTION

THERMOMETER, CLINICAL, HUMAN

(Electronic, Battery Powered, With Battery Charger)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers requirements for an electronic thermometer suitable for oral or rectal body temperature measurement in humans.

Salient characteristics:

The thermometer shall be a portable, battery powered, Fahrenheit/Celsius electronic clinical thermometer system consisting of a thermometer, battery charger/storage base, oral probe, rectal probe, disposable probe covers, and dispenser probe cover brackets. The battery charger shall operate from 110/220V, 50/60 Hz power.

The battery charger shall be in compliance with UL 544, Standard for Medical and Dental Equipment.

The thermometer shall have a predictive mode temperature range of 90.0°F to 107.9°F, plus or minus 0.2°F.

The thermometer shall have monitor mode temperature range of 88.0°F to 107.9°F, plus or minus 0.2°F.

The thermometer shall have a pulse mode range of 20 to 200 beats per minute (BPM). Pulse mode activation shall be by depressing a button on the bottom of the thermometer.

The thermometer shall have an LED charge light indicator on the display panel when the battery is recharging.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

The thermometer shall have an audible tone that sounds when temperature has been computed, when pulse has been calculated, and when an error message has occurred.

The thermometer shall have a display panel that indicates temperature readings, tissue contact, pulse rate, and error messages on a five character minimum LED display.

A rechargeable nickel-cadmium battery, that automatically recharges when placed in the battery charger/storage base shall be supplied with the thermometer. The battery shall recharge in no more than 10 hours. Battery life shall be at least five hours continuous from fully charged state.

The thermometer shall have the capability of both Celsius and Fahrenheit readings.

The thermometer shall be capable of being hung around the neck for convenience during use.

The thermometer shall have a compartment for holding a box of disposable probe covers.

The thermometer shall include one oral probe with a spring action cord approximately 2 feet long when coiled, approximately 4 feet long when stretched. It shall be capable of attachment to and detachment from the thermometer with ease.

One rectal probe with a spring action cord approximately 2 feet long when coiled, approximately 4 feet long when stretched shall also be included with the thermometer. It shall be capable of attachment to and detachment from the thermometer with ease.

The thermometer shall be provided with 1000 clean, disposable, rigid plastic probe covers. The covers shall be 6 inches long and shall be fitted onto and removed from each probe without having to touch the probe cover.

The thermometer shall be provided with a probe cover dispenser bracket which may be metal or plastic. If metal, it shall be corrosion-resisting steel, properly hardened and passivated.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:37Z
Check the source to verify that this is the currentversion before use.

Workmanship. The electronic thermometer shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One electronic thermometer, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one thermometer.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one thermometer.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:37Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to, the following:

Accuracy test. Set the thermometer in the Monitor mode and set the Fahrenheit (^{o}F) = Celsius (^{o}C) to ^{o}F or ^{o}C as desired. Install probe cover on the probe. Insert probe tip into a circulating water bath that has a known temperature between 88°F and 107.9^{o}F and is accurate to within plus or minus 0.03^{o}F. After 3 to 5 minutes, the displayed temperature should completely stop changing and shall be equal to known bath temperature to within plus or minus 0.2^{o}F.

Performance test. The thermometer shall be subjected to appropriate tests to determine that the thermometer conforms to all performance requirements of this specification.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

6515-01-187-6606

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:37Z
Check the source to verify that this is the current version before use.

A-A-54736

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA - OSS

Project Number: 6515-5343

Location: THERM17C.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54736	23 November 1992

3. DOCUMENT TITLE: THERMOMETER, CLINICAL, HUMAN (Electronic, Battery Powered, With Battery Charger)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial: — (2) AUTOVON: (If applicable)
	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340, AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 98/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.