INCH-POUND

A-A-54737
16 October 1992

COMMERCIAL ITEM DESCRIPTION

CATHETER AND NEEDLE UNIT, INTRAVENOUS, 16 GAUGE, RADIOPAQUE, DISPOSABLE (Catheter into Needle)

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a radiopaque catheter and needle unit suitable for administering intravenous medication from a single venipuncture.

Salient characteristics:

General. The catheter and needle unit shall consist of an 8 inch, 16 gauge catheter with a Luer-lock hub in a transparent sleeve, a 14 gauge by 2 inch needle, a wire stylet, a needle protector and a flow control plug.

Catheter. The catheter shall be fabricated from clear polyvinyl chloride which meets the requirements of the United States Pharmaceutical (USP) for Class IV plastics (See NOTE 2). The sleeve shall be fabricated from transparent flexible polyethylene. The catheter hub shall be fabricated from rigid or semi-rigid plastic which meets the requirements of the USP for Class II plastics.

Needle. The needle shall be fabricated from Type 304 stainless steel in accordance with American Society for Testing and Materials (ASTM) Designation F899. The needle hub shall be fabricated from either Class 3, Type 304 steel, FEP, polypropylene, or other plastic meeting USP Class II requirements.

Wire stylet. The wire stylet shall be fabricated from Class 3, austenitic stainless steel in accordance with ASTM F899.

Needle cover, needle guard, and flow control plug. The needle cover, needle guard, and flow control plug shall be fabricated from rigid plastic.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:37Z
Check the source to verify that this is the current version before use.

Style and design. The catheter shall consist of a needle, the lumen through which an indwelling catheter shall pass. The needle shall be 14 gauge, have a lumen 2 ± 1/8 inches long and be of short bevel design. The needle point tip shall be ground to maximum keenness and the needle point heel shall be dull to prevent coring. The needle hub shall make a secure fit when pushed through the needle collar, which is permanently secured to the protective sleeve. The unit shall be supplied with a needle guard for safety prior to use and a folding needle cover for covering the needle after being withdrawn, following venipuncture.

Following venipuncture and needle withdrawal, the needle cover shall enclose and lock the entire needle cannula, preventing the needle bevel from shearing the catheter.

The catheter shall be permanently secured to the Luer-lock catheter hub if of two-piece construction. If of one piece construction, the proximal end shall incorporate a Luer-lock hub. The distal edges of the catheter shall be free from burrs and cuts. The wire stylet shall move freely within the catheter and have a flow control plug permanently secured to its proximal end.

The flow control plug shall have a Luer taper to provide leakproof seal with the catheter hub. A protective sleeve, covering from the needle collar to over the flow control plug, shall maintain the sterility of the catheter unit until opened.

The catheter shall be 16 gauge, 8 ± 1/2 inches. The catheter shall be radiopaque in accordance with ASTM F640. The radiopaque material may be incorporated into the plastic either as stripes or throughout the whole catheter.

Catheter hub and flow control plug. The female Luer-lock fitting on the catheter hub and the male Luer slip fitting on the flow control plug shall meet the requirements of ISO 594.

Catheter flexibility. With the wire stylet and the needle removed, the catheter shall be wrapped 90 degrees around a 1/2 inch diameter rod without kinking or cracking. Water shall then be admitted through the catheter hub and observed for restricted flow from the distal tip.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:37Z
Check the source to verify that this is the current version before use.

Catheter security of assembly. The catheter shall not tear or separate from the catheter hub when the catheter is subjected to a minimum tensile force of 2.5 pounds.

Catheter leakage. With the needle removed, the catheter shall not leak when subjected to the liquid leakage test specified in ISO 594. The catheter hub shall not leak when observed for two minutes.

Heavy metals extractability. Plastic components coming in contact with intravenous solution or patient's blood shall be tested. The quantity of lead and cadmium shall not exceed 5 parts per million (ppm) and the amount of barium shall not exceed 10 ppm when tested in accordance with the USP (See NOTE 3).

Sterility, pyrogenicity, toxicity. The catheter and needle unit shall be nonpyrogenic and nontoxic in accordance with USP. The unit shall be sterile in accordance with USP or AAMI requirements. Sterility assurance level shall be 0.0001.

Workmanship. The catheter shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing fifty individually packaged sterile catheter and needle units, as specified, constitutes one unit. The individual package shall be a peel-open type container capable of maintaining sterility of contents unless the package is damaged or opened.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspection are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one catheter needle unit (with syringe when specified). Sampling shall be inspection level S-1 with an AQL of 1.0.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:37Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE (1): The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-174-9893	Catheter and Needle Unit Intravenous, Radiopaque, 16 Gauge by 8 inch Catheter, 50's

NOTE (2): The following is acceptable as an alternate to USP Class IV plastics lot testing: The catheter shall be initially qualified for Class IV plastics in accordance with USP, and simultaneously, tested for tissue culture cytotoxicity in accordance with HIMA "Guideline For Evaluating The Safety Of Materials Used In Medical Devices". Subsequent production lots which show greater toxicity than the initial qualification lot

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

shall be subjected to USP Class IV biological testing. In the event of a change in material formulation, USP, Class IV initial qualification testing shall again be required.

NOTE (3): Certificates of heavy metals testing. A certificate of compliance, stipulating that offered items meet the heavy metals extractable testing, shall be acceptable, provided that the contractor has supplied acceptable items to the Government within the past three years and that there have been no changes to material formulation and method of processing since actual testing was performed.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5470
Location: c:\word\doc\levin\catheter

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given

2 The submitter of this form must complete blocks 4, 5, 6, and 7.

3 The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER A-A-54737	2 DOCUMENT DATE (YYMMDD) 16 October 1992

3 DOCUMENT TITLE
CATHETER AND NEEDLE UNIT, INTRAVENOUS, 16 GAUGE, RADIOPAQUE, DISPOSABLE (Catheter into Needle)

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, Middle Initial)	b ORGANIZATION	
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7 DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME Defense Personnel Support Center DPSC-MSE	b TELEPHONE (Include Area Code) (1) Commercial 215-737-2870 (2) AUTOVON 444-2870
c ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT. Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756 2340 AUTOVON 289-2340

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.