METRIC

A-A-54751
21 September 1992

COMMERCIAL ITEM DESCRIPTION

DISH, BIOLOGICAL PREPARATION (Stender Design)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-D-36622 which is cancelled.

This Commercial Item Description covers Stender design biological preparation dishes with ground glass covers, for use in biological laboratory procedures.

Salient characteristics:

Shall be Stender type preparation dishes of the following types and sizes:

Type I. Low form design
Size (diameter X height)
37 X 25 mm
48 X 33 mm
50 X 25 mm
60 X 28 mm
62 X 36 mm

Type II. High form design
Size (diameter X height)
60 X 95 mm

A tolerance of ± 5 mm shall be applicable to the dimensions for the diameter and height.

Shall provide each dish with a loose fitting cover that includes a circular groove to fit over the rim of the dish and prevent slipping or evaporation.

Shall provide dishes and covers fabricated of clear heavy molded glass.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Shall provide dishes that are cylindrical in shape. The base of each dish shall be flat so as to rest on a flat level surface without rocking or wobbling. The top edge of each dish shall be ground to permit the applied cover to seal the dish.

Shall provide each dish with a circular flat plate cover that is 7 mm in thickness. The cover shall include a recessed circular grove that is 0.5 mm in depth and is ground to mate with the rim of the dish. The recessed circular groove with ground finish shall extend around the periphery of the dish.

Shall provide each dish with legible and permanent markings that include the name or registered trademark of the manufacturer.

Workmanship. The dishes and covers shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One dish with cover, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Inspections and tests. Inspections and tests shall be conducted on the dishes and covers in accordance with commercial practice to determine compliance with specification requirements. Sampling for inspections and tests shall also be conducted in accordance with commercial practice.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor Certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Size dish required, intermediate/exterior package quantities, labeling and marking must be specified in the contract and/or order.

NOTES: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	Specification Designated SIZE	TYPE
6640-00-422-4000	DISH, BIOLOGICAL PREPARATION, Stender.	50 x 25	I
6640-00-422-5000	DISH, BIOLOGICAL PREPARATION, Stender.	60 x 95	II

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6640-1782

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given

2 The submitter of this form must complete blocks 4, 5, 6, and 7.

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER	2 DOCUMENT DATE (YYMMDD)
	A-A-54751	21 SEPTEMBER 1992

3. DOCUMENT TITLE
DISH, BIOLOGICAL PREPARATION (Stender Design)

4 NATURE OF CHANGE (*Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed.*)

5. REASON FOR RECOMMENDATION

6. SUBMITTER		
a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a NAME	b TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	(215) 737-2870	444-2870
c ADDRESS (Include Zip Code) ATTN: MSE 2800 S. 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 *Previous editions are obsolete* 0102-LF-009-2700 eU S Government Printing Office: 1990-704-119/20055 2-1 198/

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.