INCH-POUND

A-A-54759
15 October 1992

COMMERCIAL ITEM DESCRIPTION

CANISTER AND TIMER SYSTEM

(Oropharyngeal Suction Apparatus)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a system of 3 canisters connected to a manifold and an electrical timer to convert a 115 volt suction apparatus from constant flow to intermittent flow.

Salient characteristics:

The unit shall be a three-canister and timer system for Wangensteen type intermittent suctioning of 3 patients simultaneously providing an on-off timer for intermittent interruption of power supply to a suction apparatus. It shall be suitable for use in hospital and military aeromedical aircraft.

The system shall include the following:

a. A vacuum supply hose (tubing), 3/8 inch ID by at least 6 feet long to connect the system's manifold to a suction apparatus.

b. A vacuum gauge, with connector for the supply hose, with a range from 0 to 760 mm Hg vacuum, connected to the manifold.

c. A manifold to receive vacuum from a suction apparatus and feed adjustable vacuum to 3 suction canisters.

d. Three vacuum adjustment valves, mounted on the manifold, each with a fitting for connection of the canisters' vacuum lines.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

e. Three vacuum lines, 1/4 inch ID by at least 6 feet long, for connection between the manifold's adjustable vacuum valves and 3 canisters.

f. Three disposable suction canisters, each with working capacity of at least 1200 cc having graduation markings conforming with ANSI Z79.9 and having a barb or tapered connector for 1/4 patient tubing.

g. An electrical control box to provide ON-OFF cycling of 115 volt, 50 to 400 Hz power source to a suction apparatus. The control box, with its power cord assembly, shall be designed to control Electromagnetic Interference (EMI) and Radio Frequency Interference (RFI). The control box shall include:

1. A flush mounted 115 volt receptacle to receive a power cord with a hospital grade plug to provide cycled on-off power to a surgical suction apparatus (marked to identify as timed/cycled on-off power).

2. A flush mounted 115 volt hospital grade receptacle for general purpose use (not time/cycled controlled).

3. A 115 volt cord with hospital grade plug, hard wired, to allow input power from a wall receptacle.

4. A timer to intermittently interrupt power from 115 VAC wall source providing power cycling of 15 seconds off and 45 seconds on. The switch circuit shall be rated at least 5 amps.

h. Three mounting bracket sets to mount each of the canisters on a stanchion with cross-section of either 2 inches by 2 inches or 2-3/4 inches by 2 inches.

The system shall be compatible with the oropharyngeal suction apparatus described in Commercial Item Description A-A-54039.

Workmanship. The canister and timer system shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One canister and timer system, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Operational and Service Data Manuals shall be furnished as follows:

a. Information content:

The contractor shall furnish manuals, handbooks, and/or brochures containing complete operation, installation and service/maintenance instructions (including pictures or illustrations, as necessary) with complete schematics and wiring diagrams (if unit has electric circuitry). All manuals, handbooks, and/or brochures shall be written in the English language and all schematics and wiring diagrams shall use American electrical data and connection diagrams for all applicable utilities. The instructions shall also contain a complete list of all replaceable parts showing part number, name and quantity required. All service notes, service memo, etc., pertaining to the equipment and issued subsequent to the printing of the original instructions shall also be included. These instructions shall be, as a minimum, that furnished to service engineers (or servicemen) that normally install and service the equipment for the company or distributor. When the system being procured includes, or will operate ancillary equipment, the service data shall include complete instructions and drawings which show interfacing of all system components.

b. Quantities:

Two (2) complete sets of the above information shall be packaged with each unit of the required equipment.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of I. The unit of product for examination purposes shall be one canister and timer system, as specified.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of I. The unit of product for test purposes shall be one canister and timer system, as specified.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:38Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

6515-01-356-4830

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

A-A-54759

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA - OSS

Project Number: 6515-5458

Location: TIMEF12C.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54759	2. DOCUMENT DATE (YYMMDD) 15 October 1992

3. DOCUMENT TITLE
CANISTER AND TIMER SYSTEM (Oropharyngeal Suction Apparatus)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 96/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.