METRIC

A-A-54762
19 October 1992

COMMERCIAL ITEM DESCRIPTION

PAN, BIOLOGICAL STAINING

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-P-002240B which is cancelled.

This Commercial Item Description covers one type of metal staining pan for use in laboratory procedures.

Salient characteristics:

Shall be a leakproof rectangular pan. Shall be fabricated from 23 gage (0.64 mm nominal) copper sheet, and shall include a brass tube insert for drainage purposes.

Shall conform to Figure 1 for design and dimensions.

Shall provide a pan constructed from three (3) sheets of copper. One copper sheet shall be used to form the bottom and two sides of the pan. Each of the other two (2) copper sheets shall be used to form a side end. The resulting joints of the copper sheets shall be lapped and soldered continuously.

Shall provide pan with a brass tube insert conforming to Figure 1. The brass insert shall be soldered flush with the orifice on the bottom of the pan, as indicated in Figure 1. The outer surface at the terminal end of the brass tube shall include circular grooves that are suitable for providing a friction fit with inserted rubber tubing of appropriate diameter.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Shall provide pan designed and constructed to include the following (see Figure 1):

The upper rim of the pan is bent to form continuous, half-turned edges on each of the four sides.

The bottom of the pan has sloping surfaces that slope gradually downward from each side toward the center.

Shallow grooves inside the pan that facilitate draining and are pressed into the metal and extend from each bottom corner to the center.

The center of the pan includes a circular depression with an orifice termination.

Shall provide a pan with a leakproof construction. When the brass tube opening is properly blocked, the water filled pan shall exhibit no leakage as viewed on the underside.

Shall provide a pan having smooth surfaces and is free of sharp edges.

Workmanship. The pan shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One pan, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted on the pan in accordance with commercial practice to determine compliance with specification requirements. Sampling for inspections and tests shall be in accordance with commercial practice.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor Certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	Item Identification
6640-00-402-9250	PAN, BIOLOGICAL STAINING, Rectangular

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM


Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6640-1781

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.



DIMENSIONS	TOLS.
0 TO 1.0 CM.	± 0.05 CM.
1.0 TO 5.0 CM.	± 0.2 CM.
OVER 5.0 CM.	± 0.5 CM.

NOTE:
ALL DIMENSIONS ARE IN CENTIMETERS.





SECTION A-A

FIGURE 1. PAN, BIOLOGICAL, STAINING

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE. This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:

1. DOCUMENT NUMBER: A-A-54762
2. DOCUMENT DATE (YYMMDD): 19 OCTOBER 1992

3. DOCUMENT TITLE: PAN, BIOLOGICAL STAINING

4. NATURE OF CHANGE (*Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.*)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

[redacted]

7. PREPARING ACTIVITY

a. NAME

b. TELEPHONE (*Include Area Code*)
(1) Commercial (2) AUTOVON

c. ADDRESS (*Include Zip Code*)

IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT:
Defense Quality and Standardization Office
5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466
Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 *Previous editions are obsolete* 198/29

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.