INCH-POUND

A-A-54763
19 October 1992

COMMERCIAL ITEM DESCRIPTION

PAPER, GLASSINE, WEIGHING

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-P-36709 which is cancelled.

This Commercial Item Description covers glassine paper suitable for weighing and transferring powders, crystals, and other preparations in pharmaceutical and laboratory procedures.

Salient characteristics:

The glassine paper shall be fabricated from a semi-chemical wood pulp and have a basis weight (24 X 36 - 480) of 25 ± 5 pounds. Shall have a bursting strength of not less than 15 pounds per square inch (psi). The glassine paper shall have an internal tear resistance of not less than 0.42 ounces (oz) (12 grams). The glassine paper shall have a turpentine transduation time (grease resistance) of not less than 600 seconds.

The glassine paper shall have a pH value of 7.0 ± 1.0. The water soluble acidity of glassine paper shall be not greater than 0.02 percent calculated as sulfite (SO_3=) ion, when tested herein.

The glassine paper shall be furnished in sheet form. Each sheet shall be 4-1/2 inches (in) long by 3-1/2 inches wide. The tolerance on the dimensions shall be ± 1/8 inch. The glassine paper shall have a thickness of 0.00125 ± 0.00020. The number of sheets of glassine paper per unit package may vary by ± 2 percent. The average count shall be 1000 sheets. The glassine paper shall have a smooth glazed surface on both sides and a uniform light green tint for background contrast. The paper shall be sufficiently clear and transparent to allow easy reading of 10-point bold type.

AMSC N/A FSC 8135

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Workmanship. The glassine paper shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 1000 sheets, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-2. The unit of product for examination purposes shall be the unit package.

Sampling for tests. Sampling for tests (dimensional), shall be conducted in accordance with MIL-STD-105, with an AQL of 0 (percent defective) and an inspection level of S-3. The unit of product for test purposes shall be the unit package.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to the following:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Basis weight. Basis weight shall be determined in accordance with TAPPI Test Method T410 os-61.

Bursting strength. The bursting strength shall be determined in accordance with procedures in TAPPI Standard T403 ts-63.

Tear resistance. The tear resistance shall be determined in accordance with TAPPI Test Method T414 ts-65.

Grease resistance. The grease resistance shall be determined in accordance with TAPPI Test Method T454 ts-66.

Acidity (pH). The pH value shall be determined in accordance with TAPPI Test Method T435 su-68.

Water soluble acidity. The water soluble acidity shall be determined in accordance with TAPPI Test Method T428 su-67.

Thickness. The thickness shall be determined in accordance with TAPPI Test Method T411 m-44.

Transparency. The transparency shall be determined as follows: Type the following sentence on a suitable grade of typing paper using 10 point bold type: "Glassine paper shall be sufficiently clear and transparent to allow easy reading of 10 point type." Place the glassing paper directly over the typed sheet. There shall be no difficulty in reading the typed sentence.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Contractor Certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	Item Identification
8135-00-407-3453	PAPER, GLASSINE, WEIGHING, 3-1/2 by 4-1/2 inches, 1000s

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 8135-0656

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54763	2. DOCUMENT DATE (YYMMDD) 19 OCTOBER 1992

3. DOCUMENT TITLE

PAPER, GLASSINE, WEIGHING

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON
c. ADDRESS (Include Zip Code)	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.