INCH-POUND

A-A-54766
30 October 1992

COMMERCIAL ITEM DESCRIPTION

RACK, TEST TUBE, LABORATORY, POLYPROPYLENE

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specifications MIL-R-0036837A and MIL-R-36837 which are cancelled.

This commercial item description covers autoclavable polypropylene plastic test tube racks suitable for use in medical laboratories.

Salient characteristics:

Classification. The racks shall be furnished in the following sizes:

Size 1 - 40 Tubes
Size 2 - 72 Tubes

Material. The racks shall be fabricated from suitable polypropylene material that is capable of meeting performance requirements indicated herein. If welding is used to bond components of the rack, the bonding material shall be the same material as the rack.

Design and construction. The racks shall be constructed in accordance with either Figure 1 (for size 1) or Figure 2 (for size 2), as applicable. Variations in contours of the rack from that indicated in the applicable figure are acceptable when not limited by requirements or dimensions. Each rack shall have a rectangular base of stiffened design and two upper tiers. The base and tiers shall be firmly secured to end support integral by design or by means of either mechanical joints or welding. The upper and lower tiers shall have an equal number of properly aligned holes; and the base shall have the same number of countersunk holes to accommodate the test tube bottoms. The center lines of

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

corresponding holes in the base and tiers of the rack shall be in proper alignment to assure that the inserted tubes rest vertically in the rack. The cutout for each end support shall serve as a carrying handle. The top and bottom surfaces of the end support shall be designed to permit secure stacking of the same size test tube racks.

Size 1 racks. The size 1 racks shall have 40 holes which are arranged in 4 rows with 10 holes per row equidistant from each other. Each hole shall be circular or square and shall accommodate test tubes 16 to 20 mm in diameter.

Size 2 racks. The size 2 racks shall have 72 holes which are arranged in 6 rows with 12 holes per row equidistant from each other. Each hole shall be circular or square and shall accommodate test tubes 10 to 13 mm in diameter.

Performance. Each rack shall be capable of withstanding steam sterilization by autoclaving at 250°F for a period of 30 minutes without disassembly, becoming tacky or distorting.

Workmanship. Workmanship shall be first class throughout, and each rack shall be free from defects which detract from its appearance or may impair its serviceability. All surfaces (including cutouts) shall be smooth and free of sharp edges, surface craze, mold flash, or projections and burrs (except for mold marks due to gating). Mold lines shall not be cause for rejection provided there are no flash protrusions. Burrs due to gating shall not be cause for rejection, provided there is no evidence of sharpness or projections that protrude more than 1/64 inch from the surrounding area. Each rack shall not wobble when stacked on top of another rack.

Unit. Each (EA). One rack of the size specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for theperformance of all inspection requirements as specified herein. Except as otherwise specified in the contract or

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in specification where such inspection are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one test tube rack as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one test tube rack. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers and their respective Item Identifications are covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION	SIZE
6640-00-159-2504	RACK,TEST TUBE, LABORATORY, Polypropylene, 40 Tube	1
6640-00-159-2506	RACK, TEST TUBE, LABORATORY, Polypropylene, 72 Tube	2

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6640-1780
Location: c:\word\nish\rackcid

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:38Z
Check the source to verify that this is the current version before use.





FIGURE 1 - Jumper

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.



NOTE.

AS AN ALTERNATE TO A CLEAR HOLE INDICATED IN SECTION A-A, A COUNTERSUNK HOLE CONTAINING A CROSS MEMBER SHALL BE ACCEPTABLE.

SIZE 2

FIGURE 2

ALL DIMENSIONS IN INCHES

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:

1 DOCUMENT NUMBER: A-A-54766

2 DOCUMENT DATE (YYMMDD): 30 October 1992

3 DOCUMENT TITLE: RACK, TEST TUBE, LABORATORY, POLYPROPYLENE

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5 REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)

b. ORGANIZATION

c. ADDRESS (Include Zip Code)

d. TELEPHONE (Include Area Code)
(1) Commercial
(2) AUTOVON (If applicable)

7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a NAME: Defense Personnel Support Center
ATTN: DPSC-MSE

b TELEPHONE (Include Area Code)
(1) Commercial: (215)737-2870
(2) AUTOVON: 444-2870

c ADDRESS (Include Zip Code)
2800 South 20th Street
Philadelphia, PA 19145

IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT:
Defense Quality and Standardization Office
5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466
Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — *98-290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.