INCH-POUND

A-A-54767
20 November 1992

COMMERCIAL ITEM DESCRIPTION

CATHETER AND NEEDLE UNIT, INTRAVENOUS, 14 gauge Catheter, Radiopaque, Disposable

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a radiopaque catheter and needle unit suitable for administering intravenous medication from a single venipuncture.

Salient characteristics:

Material.

Catheter. The catheter shall be fabricated from pliant Fluorinated Ethylene Propylene (FEP) or polypropylene which meets the requirements of the United States Pharmacopeia (USP) for class IV plastics. (See Note 2). When the catheter is of two-piece construction, the hub shall be fabricated from rigid or semi-rigid plastic which meets the requirements of USP for class II plastics.

Needle. The needle shall be fabricated from type 304 stainless steel in accordance with ASTM Designation F899. The needle hub shall be fabricated either from class 3, type 304 steel, FEP, polypropylene, or other plastic meeting USP class II requirements.

Needle guard and flow control plug. The needle guard and flow control plug shall be fabricated from rigid plastic.

Style and design. The unit shall consist of an outer catheter assembled over an inner hypodermic needle of short bevel design. The catheter shall be 14 gauge and shall have a free length of 2-1/4 ± 4 inches. The needle shall be 16 or 17 gauge. The needle cannula shall make a safe, sliding fit within the catheter. The proximal end of the catheter shall be permanently secured to a Luer-Lock catheter hub of either one or two piece construction, and in accordance with

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:38Z
Check the source to verify that this is the current version before use.

ISO 594. The proximal end of the needle cannula shall be permanently secured to the distal end of the needle hub. The proximal end of the needle hub shall incorporate a flash chamber and terminate in a female Luer-Slip or Luer-Lock configuration in accordance with ISO 594. A vented plug shall be included. The catheter hub shall be firmly anchored flush against the needle hub to prevent needle movement within the catheter until ready for use. The distal tip of the catheter shall have a smooth taper or step and gentle transition to ease catheter insertion and reduce vascular trauma after venipuncture. The needle point tip shall be ground to maximum keenness and the needle point heel shall be dull to prevent coring. Catheter and needle units shall be radiopaque in accordance with ASTM Designation F640.

The radiopaque material may be incorporated into the plastic either as stripes or throughout the whole catheter cannula. All units shall be supplied with a needle guard.

Catheter flexibility With the needle removed, the catheter shall be wrapped 90 degrees around a 1/2 inch diameter rod without kinking or cracking. Water shall then be admitted through the catheter hub and observed for restricted flow from the distal tip

Catheter security of assembly. The catheter shall not tear or separate from the catheter hub when the catheter is subjected to the minimum tensile force of 2.5 pounds.

Catheter leakage. With the needle removed, the distal tip of the catheter shall be sealed and 20 psig water admitted through the catheter hub. The junction of the catheter to catheter hub shall not leak when observed for 2 minutes.

Heavy metals extractability Plastic and rubber components coming in contact with intravenous solution or patient's blood shall be tested. The quantity of lead and cadmimum shall not exceed 5 parts per million (ppm) and the amount of barium shall not exceed 10 ppm when tested in accordance with the USP (See Note 3).

Sterility, pyrogenicity, toxicity. The catheter and needle unit shall be non-pyrogenic and non-toxic in accordance with USP. The catheter and needle unit shall be sterile in accordance with USP or AAMI requirements. Sterility assurance level shall be 0.0001.

Source: https://assist.dla.mil -- Downloaded: 20260-26T09:38Z
Check the source to verify that this is the current version before use.

Workmanship. The catheter and needle unit shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 200 individually packaged sterile catheter and needle units, as specified, constitutes one unit. The individual package shall be a sealed, peel-open type container capable of maintaining sterility of contents unless package is damaged or open.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an Acceptable Quality Level (AQL) of 2.5 (percent defective) and an inspection level of II. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. the

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

unit of product for inspection purposes shall be one unit of issue.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, inspection level S-1 and AQL 1.0. Test samples shall be selected from each manufacturing lot such as that each material lot is represented in the samples. The unit of product for test purposes shall be one catheter and needle unit (with syringe when specified).

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

NOTE (1): The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6515-01-169-7473	CATHETER AND NEEDLE UNIT, INTRAVENOUS, Radiopaque, 14 gauge Catheter, 2 $^{1}/_{4}$ inches

NOTE (2): The following is acceptable as an alternate to USP Class IV plastics lot testing: The catheter shall be initially qualified for Class IV plastics in accordance with the USP, and simultaneously, tested for tissue culture cytotoxicity in accordance with HIMA "Guideline For Evaluating The Safety of Materials Used in Medical Devices". Subsequent production lots which show greater toxicity than the initial qualification lot shall be subjected to USP Class IV biological testing In the event of a change in material formulation, USP Class IV initial qualification testing shall again be required.

NOTE (3): Certificates of heavy metals testing. A certificate of compliance, stipulating that offered items meet the heavy metals extractable testing, shall be acceptable, provided that the contractor has supplied acceptable items to the Government

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:38Z
Check the source to verify that this is the current version before use.

within the past three years and that there have been no changes to material formulation and method of processing since actual testing was performed.

NOTE (4): In lieu of having each material batch represented in the finished goods chemical, safety and biological test samples, the following shall be acceptable as an alternate: New raw materials shall be initially qualified through chemical, safety, and biological testing and again testing after assembly, sterilization, and aging. Subsequent incoming materials will be tested physically for material consistency. Any changes to material or processing shall be requalification tested.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-5463
Location: ML Disk #2
(RADIO)

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.

2 The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER A-A-54767	2 DOCUMENT DATE (YYMMDD) 20 November 1992

3. DOCUMENT TITLE
CATHETER AND NEEDLE UNIT, INTRAVENOUS, 14 gauge Catheter, Radiopaque, Disposable

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed.)

5 REASON FOR RECOMMENDATION

[redacted]

8. PREPARING ACTIVITY

a NAME	b TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center ATTN: DPSC-MSE	(215)7372870	444-2870
c ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT. Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 Previous editions are obsolete 19629

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.