INCH-POUND

A-A-54769
18 December 1992

COMMERCIAL ITEM DESCRIPTION

CATHETERIZATION SET, URETHRAL
Disposable, Plastic

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a complete sterile, disposable, pre-assembled, ready-to-use urethral catheterization set with catheter.

Salient characteristics:

General. The disposable catheterization set shall be a complete set of items necessary for urethral catheterization.

Characteristics. The catheterization set shall consist of, as a minimum, the following disposable components:

Component	Quantity
Balls, prepping	5
Basin, plastic, 1200 ml, anatomically shaped	1
Catheter, 14-16 Fr, plastic	1
Container, specimen, plastic	1
Cup, prep, removable	1
Drape, fenestrated	1
Forceps, plastic	1
Gloves, latex rubber, universal size	2
Packet, lubricating jelly	1
Solution, povidone-iodine 10 percent N.F. 1 fl oz. (29.5 ml)	1
Underpad, waterproof	1

Prepping balls. The prepping balls shall be natural cotton or viscose rayon and shall be approximately 1-1/2 inches in diameter.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Basin. The anatomically shaped basin shall be 9-3/4 inches long, 4-1/2 inches wide by 4-1/4 inches deep. The basin shall be fabricated of polyethylene or polystyrene and shall be marked in 100 cc increments from 200 cc to 1100 cc.

Catheter. The catheter shall be fabricated of polyvinyl chloride (PVC) plastic. Catheter shall be 14-16 Fr with a length of 17 ± 1 inches. A flared soft plastic adapter shall be provided at the proximal end and two eyes for drainage at the distal end.

Specimen container. The specimen container shall be fabricated of plastic and shall be of 4 oz., 120 cc capacity. Side of container shall be marked at 1/4 oz. and 10 cc increments. Every 1/2 oz. and 10 cc increments shall be numbered.

Prep cup. The prep cup shall be 5 inches long, 4-1/2 inches wide by 2-1/4 inches deep. Cup shall be fabricated of polyethylene plastic.

Fenestrated drape. The fenestrated drape shall have minimum dimensions of 12-1/2 inches wide by 15-1/2 inches long. Drape shall be fabricated of absorbent materials with a fenestrated center.

Forceps. The forceps shall be fabricated of plastic and shall be one piece. Serrations shall be provided at the tip for gripping and on the sides for manual use. Forceps shall be approximately 5 inches long.

Examining gloves. The examining gloves shall be of universal size fabricated of latex. Gloves shall be pre-powered and folded or packaged to facilitate donning without violating aseptic technique. The gloves shall be free of seams or welts in the palm or finger tip areas.

Lubricating jelly. The lubricating jelly shall be sterile and water soluble. A minimum of 2.7 grams shall be provided in a suitable foil packet.

Povidone-iodine solution. The povidine-iodine solution shall be 10 percent NF and shall be provided in a foil packet containing at least 27.5 ml.

Absorbent underpad. The absorbent underpad shall be coated on one side with a water impermeable barrier. The underpad shall be 19-1/2 ± 1 inches long (minimum) by 15 inches wide minimum.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:38Z
Check the source to verify that this is the currentversion before use.

Overwrap. The set components shall be placed in the basin and the basin shall be overwrapped with Central Supply Room (CSR) sterilization overwrap. The overwrapped basin shall be placed in a sealed plastic bag or a peel-open type container capable of maintaining sterility until opened or damaged. The set components shall be packed in the basin in the proper sequence of use.

Biocompatibility. The catheter shall meet the biocompatibility requirements for Class II plastics in accordance with the USP.

Sterility. The catheter set shall be sterile in accordance with the USP or AAMI guidelines. Sterility assurance level shall be 0.0001.

Workmanship. The set shall be free from defects which detract from its appearance or impair its serviceability. Set components shall be free from flash, sharp edges, tears, cuts, holes, distortion or foreign material.

Unit. Package (PG). One package containing twenty individually packaged sterile catheterization set, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspection are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one catheterization set as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one catheterization set. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-043-4008	CATHETERIZATION SET, URETHRAL, Disposable, Plastic, 20s

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:38Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5472
Location: c:\word\nish\cathset

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given

2 The submitter of this form must complete blocks 4, 5 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER A-A-54769	2 DOCUMENT DATE (YYMMDD) 18 December 1992

3 DOCUMENT TITLE

CATHETERIZATION SET, URETHRAL, Disposable, Plastic

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, Middle Initial)	b ORGANIZATION	
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7 DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME Defense Personnel Support Center ATTN: DPSC-MSE	b TELEPHONE (Include Area Code) (1) Commercial (215)737-2870	(2) AUTOVON 444-2870
c ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403 Falls Church VA 22041 3466 Telephone (703) 756 2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 *Previous editions are obsolete* '98/

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.