INCH-POUND

A-A-54771
18 December 1992

COMMERCIAL ITEM DESCRIPTION

LIGHT, HEAD, DIAGNOSTIC, LEMPERT-STORZ

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-L-37704 which is cancelled.

This commercial item description covers an adjustable headlight, of the Lempert-Storz design, suitable for use in operating and other medical procedures.

Salient characteristics:

Shall be a headlight of the Lempert-Storz design, suitable for use in general medical and surgical procedures. The unit shall be supplied in a case, complete with headband, mirror, lamp, transformer, bayonet-type locking connectors line cord with 3-prong plug and an extra lamp. The headband shall be made from durable lightweight, non-staining nylon and shall be lined with a washable cloth suspension lining. The mirror shall be made from high quality optical glass, free from defects. The mirror shall be back surfaced.

The lamp shall be of the closely wound filament type matched to the fixed voltage of the transformer. The lamps shall have a current rating of 1 amp., with voltage rating of 8 volts and a minimum life of 25 hours. The lamp shall provide a square of light in the central field of the lamps to reduce filament shadow to a minimum.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

The transformer shall have a primary input of 115 volts, 50/60 Hz and a secondary output of 8 volts AC. The transformer shall be equipped with a flexible, insulated, 3-wire cord as shown on DPSC Drawing No. 20719. A plug, in accordance with ANSI Standard C73.11 shall be attached to one end of the power cord. A cord carrying the secondary transformer voltage shall be permanently attached to the transformer and shall have a insulated shielded cable with high tensile strength. Bayonet lock connectors shall be provided for connecting and disconnecting the transformer assembly from the headlight proper.

Adjustment of the headband shall be easy and self-locking. The unit shall be easily and effectively adjustable by means of a spiral gear assembly that shall provide positive locked beam focusing. The focus of the headlight shall remain unchanged during normal movements of the operator, or when the mirror assembly is adjusted. The mirror assembly shall be adjustable by means of three supporting tension arms that shall permit both vertical and lateral adjustment. The headband shall permit quick, easy adjustment without the use of locking or pressure nuts.

The headband suspension lining shall effect weight distribution of the headlight, eliminating pressure points. The mirror shall be specially ground, with polished optical surface, full size, and be readily removed and assembled to mirror metal housing. Mirror shall allow the operator director indirect vision into a surgical cavity in the same plane as reflected light. The illumination of the headlight assembly shall be brilliant and shadow free when connected to a nominal 115 volt, 50/60 Hz, electrical source.

Identification marking. Each headband assembly and transformer of the headlight assemblies shall be permanently and legibly marked with the manufacturers name or registered trademark.

Case. The complete unit, including extra lamp shall be supplied in a suitable case. The interior of the case shall be designed to hold the components securely in place.

Instructions for use. An instruction manual containing instructions for use and maintenance shall be provided with each unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Workmanship. The diagnostic headlight shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One diagnostic headlight as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspection are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one diagnostic headlight.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 20261O-26T09:38Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests (dimensional) and illumination, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for test purposes shall be one diagnostic headlight.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-00-074-4190	LIGHT, HEAD, DIAGNOSTIC LEMPERT STORZ

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5464
Location: c:\word\doc\eric\lightcid

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.

2 The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE. This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER	2 DOCUMENT DATE (YYMMDD)
	A-A-54771	18 December 1992

3 DOCUMENT TITLE

LIGHT, HEAD, DIAGNOSTIC, LEMPERT STORZ

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite if possible Attach extra sheets as needed.)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, MI)	b ORGANIZATION
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	e DATE SUBMITTED (YYMMDD)

7 PREPARING ACTIVITY

a NAME	b TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center ATTN: DPSC-MSE	(215) 7372870	444-2870
c ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 Previous editions are obsolete

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:38Z
Check the source to verify that this is the current version before use.