INCH-POUND

A-A-54773A
15 March 1994
SUPERSEDING
A-A-54773
18 December 1992

COMMERCIAL ITEM DESCRIPTION

NEEDLE, HYPODERMIC, DISPOSABLE

The General Services Administration has authorized the use of this Commercial Item Description

This Commercial Item Description covers sterile, disposable, hypodermic needles.

Salient characteristics:

Except as specified herein the needle shall be in accordance with International Standard Organization (ISO) 7864, Sterile Hypodermic Needles for Single Use.

Types:

Type I	-	Regular bevel, Regular Wall
Type II	-	Short bevel, Regular Wall
Type III	-	Regular bevel, Thin Wall
Type IV	-	Short bevel, Thin Wall

General. The needle shall consist of a hub, cannula, and sheath (needle guard).

Material. The cannula shall be fabricated from Class 3, Type 304 stainless steel in accordance with American Society for Testing Materials (ASTM) Designation F899. The hub shall be fabricated from rigid polypropylene that meets the United States Pharmacopeia (USP) requirements for Class IV plastics or aluminum. The sheath shall be fabricated from rigid plastic. The exterior of the cannula shall be lubricated with inert non-toxic medical grade silicone compound.

Style and design. The needle shall be furnished with a removable, protective sheath. The tip of the point shall be ground to maximum keenness. The heel of the needle point shall be dull to prevent coring. The needle shall have a luer lock hub in accordance with ISO 594-2.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Cannula length (inch)	Tolerance (inch)
1/2, 5/8	+ 3/64, - 1/16
3/4 - 1-1/2	± 1/16

Gauge	Point length (inch) (Reg. Bevel)
18	0.203 - 0.249
19	0.175 - 0.215
20	0.148 - 0.180
21	0.133 - 0.163
22	0.120 - 0.146
23	0.105 - 0.129
25	0.078 - 0.103
26	0.059 - 0.095

The point length for the 18 gauge short bevel needle shall be 0.135 to 0.160 inch.

The point length for the 19 gauge short bevel needle shall be 0.112 to 0 139 inch.

Aluminum hub. Needles with aluminum hubs shall show no evidence of corrosion when covered with 3 percent sodium chloride solution at 20-25°C for one hour. There shall be no discoloration visible or any sign of reaction.

Non-toxic and pyrogen-free. The needle shall be non-toxic and pyrogen-free when tested in accordance with USP.

Sterility test. The needle shall be tested for sterility in accordance with the USP or American Association for Mechanical Instrumentation (AAMI) requirements. Sterility assurance level shall be 0.0001.

Workmanship. The needle shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing one hundred individually packaged sterile needles of one gauge only, as specified, constitutes one unit. The individual package shall be a sealed peel-open container or rigid plastic tamper proof container capable of maintaining sterility of contents unless the package is damaged or opened. The package design shall permit ready dispensing of the needle from the package by sterile techniques and by attaching the syringe to the hub of the needle without touching the needle with hands or other instruments.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:38Z
Check the source to verify that this is the current version before use.

Color Coding. The hub of the needle shall be color coded. The color designations specified in ISO 6009 are optional. If colored the needle sheath is not required to meet ISO 6009.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspection are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one needle, as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 2.5.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one needle. Sample shall be Inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE (1): The following National Stock Numbers and their respective Item Identifications are covered by this document:

NATIONAL STOCK NUMBERS	ITEM IDENTIFICATIONS	TYPES
6515-00-655-5751	NEEDLE, HYPODERMIC, DISPOSABLE, 25 Gauge, 5/8 Inch, 100s	I
6515-00-754-2834	NEEDLE, HYPODERMIC, DISPOSABLE, 18 Gauge, 1-1/2 Inch, 100s	I
6515-00-754-2835	NEEDLE, HYPODERMIC, DISPOSABLE, 22 Gauge, 1 Inch, 100s	I
6515-00-754-2836	NEEDLE, HYPODERMIC, DISPOSABLE, 20 Gauge, 1-1/2 Inch, 100s	I
6515-00-754-2837	NEEDLE, HYPODERMIC, DISPOSABLE, 26 Gauge, 1/2 Inch, 100s	I
6515-00-754-2838	NEEDLE, HYPODERMIC, DISPOSABLE, 21 Gauge, 1-1/4 Inch, 100s	I
6515-00-754-2839	NEEDLE, HYPODERMIC, DISPOSABLE, 23 Gauge, 3/4 Inch, 100s	I
6515-01-140-5327	NEEDLE, HYPODERMIC, DISPOSABLE, 19 Gauge, 1-1/2 Inch, 100s	IV
6515-01-172-7650	NEEDLE, HYPODERMIC, DISPOSABLE, 22 Gauge, 1-1/2 Inch, 100s	I
6515-00-059-5781	NEEDLE, HYPODERMIC, DISPOSABLE, 22 Gauge, 1 Inch, 100s	I

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

NATIONAL STOCK NUMBERS	ITEM IDENTIFICATIONS	TYPES
6515-00-059-5782	NEEDLE, HYPODERMIC, DISPOSABLE, 22 Gauge, 1-1/2 Inch, 100s	I
6515-00-059-5783	NEEDLE, HYPODERMIC, DISPOSABLE, 23 Gauge, 3/4 Inch, 100s	I
6515-00-059-5786	NEEDLE, HYPODERMIC, DISPOSABLE, 18 Gauge, 1-1/2 Inch, 100s	II
6515-00-734-4375	NEEDLE, HYPODERMIC, DISPOSABLE, 20 Gauge, 1 Inch, 100s	I
6515-00-762-5361	NEEDLE, HYPODERMIC, DISPOSABLE, 19 Gauge, 1-1/2 Inch, 100s	IV

NOTE (2): ANSI and ISO standards are available from the American National Standards Institute, Inc., 11 West 42nd Street, New York, NY 10036.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5680

Location: c:\word\doc\levin\needlecd

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A A 54773A	15 March 1994

3. DOCUMENT TITLE
NEEDLE, HYPODERMIC, DISPOSABLE

4. NATURE OF CHANGE (*Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.*)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. [illegible]	b. ORGANIZATION
c. ADDRESS (*Include Zip Code*)	d. TELEPHONE (*Include Area Code*) (1) Commercial (2) AUTOVON (*If applicable*)
	e. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (*Include Area Code*)
Defénse Personnel Support Center ATTN: DPSC-MSE	(1) Commercial: 215-737-2870 (2) AUTOVON: 444-2870
c. ADDRESS (*Include Zip Code*) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete. — 1982

Source: https://assist.dla.mil - Downloaded: 2026-04-26T09:38Z
Check the source to verify that this is the current version before use.