METRIC

A-A-54775
30 November 1992

COMMERCIAL ITEM DESCRIPTION

DISH, CULTURE, PETRI (Borosilicate Glass)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specification NNN-D-1478B which is cancelled.

This Commercial Item Description covers borosilicate glass Petri dishes.

Salient characteristics:

Shall be uncompartmented borosilicate glass Petri dishes for use in performing bacteriological culture growth and biological assays.

Shall be Petri dishes of the following types, classes, and sizes (see "Ordering data"):

Type I - Blown type.
Type II - Pressed type.

Class 1 - Complete dish (top and bottom sections).
Class 2 - Top section only.
Class 3 - Bottom section only.

60 x 15 mm size
100 x 10 mm size
100 x 15 mm size
100 x 20 mm size
100 x 22 mm size
150 x 15 mm size
150 x 20 mm size

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Shall be Petri dishes fabricated of clear borosilicate glass conforming to the requirements of Type 1 and Class A or B in ASTM E438. The glass shall be capable of withstanding repeated sterilization, wet or dry, without loss of optical clarity. The glass shall be suitably annealed to conform to the requirements contained in ASTM E671.

Shall supply Petri dishes with the top and bottom sections having circular shapes with their sides approximately perpendicular to the circular surfaces. The top section covering for each size Petri dish shall be dimensioned to be loose fitting over the mating bottom section. The Petri dishes of any one size shall be of uniform production so that the tops and bottoms are interchangeable.

Shall supply Type I Petri dishes with reinforced beaded edges that provide increased resistance to mechanical breakage and facilitate the centering of each bottom section dish inside its top section covering. The surfaces of the Type I Petri dishes shall have no degree of curvature, convex, or concave, exceeding 1.6 mm.

Shall supply Type II Petri dishes with the bottom sections having flat inside bottom surfaces. Each of these bottom sections shall also include raised edges that provide a level three point support to protect the exterior bottom surface from being scratched. The outer surface (side) of each bottom section shall include a marking spot that is at least 120 square mm in area. The circular surfaces of the top and bottom sections for each Type II Petri dish shall have no degree of curvature, convex or concave, exceeding 0.5 mm.

Shall supply Type I and II Petri dishes with top and bottom sections having minimum wall thicknesses of 1 mm. The dimensions for each size Petri dish shall be as specified in Table I.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

TABLE I. Sizes and dimensions

Nominal Size mm	Minimum wall Thickness mm	TOP 1/ Inside Depth mm	TOP 1/ Inside Dia. mm	BOTTOM 1/ Outside Depth mm	BOTTOM 1/ Outside Dia. mm
60 x 15	0.75	14	57.5	16	57
100 x 10	0.75	9	95.5	11	95
100 x 15	0.75	14	95.5	16	95
100 x 20	0.75	19	95.5	21	95
100 x 22 2/	0.75	--	----	22	95
150 x 15	0.75	14	145.53/	16	1453/
150 x 20	0.75	19	145.53/	21	1453/

1/ Tolerances: Depth, top - 2 mm, bottom + 2 mm. Diameter, top + 3 mm, bottom - 3 mm.

2/ Type II, class 3.

3/ As alternate, inside/outside diameter dimensions of 144.5/144 mm acceptable.
Shall supply Class 1,2, and 3 Petri dishes with legible and permanent markings that include the following:

Each top and bottom section contains the name or registered trademark of the manufacturer or supplier.

The top and bottom sections are distinguished from each other by a suitable marking method that facilitates error-free sorting and permits easy identification of the tops and bottoms when assembling pairs.

Workmanship. The Petri dishes shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing 12 Petri dishes of one type, class and size only, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and testing. Inspections and testing shall be conducted on the Petri dishes in accordance with commercial practice to determine compliance with specification requirements. Sampling for the inspections and testing shall be conducted in accordance with commercial practice.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Contractor Certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Other publications. The following documents form a part of this specification to the extent that is specified herein. Unless otherwise indicated, the effect on date of invitation for bids or request for proposal shall apply.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

ASTM

C148 Standard Specification for Polariscope Examination of Glass Containers.

E438 Standard Specification for Glasses in Laboratory Apparatus.

E671 Standard specification for Maximum Permissible Thermal Residual Stress in Annealed Glass Laboratory Apparatus.

(Application for copies should be addressed to ASTM, 1916 Race Street, Philadelphia, PA 19103.)

Ordering data. Type, class, and size Petri dish required, unit/intermediate/exterior package quantities, labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

		SPECIFICATION DESIGNATIONS:		
NSN	Item Identification	TYPE	CLASS	SIZE
6640-01-351-0065	DISH, CULTURE, PETRI, Top and Bottom, Complete, 12s.	I	1	100 x 15 mm

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6640-1803

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54775	30 NOVEMBER 1992

3. DOCUMENT TITLE
DISH, CULTURE, PETRI (Borosilicate Glass)

4. NATURE OF CHANGE (*Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.*)

5. REASON FOR RECOMMENDATION

6. SUBMITTER	
a. [illegible]	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY	
a. NAME Defense Personnel Support Center	b. TELEPHONE (Include Area Code) (1) Commercial (215) 737-2870 (2) AUTOVON 444-2870
ADDRESS (Include Zip Code) ATTN: MSE 2800 S. 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.