INCH-POUND

A-A-54778
15 January 1993

COMMERCIAL ITEM DESCRIPTION

ANALYZER, CLINICAL CHEMISTRY, DRY

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a solid state clinical chemistry analyzer employing the reagent format methodology.

Salient characteristics:

The analyzer shall be a solid state clinical chemistry analyzer employing the dry reagent format methodology. A dry reagent format is a system wherein all reagents required for reaction and measurement are totally self-contained and are activated when the specimen sample is placed onto a test slide, pack, or tablet.

The analyzer shall be capable of performing a full range of colorimetric, potentiometric, and kinetic chemistry methodologies.

The unit shall also be capable of performing discrete, batch, or STAT testing.

Shall have a minimum throughput of 50 tests per hour.

A pipette(s) or equivalent device for patient sample delivery shall be provided.

The analyzer shall have an integral printer that provides a hard copy of test results and calibration parameters, and a keyboard for data entry, patient ID, and operating commands.

AMSC N/A FSC 6630

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

The clinical chemistry analyzer shall have the capability to display status messages that guide in the proper operation of the analyzer and warning messages that alert the operator to errors or oversights.

A system that warns the operator of any unacceptable results shall be provided.

Specimen sample volume requirements for each test shall not exceed 12 microliters.

The analyzer shall perform, as a minimum, the following tests:

Test	Range
Albumin	3-5 g/dL
Amylase	5-900 U/L
Creatinine	0.1-15 mg/dL
Glucose	20-450 mg/dL
Urea Nitrogen	1-100 mg/dL
Phosphorus	0.5-13 mg/dL
Total Bilirubin	0.1-20 mg/dL
Total Protein	2-11 g/dL
Aspartate Aminotransferase (AST)	4-950 U/L
Alanine Aminotransferase (ALT)	3-950 U/L
Creatine Kinase (CK)	20-1600 U/L
Calcium	3-15 mg/dL
Sodium	95-215 mmol/L
Potassium	1-11 mmol/L
Chloride	65-140 mmol/L
Carbon Dioxide	5-50 mmol/L

Precision and accuracy of the test procedures shall be in keeping with nationally accepted clinical laboratory accreditation standards.

The dimensions of the unit, or modules that make up the unit, shall not exceed 42 inches in width, 14 inches in depth, and 7 inches in height.

The weight of the unit, or modules that make up the unit, shall not exceed 55 lbs.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

The analyzer shall be provided with standard RS232C serial port to permit computer interface.

The clinical chemistry analyzer shall use 120V, 60 Hz electric power unless otherwise specified in the contract or purchase order.

Workmanship. The clinical chemistry analyzer shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One analyzer, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one analyzer.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

6630-01-309-4513

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA - OSS

Project Number: 6630-0528

Location: ANLZR14C.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE. This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:

1 DOCUMENT NUMBER: A-A-54778

2 DOCUMENT DATE (YYMMDD): 15 January 1993

3 DOCUMENT TITLE: ANALYZER, CLINICAL CHEMISTRY, DRY

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a. NAME (Last, First, Middle Initial)

b. ORGANIZATION

c. ADDRESS (Include Zip Code)

d. TELEPHONE (Include Area Code)
(1) Commercial
(2) AUTOVON

7. DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a. NAME
DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DPSC-MSE

b. TELEPHONE (Include Area Code)
(1) Commercial: AC 215-737-2870
(2) AUTOVON: AV 444-2870

c. ADDRESS (Include Zip Code)
2800 SO. 20TH STREET
PHILADELPHIA, PA 19145

IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT
Defense Quality and Standardization Office
5203 Leesburg Pike Suite 1403 Falls Church VA 22041-3466
Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 98/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.