REVIEW ACTIVITIES: ARMY- MD
NAVY - AS, OS, YD
USER ACTIVITIES: ARMY - ME, PS
NAVY - MD

FED. SUP CLASS
6640



PART NUMBER	CAPACITY (ml)	STOPPER SIZE ACCOMMODATED
MS36066-1	125	5
MS36066-2	250	6
MS36066-3	500*	7
MS36066-4	500	7
MS36066-5	1000	8
MS36066-6	4000	12

*Part Number MS36066-3 does not have a tubulation.

MATERIAL: Borosilicate glass in accordance with Federal Specification DD-G-541, Type I, Class A.

(A) DESIGN: Flasks shall be in accordance with Federal Specification NNN-F-240, Type V.

Mouths of the flasks shall be designed to accommodate standard rubber stoppers as specified in Federal Specification ZZ-S-751.

Flasks shall be of heavy walled construction to provide necessary mechanical strength.

(A) NOTES:

1. Procurement document for these items with the exception of MS36066-6, shall be Federal Specification NNN-F-240.
2. Flasks are not supplied with stoppers.
3. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements as specified herein.
4. For design feature purposes this standard takes precedence over procurement documents referenced herein.
5. Referenced documents shall be of the issue in effect on date of invitations for bid.

APPROVED 29 Jun 1960 REVISED (A) 15 Aug 1969

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. MU Other Cust MS 82	TITLE FLASK, FILTERING	MILITARY STANDARD MS36066	
PROCUREMENT SPECIFICATION See Note 1	SUPERSEDES:	SHEET 1 OF 1	

DD FORM 1 SEP 67 672-1 (Coordinated)

PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE.

☆ U. S. GOVERNMENT PRINTING OFFICE: 1969-393-845/5-1514

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:33Z
Check the source to verify that this is the current version before use.