FED. SUP CLASS
6640



PART NUMBER	CAPACITY ml	DIMENSIONAL DATA mm	
		LENGTH	OUTSIDE DIAMETER
MS36165-1	100	195-203	36.00 to 37.75

MATERIAL: Borosilicate glass in accordance with Federal Specification DD-G-541, Type 1, Class A.

DESIGN: Centrifuge tube shall be in accordance with Federal Specification NNN-T-780, Type 1 and in conformance with ASTM D91, D96, D1347, and D1796.

Centrifuge tube shall be tapered to a smaller diameter cylindrical stem, with a round bottom.

Top or open end of the tube shall have a beaded rim.

Graduation intervals shall be:		Capacity tolerance shall be:	
0.05 ml	0 to 0.5 ml	0 to 0.1 ml	± 0.02 ml
0.10 ml	0.5 to 2.0 ml	0.1 to 0.3 ml	± 0.03 ml
0.20 ml	2.0 to 3.0 ml	0.3 to 0.5 ml	± 0.05 ml
0.5 ml	3.0 to 5.0 ml	0.5 to 1.0 ml	± 0.05 ml
1.0 ml	5.0 to 10.0 ml	1.0 to 2.0 ml	± 0.10 ml
5.0 ml	10.0 to 25.0 ml	2.0 to 3.0 ml	± 0.10 ml
25.0 ml	25.0 to 100 ml	3.0 to 5.0 ml	± 0.20 ml
		5.0 to 10.0 ml	± 0.50 ml
		10.0 to 25.0 ml	± 1.00 ml
		25.0 to 100.0 ml	± 1.00 ml

Main graduations shall be numbered at 1/2, 1, 2, 3, 4, 5, 6, 8, 10, 15, 20, 25, 50, 75, and 100 ml.

Graduation markings and numbers shall be clearly legible and shall remain so under ordinary usage.

NOTES:
1. For design feature purposes this standard takes precedence over procurement document referenced herein.
2. Referenced documents shall be of the issue in effect on date of invitations for bid.

(A) ENTIRE STANDARD REVISED

APPROVED 30 DEC 1960 REVISED (A) 20 June 1969

USER ACTIVITIES: ARMY - EL, MF

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive uses shall be made from this document.

P.A. MU	TITLE	MILITARY STANDARD
Other Cust MS	TUBE, CENTRIFUGE oil type	MS36165
PROCUREMENT SPECIFICATION NNN-T-780	SUPERSEDES	SHEET 1 OF 1

DD FORM 1 SEP 67 672-1 (Condensed)

PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE.

☆ U. S. GOVERNMENT PRINTING OFFICE: 1969-353-663/3-46

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:35Z
Check the source to verify that this is the current version before use.