FED. SUP CLASS
6640

USER ACTIVITIES: ARMY - EL, ME



PART NUMBER	OUTSIDE DIAMETER ± 0.5 mm	LENGTH ± 2.5 mm	WALL THICKNESS ± 0.25 mm
MS36183-1	Canceled Inactive		
MS36183-2	Canceled Inactive		
MS36183-3	18	150	1.75
MS36183-4	25	200	1.75

MATERIAL: Borosilicate glass in accordance with Federal Specification DD-G-541, Type I, Class A.

DESIGN: Test tubes shall be in accordance with Federal Specification NNN-T-189, Type I, Style II.

Test tubes shall be cylindrical with one end having a rounded bottom and the top or open end having a lipless, fire-polished edge.

NOTES:
1. For design feature purposes this standard takes precedence over procurement document referenced herein.
2. Referenced documents shall be of the issue in effect on date of invitations for bid.

APPROVED 30 DEC 1960 REVISED (A) 20 June 1969

(A) ENTIRE STANDARD REVISED

P.A. MU Other Cust MS	TITLE TEST TUBE ignition, heavy wall		MILITARY STANDARD MS36183
Procurement Specification NNN-T-189	SUPERSEDES:		SHEET 1 OF 1

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

DD FORM 1 SEP 63 672 (Coordinated)

EDITION OF 1 SEP 57 MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:35Z
Check the source to verify that this is the current version before use.