MILITARY SPECIFICATION

CLAMP, ARTERY, BULLDOG, GLOVER

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers artery clamps, of the Glover design, for accomplishing hemostasis in surgical procedures.

1.2 Classification. Clamps shall be furnished in the following types and sizes as indicated below:

Type	Size	Centimeters
1 - Straight jaws	A	9
2 - Curved jaws	B	6.5

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

FEDERAL

GG-I-526 - Instruments, Dental and Surgical; General Specifications.
PPP-B-566 - Boxes, Folding, Paperboard.
PPP-B-585 - Boxes, Wood, Wirebound.
PPP-B-601 - Boxes, Wood, Cleated-Plywood.
PPP-B-621 - Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636 - Boxes, Shipping, Fiberboard.
PPP-B-676 - Boxes, Set-Up.

MILITARY

MIL-P-116 - Preservation, Methods of.
MIL-L-10547 - Liners, Case, and Sheet Overwrap; Water-Vaproproof, or Waterproof, Flexible.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Hq, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, Code ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:58Z
Check the source to verify that this is the currentversion before use.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.

MIL-STD-129 - Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC) Drawing

No. 20714 CLAMP, ARTERY, BULLDOG, GLOVER, STRAIGHT AND CURVED JAWS.

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM E 18 - Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials.

ASTM E 92 - Vickers Hardness of Metallic Materials, Test for.

ASTM E 140 - Standard Hardness Conversion Tables for Metals.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.)

(Technical society and technical association specifications and standards are generally available for reference from libraries. They are also distributed among technical groups and using Federal agencies.)

3. REQUIREMENTS

3.1 General requirements. The clamps shall be in accordance with the applicable requirements of GG-I-526 and as specified herein.

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:58Z
Check the source to verify that this is the currentversion before use.

3.2 Material. The clamps shall be fabricated from class 4, type 410, 416 or 420 material. Adjusting screw shall be fabricated from class 3, type 302, 303 or 304 material. Spring shall be fabricated from class 3, type 302 or 303, bright finish, spring temper wire.

3.3 Hardness. After proper stress relieving, Rockwell hardness of type 410 and 416 material shall be C36 to C41, and C40 to C45 for type 420 material. Rockwell hardness of the adjusting screw shall be C25 to C30. Hardness shall be tested as specified in 4.4.1. The requirements of GG-I-526 shall apply.

3.4 Style, design and dimensions. Shall be a Bulldog artery clamp of the Glover design. Tension of the clamp shall be regulated by means of an integral adjusting screw and spring at the proximal end of the clamp. One jaw shall be provided with two rows of well defined teeth throughout the entire length of the jaw. The other jaw shall possess a single row of well defined teeth. The depth of all teeth shall be consistent. Clamps specified herein shall be of two types and two sizes. Type 1 clamps shall have straight jaws and type 2 clamps shall have curved jaws. Both types shall be of size A (nominal overall length of 9 centimeters) or size B (nominal overall length of 6.5 centimeters). All clamps shall provide a firm, non-slip, triangular grasp, for accomplishing hemostasis in delicate tissue, without causing trauma to the tissue. The clamps shall be in strict accordance with DPSC Drawing Number 20714, for appropriate type, size, stock number and item identification.

3.5 Finish. Finish of the clamps proper (excluding spring and adjusting screw) shall be satin.

3.5.1 Passivation. All material shall be properly passivated.

3.6 Identification marking. In addition of GG-I-526 requirements, clamps of foreign manufacture shall also be marked "stainless".

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specifications where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Shall be in accordance with GG-I-526.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:58Z
Check the source to verify that this is the current version before use.

4.1.2 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the corrosion-resisting steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number (NSN), the item identification, the name of the component and material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as indicated in table I. The unit of product for examination purposes shall be one clamp.

TABLE I. Sampling criteria for examination.

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in table II. The unit of product for test purposes shall be one clamp. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

TABLE II. Sampling criteria for tests.

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL
Clamp	Hardness	3.3 & GG-I-526	4.4.1	S-2	1.5
	Dimensions	3.4	-----	S-2	2.5
	Copper sulfate test (for class 4 material)	GG-I-526	GG-I-526	S-2	1.0
	Boil test (for all material)	GG-I-526	GG-I-526	S-2	1.0

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:58Z
Check the source to verify that this is the current version before use.

4.2.3 Raw materials. A sufficient quantity of each lot of each metal used in the mnaufacture of the clamps shall be selected for chemical analysis.

4.3 Examination.

4.3.1 Examination of clamps. The clamps shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in tables I and III. Examination, however, shall not be restricted to defects listed in table III. Any clamps in the sample containing one or more defects shall be rejected, and if the number of defective clamps in any sample exceeds the acceptance number for the sample, the lot represented by that sample shall be rejected.

TABLE III. Classification of defects

Categories	Defects
Critical	None defined
Major	
101	Clamp not of the specified type and size for appropriate stock number.
102	Teeth not well defined.
103	Teeth not of uniform depth.
104	Teeth do not mesh when clamp is opened and closed.
105	Adjusting screw and spring not capable of easy tension adjustment by hand force.
106	Clamps not free from cracks, pits, burrs, sharp edges, rough areas, tool marks, scale and corrosion.
107	Finish not as specified.
108	Adjusting screw and spring not adequately secured to clamp.
109	Serrations at proximal end missing, incomplete, or not well defined.
Minor	
201	Identification markings not present, not correct, not complete or not legible.

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:58Z
Check the source to verify that this is the current version before use.

4.3.2 Packaging inspection. The inspection of the preservation-packaging and interior package marking shall be in accordance with the group A and B quality conformance inspection requirements of MIL-P-116. The inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Hardness test. Hardness shall be determined in accordance with ASTM Std. Method E18 or E92. Conversion of Vickers hardness values to Rockwell values shall be in accordance with ASTM Std. Hardness Conversion Tables E140.

5. PACKAGING

5.1 Preservation-packaging. Preservation-packaging shall be level A or C as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Cleaning, drying and preservative. Each clamp of the type and size specified shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116. Preservative shall not be applied.

5.1.1.2 Unit packaging. Each clamp of the type and size specified shall be wrapped in neutral tissue or other suitable anticorrosive material. Each clamp shall be further packaged in accordance with MIL-P-116, method 1C-1.

5.1.1.3 Intermediate packaging. Twelve (12) clamps of the type and size specified shall be packaged in a box of appropriate size and design constructed in accordance with PPP-B-566, PPP-B-676 or PPP-B-636, type CF, class domestic. Closure shall be in accordance with the appendix of the applicable box specification.

5.1.2 Level C. Clamps shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.1).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:58Z
Check the source to verify that this is the current version before use.

5.2.1 Level A. Two hundred eighty eight (288) clamps of the type and size specified shall be packed in a box designed for a type 2 load, and constructed in accordance wth PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2 or PPP-B-636, class weather resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof in accordance with the appendix of PPP-B-636.

5.2.2 Level B. Two hundred eighty eight (288) clamps of the type and size specified shall be packed in a box designed for a type 2 load, and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, type CF, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Unit, intermediate and exterior container. Each unit, intermediate and exterior container sall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents shall specify the following:

(a) Number, title and date of this specification.
(b) Type and size required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(d) Applicable National Stock Number.

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:58Z
Check the source to verify that this is the currenversion before use.

6.2 This specification covers the following items appearing in the Federal Supply Catalog:

National Stock Number	Item Identification	Specification Designation Type	Size
6515-00-690-3215	CLAMP, ARTERY, Bulldog, Glover, Straight Jaws, 9 cm, 4 cm JAW	1	A
6515-00-690-3216	CLAMP, ARTERY, Bulldog, Glover, Curved Jaws, 9 cm, 4 cm Jaw	2	A
6515-00-890-1681	CLAMP, ARTERY, Bulldog, Glover, Straight Jaws, 6.5 cm, 2.7 cm Jaw	1	B
6515-00-890-1682	CLAMP, ARTERY, Bulldog, Glover, Curved Jaws, 6.5 cm, 2.7 cm Jaw	2	B

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-2019

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:58Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-01-21T10:58Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION (Mark one)
	☐ VENDOR
b. ADDRESS (Street, City, State, ZIP Code)	☐ USER
	☐ MANUFACTURER
	☐ OTHER (Specify): _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER (Last, First, MI) – Optional	b. WORK TELEPHONE NUMBER (Include Area Code) – Optional
c. MAILING ADDRESS (Street, City, State, ZIP Code) – Optional	8. DATE OF SUBMISSION (YYMMDD)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil - Downloaded: 2025-01-21 10:58Z
Check the source to verify that this is the current version before use.