MILITARY SPECIFICATION

DESTRUCTION UNIT, SYRINGE-HYPODERMIC NEEDLE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1.1 SCOPE

1.2 Scope. This specification covers a small, portable, self-contained device suitable for safe destruction of hypodermic syringes and hypodermic needles.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-B-566 - Boxes, Folding, Paperboard.
PPP-B-585 - Boxes, Wood, Wirebound.
PPP-B-601 - Boxes, Wood, Cleated-Plywood.
PPP-B-621 - Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636 - Boxes, Shipping, Fiberboard.
PPP-B-676 - Boxes, Set-Up.

MILITARY

MIL-L-10547 - Liners, Case and Sheet, Overwrap Water-Vaporproof, or Waterproof, Flexible.
MIL-N-36124 - Needles, Hypodermic, Disposable.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Hq, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document, or by letter.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:55Z
Check the source to verify that this is the current version before use.

STANDARDS

FEDERAL

FED-STD-66	- Steel, Chemical Composition and Hardenability.
FED-TEST METHOD-STD-151	- Metals, Test Methods.

MILITARY

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer).

3. REQUIREMENTS

3.1 First article. When specified in the procurement document, the contractor shall furnish sample unit(s) for first article inspection and approval (see 4.1.2.1 and 6.2). The first article shall function properly after being tested as specified in 4.4.2.1).

3.2 Materials.

3.2.1 Corrosion-resisting steel. Corrosion-resisting steel shall comply with FED-STD-66 when tested as specified in 4.4.1.

3.2.1.1 Shearing mechanism. The shearing mechanism shall be fabricated of AISI 300 series of FED-STD-66, except as specified below.

3.2.1.2 Blade and needle aperture bushing. The blade and the needle aperture bushing shall be fabricated of AISI 400 series of FED-STD-66 and shall have a Rockwell Hardness of at least 86.5 on the 15-N scale.

3.2.2 Plastic. The case, cover, and handles shall be fabricated of polypropylene plastic. Color shall be red or orange for warning indication.

3.3 Design and construction. The destruction unit shall consist of a plastic case with a removable cover. The case shall contain a paperboard cup for collecting the sheared-off needles and syringe tips. The cover shall include two apertures and a shearing device. The complete unit shall weigh not more than one pound. The unit shall be capable of withstanding repeated steam autoclaving at 250° F for 20 minutes without showing any evidence of failure or deterioration. The unit shall perform as required after completion of the repeated steam autoclaving.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:55Z
Check the source to verify that this is the current version before use.

3.3.1 Case. The case shall be constructed of polypropylene plastic, at least .050 inch thick. It shall be square or rectangular in shape, with each side not less than 3½ inches nor more than 5 inches long and not less than 3 inches nor more than 5 inches high. The sides shall be straight or tapered inward toward the bottom.

3.3.1.1 Cup. The case shall contain a disposable paperboard cup of standard commercial design with a capacity of 6 ounces or more. The cup shall be retained in a location such that needles and syringe tips shall fall into it when clipped-off by the shearing mechanism.

3.3.2 Cover. The cover shall be constructed of the same plastic as the case, at least .050 inch thick, and shall lock snugly onto the case. The cover shall include the shearing mechanism for the needles and syringe tips.

3.3.2.1 Shearing mechanism. The shearing mechanism shall be constructed of corrosion-resisting steel and shall include two apertures: a needle aperture, with a bushing which will accept needles of 18 gage or smaller, and an aperture 3/8 ± 1/16 inch in diameter for the insertion of syringe tips. It shall also include a blade and two plastic handles.

3.3.2.1.1 Blade. A corrosion-resisting steel blade shall be attached securely to the shearing mechanism at the pivot point by a permanent rivet. As an alternate, the blade may be detachable for sharpening or replacement. The blade shall have appropriate cutting edges for needles and syringe tips. The blade shall be operated by squeezing together a plastic handle attached to the blade and a fixed plastic handle attached to the mechanism. A corrosion-resisting steel spring shall return the blade to its open position. The shearing mechanism shall be capable of destroying 10,000 needles and 10,000 syringe tips when tested as specified in 4.4.2. The spring shall function properly throughout the testing specified in 4.4.2.

3.4 Marking. Each unit shall be marked with the following information. The marking shall be clear, permanent, and legible. It shall be capable of withstanding repeated autoclaving, at 250° F for 20 minutes, without deteriorating or becoming detached from the unit.

3.4.1 Identification. Identification shall consist of the contract or purchase order number and the manufacturer's name or registered trademark.

3.4.1.1 Operating directions. Directions, marked permanently and legibly on the top of the unit, shall indicate clearly and conspicuously the needle aperture, the plastic barrel aperture, and the sequence of cutting. Appropriate cleaning and operating instructions may also be present upon the unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:55Z
Check the source to verify that this is the current version before use.

3.5 Service data. The contractor shall include, with each unit, complete instructions for operating and cleaning. Instructions shall be on a separate sheet, printed on the unit package, or printed on a label on the unit itself.

3.6 Workmanship. Workmanship shall be first class throughout. Item shall be free from defects which detract from its appearance or may impair its serviceability. All flash, burrs, and sharp edges shall be removed from the item.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2.1 Classification of inspection. The inspection requirements specified herein are classified as follows:

a. First article inspection.
b. Quality conformance inspection.

4.1.2.1.1 First article inspection. First article inspection shall be performed on the destruction unit when a first article sample is required (see 3.1). This inspection shall include examination of 4.3 and tests of 4.4.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:55Z
Check the source to verify that this is the current version before use.

4.1.2.1.2 Quality conformance inspection. Quality conformance inspection shall include the examination of 4.3 and tests of 4.4 with the exception of 4.4.2.1. The Government reserves the right to perform the test specified in 4.4.2.1 to determine compliance of a product and or a particular lot with the requirement of 3.3.2.1.1.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials may be furnished in lieu of actual performance of such testing by the contractor, provided lot identy has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I. The unit of product for sampling purposes shall be one destruction unit.

TABLE I. Sampling for examination.

	Inspection level	AQL (percent defective)
For visual examination		
Major defects	II	1.0
Total defects (Major and Minor combined)	II	4.0
For dimensional examination	S-1	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II for an acceptance number of zero. The unit of product for sampling purposes shall be one destruction unit.

TABLE II. Sampling for tests.

Component (Unit of product)	Characteristic	Requirement	Test procedure	Inspection level
Shearing mechanism	Chemical Composition	3.2.1	4.4.1	See 4.2.3
	Extended operation	3.3.2.1.1	4.4.2.1	First Article only

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:55Z
Check the source to verify that this is the current version before use.

4.2.3 Raw materials. A sufficient quantity of each lot of each metal used in the manufacture of the destruction unit shall be selected for chemical analysis.

4.3 Examination. The destruction unit shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in Table II. Examination shall not be restricted to the classified possible defects listed below.

TABLE II. Classification of defects.

Categories	Defects
Major	
101	Any component part missing or not usable.
102	Components not properly aligned.
103	Surface torn, cracked, ruptured, or deformed.
104	Exposed or accessible edge sharp.
105	Exposed edge has flash or burrs.
106	Shearing mechanism does not function properly.
107	Cover does not include specified directions printed legibly.
Minor	
201	Exposed surface deformed, marred, or scratched.
202	Color not as specified.
203	Unit not clean, not free of grease, oil or foreign matter.
204	Identification marking missing or illegible.

4.3.2 Dimensional examination. The destruction unit shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.3.3 Packaging inspection. The inspection of packaging, packing, and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:55Z
Check the source to verify that this is the current version before use.

4.4.1 Chemical composition. Four ounces from each lot of metal shall be tested in accordance with FED TEST METHOD STD 151.

4.4.2 First article test .

4.4.2.1 Extended operation. When specified (see 3.1) the destruction unit shall be steam autoclaved three times at 250° F for 20 minutes each. Corrosion-resisting steel tubing conforming to the type specified for disposable hypodermic needles in MIL-N-36124 shall then be fed, by any convenient means, into the needle aperture and clipped off. Tubing shall consist of that specified for needle gages 18 through 25, with at least 25% consisting of gage 18. The clipping may be accomplished manually or by means of a suitable mechanical arrangement. A minimum of 10,000 clippings shall be made. A similar number of clippings shall be made of plastic tubing, of the type specified in MIL-N-36124 for hubs, inserted into the plastic-barrel aperture of the unit.

5. PACKAGING.

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit packaging. Each destruction unit shall be packaged in a box of appropriate size constructed in accordance with PPP-B-566 (except thickness of board shall be .016, minimum) PPP-B-676, or PPP-B-636, type CF, class domestic. Box shall be closed as specified in the appendix of the applicable box specification.

5.1.1.2 Intermediate package. Ten (10) destruction units shall be packed in a snug fitting box constructed in accordance with PPP-B-566, PPP-B-636, class domestic or PPP-B-676. Closure of the box shall be as specified in the appendix of the applicable box specification.

5.1.2 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be Level A, B or C as specified (see 6.2).

5.2.1 Level A. Forty (40) destruction units shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:55Z
Check the source to verify that this is the current version before use.

5.2.1.1 Case liner. Each level A wood box shall be lined with a water-proof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in the appendix of PPP-B-636.

5.2.2 Level B. Forty (40) destruction units shall be packaged in an exterior container designed for a type 2 load and contructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II as specified in PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking. Each unit, intermediate package and exterior container shall be marked as specified in MIL-STD-129.

6. NOTES

6.1 Intended use. The destruction unit is intended for the safe destruction of hypodermic syringes and hypodermic needles in military medical facilities.

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents.

a. Title, number and date of this specification.
b. Selection of applicable levels of packaging and packing, (See 5.1 and 5.2).
c. When a first article is required (see 3.1, 4.1.2.1 and 6.3)

6.3 First article. When a first article is required, it shall be tested and approved under the appropriate provisions of 7-104.55 of Armed Services Procurement Regulation. The contract or purchase order specifies the quantity, procedure for delivery and other specific instructions for inspection and approval of the first article.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:55Z
Check the source to verify that this is the current version before use.

MIL-D-37528

6.4 This specification covers the following item(s) appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6530-00-152-0047	DESTRUCTION UNIT, SYRINGE-HYPODERMIC NEEDLE, PORTABLE

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6530-1399

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:55Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
DESTRUCTION UNIT, SYRINGE-HYPODERMIC NEEDLE MIL-D-37528

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:55Z
Check the source to verify that this is the current version before use.