MILITARY SPECIFICATION

LIGHT, DESK, WARD

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a small, high-intensity desk light for use in hospital wards.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-B-636 - Boxes, Shipping, Fiberboard.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

UNDERWRITERS' LABORATORIES, INC.

UL 153 - Portable Electric Lamps.
UL 1262 - Laboratory Equipment.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6230

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:59Z
Check the source to verify that this is the current version before use.

(Application for copies should be addressed to the Underwriters' Laboratories, Inc., 207 East Ohio Street, Chicago, IL 60611, 333 Pfingsten Road, Northbrook, IL 60062, 1285 Walt Whitman Road, Melville, LI, NY 11746, 1655 Scott Blvd., Santa Clara, CA 95050.)

3. REQUIREMENTS

3.1 First article. When specified in the procurement document, the contractor shall furnish sample unit(s) for first article inspection and approval (see 4.1.2.1 and 6.2).

3.2 Fire and casualty hazards.

3.2.1 The contractor shall submit proof, prior to delivery, that the items to be delivered under the contract conform to the applicable requirements of Underwriters' Laboratories, Inc. (UL) Standard of Safety UL 153. Proof shall consist of one of the following:

A. UL label affixed to the product, provided the contractor agrees to have Underwriters' Laboratories, Inc. forward a copy of the UL report directly to the Contracting Officer, or the Government Inspector, upon the latter's request.

B. Laboratory report from a nationally recognized independent laboratory. If an independent laboratory is used, it must be adequately equipped and competent to determine compliance with UL standards. The Government reserves the right to require a facilities report and to conduct a facilities survey. Any laboratory report submitted shall include actual test methods, data taken, and results of tests. Detailed findings of examinations necessary to determine compliance shall also be included. The test report shall also have a section that describes the test sample in such a way that the report can be used to insure identity of production items with the test sample. This section shall be supplemented with construction information such as materials used, spacings between electrical components, types of bushings, photographs, wiring diagrams, sketches, and parts list, all appropriately identified. Compliance with this paragraph does not excuse compliance with all other requirements of the contract and specifications.

3.3 Materials. All material used in the construction of the lights shall be of the highest quality throughout and suitable in all respects for the purpose intended.

3.4 Design and construction. The light shall consist of a base and a lamphead, connected by an adjustable, articulated arm, and a power cord and plug.

3.4.1 Base. The base shall have maximum dimensions of 3 inches wide, 5 inches deep and 2-1/2 inches high. It shall contain a three-position switch with a metal "hi-lo-off" indicator plate, and a transformer. The base shall be fabricated of aluminum, covered with a black, brown, or walnut-colored vinyl laminate, and plated steel. The bottom of the base shall have a slotted hole for wall mounting. The base shall rest on four raised metal nipples or on four felt foot pads. The base shall remain stable on a level surface when the outer edge of the reflector is 12 inches in front of the forward edge of the base (with the reflector facing downward).

Source: https://assist.dla.mil -- Downloaded: 2026l0-21T10:59Z
Check the source to verify that this is the currentversion before use.

3.4.2 Arm assembly. The arm assembly shall be fabricated from plated, rectangular, steel tubing at least .031 inch thick (No. 22, Manufacturers' Standard gage). Tubing shall be not less than 1/4 inch nor more than 1/2 inch on each side. The arm shall consist of two sections: a lower arm 6-1/2 ± 3/4 inches long and an upper arm 8-1/2 ± 3/4 inches long (measured between pivot centers). The joints connecting the lower arm to the base and the upper arm to the lower arm shall each allow rotation through an arc of at least 150° in the vertical plane. The lamphead shall be connected to the upper arm by a joint that allows the lamphead to rotate through an arc of at least 90° in the vertical plane, and at least 45° to the right, and 45° to the left, of the center axis.

3.4.3 Lamphead. The lamphead shall consist of the reflector, the lamphead (socket) and the lamp.

3.4.3.1 Reflector. The reflector shall be not more than 4-1/2 inches in diameter nor more than 3-1/2 inches from front to back. The reflector shall be made of aluminum, not less than .03 inch thick.

3.4.3.2 Lamp. The light shall be supplied with a lamp and two spares. The lamp shall be an incandescent, single-contact, bayonet type.

3.4.4 Power cord. The light shall be supplied with a two-conductor power cord (an optional grounding conductor may be added), No. 18 American Wire gage or larger, terminating in a molded, two or three-prong plug. The cord shall extend at least 5 feet from the base of the light. The cord shall enter the base through a strain-relief bushing. The ungrounded conductor shall be connected to the switch so that the screw shell and the center contact of the socket shall be electrically dead when the switch is in the "off" position. The cord and plug shall conform to UL 153.

3.5 Finish.

3.5.1 Base. Exposed metal portions of the base shall be plated with a minimum thickness of .001 inch of brass, nickel, chromium, cadmium, or any combination thereof. The finish shall be protected by a coating of clear lacquer or epoxy.

3.5.2 Arms. The arms shall be plated with a minimum thickness of .0005 inch of brass, nickel, chromium, cadmium, or any combination thereof. The finish shall be further protected by a coating of clear lacquer or epoxy.

3.5.3 Reflector. The reflector shall be degreased and prepared for painting by etching or an equivalent process. The interior shall be painted with white, baked-on enamel, applied in accordance with the paint manufacturer's directions. The exterior shall be painted with brown or black, baked-on enamel, smooth or wrinkle finish, applied in accordance with the paint manufacturer's directions.

3.5.4 Hardware. All exposed hardware shall be plated to resist corrosion, or shall be made of corrosion-resisting metal.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:59Z
Check the source to verify that this is the current version before use.

3.6 Light intensity. The light intensity at any point within a 12-inch by 14-inch rectangular area shall be not less than 70 foot candles on the high-intensity setting nor 30 foot candles on the low intensity setting. The maximum intensity within the area should not exceed the minimum intensity by more than 3 to 1. Light intensity shall be measured as specified in 4.4.1.

3.7 Dielectric withstand. The light shall show no sign of dielectric breakdown when tested as specified in 4.4.2.

3.8 Leakage current. The leakage current of the light shall not be more than 0.5 milliamperes when tested as specified in 4.4.3.

3.9 Identification marking. Each light shall have a nameplate, securely affixed to the underside of the base, with the following information, as a minimum:

a. Manufacturer's name or registered trademark.
b. Manufacturer's model number.
c. Operating volts, cycles, and watts.

3.10 Workmanship. Workmanship shall be sound in all details and free from defects which affect the operation, durability, or appearance of the lamp. Adequate strain relief shall be provided throughout, effective in all positions of adjustment. Wire raceways shall be free of burrs and sharpness. All connections shall be firmly and securely made. Insulation shall be sound, with only essential stripping made for terminal connections. Stranded conductors shall not have loose strands at terminals.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:59Z
Check the source to verify that this is the currentversion before use.

4.1.2.1 Classification of inspection. The inspection requirements specified herein are classified as follows:

a. First article inspection.
b. Quality conformance inspection.

4.1.2.1.1 First article inspection. First article inspection shall be performed on the light when a first article sample is required (see 3.1). This inspection shall include examination of 4.3 and tests of 4.4.

4.1.2.1.2 Quality conformance inspection. Quality conformance inspection shall include the examination of 4.3 and tests of 4.4.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government and provided also that the test methods and the results obtained comply with those described in this specification. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I. The unit of product for sampling purposes shall be one light.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Total defects (Major and Minor combined)	II	4.0
For dimensional examination	S-1	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II, for an acceptance level of zero. The unit of product for sampling purposes shall be one light.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:59Z
Check the source to verify that this is the current version before use.

TABLE II. Sampling for tests.

Characteristic	Requirement	Test procedure	Inspection level
Light intensity	3.6	4.4.1	S-2
Dielectric withstand	3.7	4.4.2	100%
Leakage current	3.8	4.4.3	100%
Polarity	3.4.4	4.4.4	100%

4.3 Examination. The light shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with, but not restricted to, the classification of defects listed in table III.

TABLE III. Classification of defects.

Categories	Defects
Major	
101	Any component does not function as specified.
102	Extension arms not easily adjustable throughout their entire range.
103	Lamphead not easily adjustable throughout its entire range.
104	Light does not remain fixed when set in any position.
105	Base not stable on level surface when outer edge of reflector is extended 12 inches out from front of base.
106	Strain relief bushing does not hold cord securely in place.
107	Exposed metal surface bent, dented, or cracked.
108	Exposed surface has burrs or sharpness.
109 1/	Internal wiring bare of insulation.
110 1/	Wire raceway has burrs or sharp edges.
111 1/	Adequate strain relief not provided throughout.
Minor	
201	Identification marking not present, correct, legible, and securely affixed.
202	Surface scratched or marred.
203	Finish not as specified.
204	Two spare lamps not furnished.
205	Cord less than 5 feet in length.

1/ Major 109, 110, 111 should be examined just prior to final assembly to avoid damage to finished lights.

4.3.2 Dimensional examination. The light shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:59Z
Check the source to verify that this is the current version before use.

4.3.3 Packaging inspection. The inspection of the unit package, packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Light intensity. The light intensity should be tested in a darkened room, using a suitable photometer or reflectometer (6.5) which is accurate to within ± 5 percent. Ambient conditions shall conform to those specified by the manufacturer of the meter. Determination of light intensity within the 12 by 14 inch area shall be made without changing the adjustment of the light.

4.4.2 Dielectric withstand. The light shall be tested for dielectric withstand as specified in UL 153.

4.4.3 Leakage current. The light shall be tested for leakage current as specified in UL 1262.

4.4.4 Polarity. A continuity check shall be made to determine if the switch is wired directly to the ungrounded conductor as specified in UL 153.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Unit packaging. Each desk light shall be wrapped in kraft paper or other suitable non-corrosive wrapping material and packaged in a box conforming to PPP-B-636, class domestic. The light shall be cushioned within the box by use of fiberboard pads or fittings or other suitable material to prevent movement and damage during shipment and handling. The two spare lamps shall be wrapped and cushioned with tissue paper or other suitable material and packaged in a chip or paperboard box. The box shall be marked to identify contents and quantity and placed in the box containing the desk light. Closure of the unit package shall be as specified in the appendix of PPP-B-636.

5.1.2 Level C. Each desk light along with two spare lamps shall be packaged in a manner that will afford adequate protection against corrosion, deterioration and physical damage during shipment from the supply source to the first receiving activity.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:59Z
Check the source to verify that this is the currentversion before use.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.1).

5.2.1 Level A. Six unit packages shall be packed in a shipping container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.

5.2.2 Level B. Six unit packages shall be packed in a shipping container designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.2.3 Level C. The packaged lights shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the carrier's rules and regulations.

5.3 Marking. Each unit package and shipping container shall be marked as specified in MIL-STD-129.

6. NOTES

6.1 Intended use. The light is intended for use in hospital wards.

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents.

a. Title, number and date of this specification.
b. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
c. When a first article is required (see 3.1, 4.1.2.1 and 6.3).

6.3 First article. When a first article is required, it shall be tested and approved under the appropriate provisions of the Defense Acquisition Regulation/ Armed Services Procurement Regulation. The contract or purchase order specifies the quantity, procedure for delivery and other specific instructions for inspection and approval of the first article.

6.4 This specification covers the following item(s) appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6230-00-809-4284	LIGHT, DESK, WARD, 115 volt, 60 Hz

6.5 Suitable meters for the light intensity test include Baumgarten Reflectometer, Taylor Brightness Meter, Macbeth Illumination Meter, and Weston Illumination Meter.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activity:
DoD - MB

Preparing activity:
DLA-DM

Project No. 6230-0234

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:59Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS: This form is provided to solicit beneficial comments which may improve this document and enhance its use. DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity. A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed.
NOTE: This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

DOCUMENT IDENTIFIER *(Number)* AND TITLE
MIL-L-37560 LIGHT, DESK, WARD

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1. ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID, RESTRICTIVE, LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW.

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDED WORDING CHANGE

C. REASON FOR RECOMMENDED CHANGE(S)

2. REMARKS

SUBMITTED BY *(Printed or typed name and address — Optional)*

TELEPHONE NO.

DATE

DD FORM 1 OCT 76 1426 EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED.

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:59Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

POSTAGE AND FEES PAID



HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:59Z
Check the source to verify that this is the current version before use.