MILITARY SPECIFICATION

RETRACTOR, MASTOID, JANSEN, 3-VS-3 BLUNT PRONGS

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers one type of mastoid retractor, suitable for retraction of musculature surrounding the mastoid bone.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

FEDERAL

GG-I-526	- Instruments, Dental and Surgical; General Specifications.
PPP-B-585	- Boxes, Wood, Wirebound.
PPP-B-601	- Boxes, Wood, Cleated-Plywood.
PPP-B-621	- Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	- Boxes, Shipping, Fiberboard.
PPP-B-676	- Boxes, Set up.

MILITARY

MIL-P-116	- Preservation, Methods of.
MIL-L-10547	- Liners, Case, and Sheet Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Hq., Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, Code ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:58Z
Check the source to verify that this is the current version before use.

DRAWING

Defense Personnel Support Center (DPSC)

NO. 10110 RETRACTOR MASTOID, JANSEN, 3-VS-3 BLUNT PRONGS

(Copies of specifications, standards and drawing required by suppliers in connection with specific procurement functions should be obtained from procuring activities or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise specified, the issue in effect on date of invitation for bids or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM E 18 - Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials, Tests for.

ASTM E 92 - Vickers Hardness of Metallic Materials, Test for.

ASTM E 140 - Standard Hardness Conversion Tables for Metals.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.)

(Technical society and technical association specifications and standards are generally available for reference from libraries. They are also distributed among technical groups and using Federal agencies.)

3. REQUIREMENTS

3.1 General requirements. The retractors shall be in accordance with applicable requirements of GG-I-526, and as specified herein.

3.2 Material. The retractors shall be fabricated from class 4, type 420 material.

3.3 Hardness. After proper stress relieving, Rockwell hardness of the arms and handle shall be C 40 to C 47. The requirements of GG-I-526 shall apply.

3.4 Style, design and dimensions. Shall be a mastoid retractor, Jansen design, double-hooked, blunt teeth, toggle-handle design, suitable for retraction of musculature surrounding the mastoid bone during mastoidectomy surgical procedures. The arms of the retractor shall be interchangeable from left to right socket position of the handle. Shall be in accordance with DPSC Drawing 10110.

3.5 Finish. Finish shall be mirror, or mirror and satin.

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:58Z
Check the source to verify that this is the current version before use.

3.5.1 Passivation. All components of the retractor shall be properly passivated.

3.6 Identification marking. In addition to GG-I-526 requirements, retractors of foreign manufacture shall also be marked "stainless".

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, shall be in accordance with GG-I-526.

4.1.3 Certificates of quality. Certificates of quality supplied by the manufacturer of the corrosion-resisting steel may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, National Stock Number (NSN), the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as indicated in table I. The unit of product for examination purposes shall be one retractor.

TABLE I. Sampling for examination

Type defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:58Z
Check the source to verify that this is the currentversion before use.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in table II. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics. The unit of product for test purposes shall be one retractor.

TABLE II. Sampling criteria for tests

Component	Characteristic	Reqmt.	Test procedure	Inspection level	AQL
Retractor	Hardness	3.2 & GG-I-526	4.4.1	S-2	1.5
	Dimensions	3.4	-----	S-1	2.5
	Copper sulfate test	GG-I-526	GG-I-526	S-2	1.0
	Boil test	GG-I-526	GG-I-526	S-2	1.0

4.2.3 Raw materials. A sufficient quantity of each lot of metal used in the construction of the forceps shall be selected for chemical analysis.

4.3 Examination. The retractors shall be examined to determine compliance with all the requirements contained in this specification. Examination, however, shall not be restricted to defects listed in table III. Any retractors in the sample containing one or more defects shall be rejected, and if the number of defective retractors in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

TABLE III. Classification of defects

Categories	Defects
Critical	None defined.
Major	
101	Retractor not of the specified design.
102	One or more prongs possess sharp points.
103	Prongs not of uniform length and uniform thickness.
104	Handle threads not well defined.
105	Arms not interchangeable in any socket position of the handle.
106	Action of the retractor too loose or too tight.
107	Retractor not free from burrs, sharp edges, pits, cracks, holes, foreign matter, and corrosion.
108	Finish not uniform.
Minor	
201	Identification markings not present, not correct, not complete and not legible.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:58Z
Check the source to verify that this is the current version before use.

4.3.1 Packaging inspection. The inspection of the preservation-packaging and interior package marking shall be in accordance with the group A and B quality conformance inspection requirements of MIL-P-116. The inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Hardness test. Hardness test shall be determined in accordance with ASTM standard Method E18 or E92. Conversion of Vickers hardness values to Rockwell Hardness values shall be in accordance with ASTM Standard Hardness Conversion Tables E140.

5. PACKAGING

5.1 Preservation-packaging. Preservation-packaging shall be level A or C as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Cleaning drying and preservative. Each retractor shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116. Preservative shall not be applied.

5.1.1.2 Unit packaging. Each retractor shall be wrapped in neutral tissue or other suitable anticorrosive materials. Each retractor shall be further packaged in accordance with MIL-P-116, method 1C-1.

5.1.1.3 Intermediate packaging. Six (6) retractors shall be packaged in a box of appropriate size and design constructed in accordance with PPP-B-676 or PPP-B-636, type CF, class domestic. Closure shall be in accordance with the appendix or the applicable box specification.

5.1.2 Level C. Retractors shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.1).

5.2.1 Level A. Two hundred eighty-eight (288) retractors shall be packed in a box designed for a type 2 load, and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2 or PPP-B-636, class weather resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:58Z
Check the source to verify that this is the current version before use.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof in accordance with the appendix of PPP-B-636.

5.2.2 Level B. Two hundred eighty-eight (288) retractors shall be packed in a box designed for a type 2 load, and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, type CF, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Unit, intermediate and exterior container. Each unit, intermediate and exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title number, and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(c) National Stock Number.

6.2 National Stock Number coverage. This specification is intended to cover NSN 6515-00-361-3950 RETRACTOR, MASTOID, 3-VS-3 Blunt Prongs, as indicated in the Federal Supply Catalog. It does not include all types or sizes as indicated by the title nor all types which are commercially available.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-2016

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:58Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-R-37544 RETRACTOR, MASTOID, JANSEN, 3-VS-3 BLUNT PRONGS

NAME OF ORGANIZATION AND ADDRESS	CONTRACT NUMBER
	MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*	TELEPHONE NO.
	DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:58Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:58Z
Check the source to verify that this is the current version before use.