MIL-S-37545
17 August 1976

MILITARY SPECIFICATION

SCISSORS, GENERAL SURGICAL, STRAIGHT, 7 INCH

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a straight scissors suitable for general surgical operative procedures.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

FEDERAL

GG-I-526	- Instruments, Dental and Surgical; General Specifications.
PPP-B-585	- Boxes, Wood, Wirebound.
PPP-B-601	- Boxes, Wood, Cleated-Plywood.
PPP-B-621	- Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	- Boxes, Shipping, Fiberboard.

MILITARY

MIL-P-116	- Preservation, Methods of.
MIL-L-10547	- Liners, Case, and Sheet Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

MILITARY

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Hq., Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, Code ATT, 2800 S. 20th St., Philadelphia, PA 19101 by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:58Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM E 18 - Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials

ASTM E 92 - Vickers Hardness of Metallic Materials, Test for

ASTM E 140 - Standard Hardness Conversion Tables for Metals.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.)

(Technical society and technical association specifications and standards are generally available for reference from libraries. They are also distributed among technical groups and using Federal agencies.)

3. REQUIREMENTS

3.1 General requirements. The scissors shall be in accordance with the applicable requirements of GG-I-526 and as specified herein.

3.2 Material. The scissors shall be fabricated from Class 4, Type 420 or 420F material, except that the carbon content shall be 0.30 percent minimum. As an alternate, the chromium content shall be 12.00 to 15.00 percent, and the phosphorous content 0.045 percent maximum.

3.3 Hardness. After proper stress relieving, the scissors shall have a Rockwell hardness of C 50 to C 55 when tested as specified in 4.4.1. The requirements of GG-I-526 shall apply.

3.4 Style, design and dimensions. Shall be an operating scissors with straight blades. Points of each blade shall be blunt, suitable for general surgical procedures. The scissors shall be in accordance with Figure 1.


3.5 Finish. Finish shall be mirror, except for the inside of the blades which shall be crocus finish.


3.5.1 Passivation. The entire item shall be properly passivated.


3.6 Cutting edges. Scissors shall cut a continuous length of at least 1-3/4 inches of the test material, when tested as specified in 4.4.2.

3.7 Identification marking. In addition to GG-I-526 requirements, scissors of foreign manufacture shall also be marked "stainless".

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:58Z
Check the source to verify that this is the current version before use.







NOTES:

1. SCREW SHALL BE PERMANENTLY SECURED IN A FIXED POSITION AND SHALL NOT LOOSEN DURING NORMAL USE.
2. UNLESS OTHERWISE SPECIFIED TOLERANCES ON FRACTIONS — ± 1/32.

ALL DIMENSIONS IN INCHES

FIGURE 1. Scissors, general surgical, straight, 7 inch.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:58Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specifications where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Shall be in accordance with GG-I-526.

4.1.2 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the corrosion-resisting steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number (NSN), the item identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as indicated in table I. The unit of product for examination purposes shall be one scissors.

TABLE I. Sampling criteria for examination.

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in table II. The unit of product for test purposes shall be one scissors. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:58Z
Check the source to verify that this is the current version before use.

TABLE II. Sampling criteria for tests.

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL
Scissors	Hardness	3.3 & GG-I-526	4.4.1	S-2	1.5
	Dimensions	3.4	-----	S-2	2.5
	Copper sulfate test	GG-I-526	GG-I-526	S-2	1.0
	Boil test	GG-I-526	GG-I-526	S-2	1.0
	Scissors performed test	3.6 & GG-I-526	4.4.2 & GG-I-526	II	1.0

4.2.3 Raw materials. A sufficient quantity of each lot of metal used in the manufacture of the scissors shall be selected for chemical analysis.

4.3 Examination.

4.3.1 Examination of scissors. The scissors shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in tables I and III. Examination, however, shall not be restricted to defects listed in table III. Any scissors in the sample containing one or more defects shall be rejected, and if the number of defective scissors in any sample exceeds the acceptance number for the sample, the lot represented by that sample shall be rejected.

TABLE III. Classification of defects.

Categories	Defects
Critical	None defined
Major	
101	Scissors not of the design specified.
102	Points of scissors sharp.
103	Scissors not of finish specified.
104	Scissors not free from cracks, burrs, pits, rough areas, scale and corrosion.
Minor	
201	Identification markings not present, not correct, not complete and not legible.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:58Z
Check the source to verify that this is the current version before use.

4.3.2 Packaging inspection. The inspection of the preservation-packaging and interior package marking shall be in accordance with the group A and B quality conformance inspection requirements of MIL-P-116. The inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Hardness test. Hardness shall be determined in accordance with ASTM Std. Method E 18 or E 92. Conversion of Vickers hardness values to Rockwell values shall be in accordance with ASTM Std. Hardness Conversion Tables E 140.

4.4.2 Scissors performance test. The test material shall be rubber dam meeting the requirements specified for National Stock Number 6520-00-754-2769, Rubber Dam, Medium, 6 inches square. Three separate, consecutive tests shall be performed by each scissors as described and specified in GG-I-526.

5. PACKAGING

5.1 Preservation-packaging. Preservation-packaging shall be level A or C as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Cleaning and drying. Each scissors shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.1.2 Preservative. Preservative shall not be applied.

5.1.1.3 Unit packaging. Each scissors shall be wrapped in neutral tissue or other suitable anticorrosive material. Scissors shall be further packaged in accordance with MIL-P-116, method 1C-1.

5.1.1.4 Intermediate packaging. Six (6) scissors shall be packaged in a box of appropriate size constructed in accordance with PPP-B-636, type CF, class domestic. Closure of box shall be as specified in the appendix of PPP-B-636.

5.1.2 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be Level A, B or C as specified (see 6.1).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:58Z
Check the source to verify that this is the current version before use.

5.2.1 Level A. Two hundred sixteen (216) scissors shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in the appendix of PPP-B-636.

5.2.2 Level B. Two hundred sixteen (216) scissors, shall be packaged in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II as specified in PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking. Each unit, intermediate package and exterior container shall be marked as specified in MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number, title and date of this specification.
(b) Selection of applicable levels of packaging and packing required (see 5.1 and 5.2).
(c) National Stock Number (NSN).

6.2 NSN coverage. This specification covers the following item appearing in the Federal Supply Catalog: NSN 6515-00-365-0640 Scissors, General Surgical, Straight, 7 inch.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-2017

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:58Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-S-37545 - SCISSORS, GENERAL SURGICAL, STRAIGHT, 7 INCH

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:58Z
Check the source to verify that this is the current version before use.