A-A-53797
13 February 1989

COMMERCIAL ITEM DESCRIPTION

IMPRESSION TRAY, DENTAL (Water-cooled)

The General Services Administration has authorized the use of this commercial item description.

Salient characteristics:

This commercial item description covers dentulous type, water-cooled, rim lock impression tray and tubing assembly suitable for taking oral impressions with reversible agar hydrocolloid impression materials. The impression trays and tubing assembly shall be as specified herein:

Class 1 - Trays, Impression

Class 2 - Tubing Assembly

The impression trays shall be furnished for upper or lower, dentulous impressions, and in the sizes (except for the tubing assembly) as listed in the "NOTE" section for the applicable National Stock Number and Item Identification.

Material.

Tray body, handle, and entrance/exit tubes. The body, the handle, and the entrance/exit tubes shall be fabricated of brass conforming to Copper Alloy No. 260 as specified in ASTM B36 for composition only.

The tray body and the liquid entrance/exit tubes shall be an annealed temper and the handle shall be half-hard temper.

Rim wire. The peripheral rim and palatal area wire shall be a soft, solid, copper wire.

Solder. The solder used for attaching the two halves of the tray, the handle, the liquid entrance/exit tubes, and the wire to the tray body shall be 95 percent tin and 5 percent antimony as specified in ASTM B32 for composition only.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Tubing assembly hose. The tubing assembly hose shall be fabricated of neoprene.

Style, design and dimensions. The dental impression trays shall be suitable for use in taking oral impressions with reversible agar impression materials. The style, design and dimensions shall be in accordance with Defense Personnel Support Center (DPSC) Drawing 24512.

Class 1 - Impression trays shall be dentulous type, water-cooled, rim lock trays. The trays shall not be perforated so as to allow for maximum cooling and easy cleaning. The trays shall have an internal design that ensures a uniform flooding throughout. The liquid entrance/exit tubes shall be soldered to and extend beyond the edges of the handle. The upper trays shall have a wire rim in the palatal area to reinforce the retention of the impression material. Each tray shall be stamped with the appropriate size designation and shall have a hole in the handle for board hanging. The trays shall be suitable for use with Class 2 - Tubing Assembly. The tubing assembly shall consist of two hoses connected to a one-piece, molded connector for the trays. The other end of the tubing assembly shall have an appropriate fountain attachment on the entrance hose which shall be connected to the dental operating unit water outlet. The other hose shall be the exit hose and shall be placed in the cuspidor. Each hose of the tubing assembly shall be 48 inches in length, minimum.

Construction. The impression tray shall be of two piece construction, one upper half and a lower half. The lower half of the tray shall have the handle, the liquid entrance/exit tubes and the wire used to direct the water flow secured by soldering. The upper half of the tray shall be placed onto the finished lower half and shall be secured by soldering around the periphery. The wire rim around the periphery of the assembled tray and the wire in the palatal area shall be secured by soldering.

Finish. Trays shall be nickel or chromium over nickel plated. Trays shall have a smooth satin or bright polished finish. Finish on edges and surfaces shall be uniform and free of burrs, sharp edges, crevices, rough areas and cracks. Wire rims shall be smoothly and uniformly contoured. Soldered joints shall be continuous and free from projections, fractures and voids. Plating shall be uniform in appearance. Plating shall be free from blisters, visible porosity, peeling, flaking, or other defects. Soldered connections shall be plated over.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Identification marking. Trays made in U.S.A. shall be marked with the name or registered trademark of the contractor. Trays of foreign (other than U.S.) manufacture shall be marked with the name of origin, together with the name or registered trademark of the contractor. All marking shall be on the top face of the handle and shall be capable of being read when the tray is hanging on the trayboard. All marking shall be permanent and legible. Each tray shall be marked on the handle with the letter or letters listed under marking for the applicable size.

Marking	Size
S	Small
M	Medium
L	Large
XL	Extra Large

Workmanship. The trays shall be free of defects which detract from their appearance or impair their serviceability and intended use. The trays shall be uniform in quality and shall be smooth and free from burrs, seams, cracks, and free from blisters. Solder joints shall be uniform, continuous and free of flux residue or foreign material.

Unit. Each (EA). One impression tray of the class and size, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification, where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:24Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one tray.

Examination. Examination shall include, but shall not be limited to the following:

Defects
Style and design for the applicable size tray not as specified.
Tray not constructed as specified.
Edges and surfaces not free of burrs, sharp edges, or cracks.
Plating not free from blisters or peeling.
Soldered connections not plated over.
Wire rim not securely soldered to tray body.
Handle not securely soldered to tray body.
Liquid entrance/exit tubes not securely soldered to the handle.
Dimensions not as specified for the applicable size.
Identification marking not present, not properly located, illegible, or incomplete.
Letter or letters designating size missing or incorrect for the applicable size.
Trays not free from dirt, polishing compound, or foreign matter.
Finish not smooth or brightly polished.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

National Stock Number	Item Identification
6520-01-119-7681	TUBING ASSEMBLY, IMPRESSION TRAY, DENTAL

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

6520-01-120-2650	IMPRESSION TRAY, DENTAL, Dentulous, Lower, Small
6520-01-120-2651	IMPRESSION TRAY, DENTAL, Dentulous, Lower, Medium
6520-01-120-2652	IMPRESSION TRAY, DENTAL, Dentulous, Lower, Large
6520-01-120-2653	IMPRESSION TRAY, DENTAL, Dentulous, Lower, Extra Large
6520-01-120-2654	IMPRESSION TRAY, DENTAL, Dentulous, Upper, Small
6520-01-120-2655	IMPRESSION TRAY, DENTAL, Dentulous, Upper, Medium
6520-01-120-2656	IMPRESSION TRAY, DENTAL, Dentulous, Upper, Large
6520-01-120-2657	IMPRESSION TRAY, DENTAL, Dentulous, Upper, Extra Large

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2077

06008

Location: NCR Disk #6

☆U.S. GOVERNMENT PRINTING OFFICE: 1989 - 604-833/12279

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.