A-A-53798
31 August 1989

COMMERCIAL ITEM DESCRIPTION

PLUGGER SET, FINGER ENDODONTIC

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers endodontic finger plugger sets for use in condensation of gutta percha during endodontic procedures.

Salient characteristics:

Material. The operative end of each plugger shall be fabricated from corrosion-resistant class 3 type 302 steel in accordance with ASTM F899. The handle shall be fabricated from nickel plated brass, or as an alternate rigid plastic designed to withstand autoclave and dry heat sterilization procedures.

Hardness. The operative end of the plugger shall have a Rockwell hardness of C49 to C52 when tested for hardness in accordance with ASTM E18, ASTM E92 and ASTM E140.

Style, design, and dimensions. Each set of pluggers shall consist of four instruments. The plugger shall be a hand operated endodontic instrument used to pack and condense a restorative material within a prepared cavity. The plugger shall be securely held within the handle. The four pluggers shall increase in diameter from small to large. The plugger shall have a knurled handle, designed to provide a secure grip. The operative end shall not loosen or break during use. The plugger shall show no signs of corrosion when tested for corrosion as specified herein. The style, design and dimensions of the pluggers shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 22401.

Finish. The operative end of the plugger shall have a fine polished mirror finish, to preclude the adherence of filling material.

Marking. The plugger shall be permanently marked with the size represented by the appropriate letter on the end of the handle and color coded with a circular band on the handle. The size designation and color coding shall be as follows:

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Color	Size
Yellow	A
Red	B
Blue	C
Green	D

Workmanship. The pluggers shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing four pluggers, as specified, constitutes one unit.

Unit package. Each plugger set shall be packaged in a rigid container designed to prevent damage during shipment or storage.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of supplies to which such records relate.

Certificates of quality. Certificates of quality for raw material, supplied by the manufacturer of the steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Corrosion-resistance. The plugger shall be autoclaved under a pressure of 19 kPa ± 2.1 kPa (2.8 psi ± 0.3 psi) at 132°C ± 5°C (270°F ± 9°F) for twenty minutes and dry heat sterilized at 160°C ± 5°C (320°F ± 9°F) for one hundred and twenty minutes. All exposed surfaces shall be examined for signs of corrosion.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and an inspection level of II. The unit of product for examination purposes shall be one plugger. Examination shall include but not be limited to the defects listed below.

Defects

Operative end not securely attached to handle.
Operative end not smoothly finished.
Instrument not free from cracks, pits, burs, sharp edges, scale, rough areas or corrosion.
Finish of the instrument not as specified.
Style and design not as specified.
Identification markings not present, not complete, not correct, or not legible.
Operative end not concentric with center of handle.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 at an inspection level of S-1. The unit of product for test purposes shall be one finger plugger.

Contractor certification. The contractor shall ceertify that the product offered meets the salient characteristics of this description and conforms to the producers own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

National Stock Number	Item Identification
6520-01-257-4656	PLUGGER SET, Finger, Endodontic, 21 mm.
6520-01-257-4657	PLUGGER SET, Finger, Endodontic, 25 mm.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

PROJECT NO. 6520-2082

LOCATION: NCR DISK #5

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.