A-A-53799
31 August 1989

COMMERCIAL ITEM DESCRIPTION

SPREADER SET, FINGER, ENDODONTIC

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers endodontic finger spreader sets for use in the condensation of gutta percha during endodontic procedures.

Salient characteristics:

Material. The operative end of each spreader shall be fabricated from corrosion-resistant class 3 type 302 steel in accordance with ASTM F899. The handle shall be fabricated from nickel plated brass.

Hardness. The operative end of the spreader shall have a Rockwell hardness of C49 to C52 when tested for hardness in accordance with ASTM E18 or ASTM E92 and E140.

Style, design and dimensions. Each set of spreaders shall consist of four different diameter instruments. The spreader shall be a hand-operated, pointed and tapered endodontic instrument designed for compressing the filling material laterally against the walls of a tooth. The end of the spreader shall be securely held within the handle. The spreader shall have a knurled handle designed to provide a secure grip. The operative end shall not loosen or break during use. The spreader shall show no signs of corrosion when tested for corrosion as specified herein. The style, design and dimensions of the spreaders shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 22402.

Finish. The operative end of the spreader shall have a fine, smooth polished mirror finish, to preclude the adherence of filling material.

Marking. The spreader shall be permanently marked with the size, represented by the appropriate number, on the end of the handle and color coded with a circular band on the handle. The size designation and color coding shall be as follows:

AMSC N/A FSC 6520
DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Color	Size
Purple	10
Yellow	20
Blue	30
Black	40

Workmanship. The spreaders shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Set (SE). One package containing one set of four spreaders, one each of sizes #10, #20, #30 and #40, as specified, constitutes one unit.

Unit package. Each set shall be packaged in a rigid container designed to prevent damage during shipment or storage.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of supplies to which such records relate.

Certificates of quality. Certificates of quality for raw material, supplied by the manufacturer of the steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Corrosion resistance. The spreader shall be autoclaved under a pressure of 19kPa ± 2.1 kPa (2.8 psi ± 0.3 psi) at 132°C ± 5°C (270°F ± 9°F) for twenty minutes and dry heat sterilized at 160°C ± 5°C (320°F ± 9°F) for one hundred and twenty minutes. All exposed surfaces shall be examined for signs of corrosion.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an inspection level of II. Unit of product for examination purposes shall be one spreader.

Examination. Examination shall include but shall not be limited to the following defects.

Defects.

Operative end not securely attached to handle.
Operative end not smoothly finished.
Instrument not free from cracks, pits, burs, sharp edges, scale, rough areas or corrosion.
Style and design not as specified.
Identification markings not present, not complete, not correct, or not legible.
Handle not serrated.
Operative end not concentric with center of handle.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 at an inspection level of S-1. The unit of product for test purposes shall be one spreader.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-01-261-5898	SPREADER SET, FINGER, ENDODONTIC

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2083

LOCATION: NCR DISK #5

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.