INCH-POUND

A-A-53803B
15 April 1992
SUPERSEDING
A-A-0053803A (DM)
30 April 1991

COMMERCIAL ITEM DESCRIPTION

SYRINGE, HYPODERMIC, General Purpose, Disposable

The General Services Administration has authorized the use of this commercial item description as a replacement for commercial item description A-A-53796 which is cancelled.

This commercial item description covers sterile, disposable, plastic Luer-slip and Luer-lock hypodermic syringes suitable for a single injection.

Salient Characteristics:

General. Syringes shall be furnished in the following types and sizes:

Type I Luer Slip

Type II Luer Lock

Sizes:
1. 3 ml
2. 5 or 6 ml
3. 10 or 12 ml
4. 20 or 25 ml
5. 30 or 35 ml
6. 50 or 60 ml

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

The sizes and graduation markings shall be in accordance with Table I.

TABLE - I

GRADUATION MARKINGS

Sizes	Main	Sub
1	0.5	0.1
2	1	0.2
3	1	0.2
4	5	1
5	5	1
6	5/10	1

Material. Syringe barrel and plunger shall be rigid polypropylene which meets the United States Pharmaceutical (USP) requirements for Class IV plastics. The interior of the barrel shall be lightly coated with an inert silicone lubricant. The plunger tip shall be inert material rubber, neoprene, or thermo-plastic elostomer, which meets the USP requirements for Elastomeric Closures for Injections. Plastic plunger tips shall meet the USP requirements for Class IV plastics.

Style, design and dimensions. Syringe barrel tip shall be in accordance with ANSI/HIMA MD70.1. Syringe barrel shall be flanged to provide finger grip and anti-roll feature. Inside of the barrel shall have a smooth, uniform surface, free from pits, grind marks, bubbles, flash, or other imperfections. Plunger shall glide through the barrel while maintaining fluid-tight fit. The plunger tip and barrel shall both be beveled and form a flush fit at zero volume. Barrel shall be transparent and the plunger tip shall have two visible rings touching the sides of the barrel.

The outside surface of the barrel shall be marked with graduations as specified in Table 1. Main graduation markings shall be at least 33 percent longer than subgraduations. Each main graduation shall also have the numercial volume indicated. Barrel shall also have the name or registered trademark of the contractor. Actual fluid volume shall be determined by the position of the distal ring of the plunger tip.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:24Z
Check the source to verify that this is the current version before use.

Accuracy test. Syringe shall deliver ± 5 percent of the capacity indicated at each main graduation mark when tested against a calibrated burette. As an alternate the determination of capacity test listed in ISO 7886 is acceptable.

Leakage test. The water pressure test system shall have a water pressure gage and a rubber plug. The needle shall be secured to the Luer-tip of the syringe by applying it with a 5 pound axial force and while maintaining this compressive stress twisting the needle hub 45^0 relative to the Luer tip. This method of attachment does not apply to syringes with Luer locks. The needle of the hypodermic shall be inserted into the plug and the water pressure brought up to 45 psig.

The plunger shall then be withdrawn to the 50 percent point of the graduated capacity where it shall be allowed to remain for 10 seconds. The plunger shall then be withdrawn to the 75 percent point and held there for another 10 seconds. The plunger shall then be withdrawn to the 100 percent point for a final 10 seconds.

During each 10-second period there shall not be enough water to pass the connection between the needle hub and the Luer tip of the syringe or needle-syringe junction and between the plunger tip and barrel of the syringe to produce a falling drop. After each successive test, any water that may have passed shall be wiped with a dry cloth and the next test applied. The needle of the hypodermic shall be withdrawn from the rubber plug and the water ejected at the rate of 1/2 cc/seconds until the plunger reaches the zero mark. The needle shall not leak or separate at its junction with the syringe.

Sterility, pyrogenicity, toxicity. The syringe shall be nonpyrogenic and nontoxic in accordance with USP. The syringe shall be sterile in accordance with USP or AAMI requirements. Sterility assurance level shall be 0.0001.

Instructions. Each unit package shall be provided with instructions describing the proper techniques for dispensing the sterile syringes and needles, as applicable, from the containers and preparing them for use.

Workmanship. The syringe shall be free from defects which detract from its appearance or impair its serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Unit. Package (PG).

For Sizes 1, 2 and 3 - One package containing one hundred sterile syringes with instructions for use, as specified, constitutes one unit.

For Size 4 - One package containing forty sterile syringes with instructions for use, as specified, constitutes one unit.

For Size 5, Type I and Size 6 - One package containing sixty sterile syringes with instructions for use, as specified, constitutes one unit.

For Size 5, Type II - One package containing twenty-five sterile syringes with instructions for use, as specified, constitutes one unit.

The individual package for the above sizes shall be a sealed, peel-open type or rigid plastic tamper proof container capable of maintaining sterility of contents unless package is damaged or opened.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:24Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be inspection level S-1 and AQL 1.0 in accordance with MIL-STD-105. The unit of product for test purposes shall be one syringe. (See Note 2)

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers'own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

NOTE (1): The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	SIZES	TYPES
6515-00-462-7348	SYRINGE, HYPODERMIC, Disposable, 3 ml, Luer-Lock, 100s	1	II
6515-00-754-0406	SYRINGE, HYPODERMIC, Disposable, 5 or 6 ml, Luer-Slip, 100s	2	I
6515-00-754-0412	SYRINGE, HYPODERMIC, Disposable, 10 or 12 ml, Luer-Slip, 100s	3	I

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

NSN	ITEM IDENTIFICATION	SIZES	TYPES
6515-01-273-1803	SYRINGE, HYPODERMIC, Disposable 20 or 25 ml, Luer Lock, 40s	4	II
6515-01-165-6742	SYRINGE, HYPODERMIC, Disposable, 30 or 35 ml, Luer Lock, 25s	5	II
6515-01-280-2320	SYRINGE, HYPODERMIC, Luer-Slip Tip, Disposable, 60 ml, 50s or 60s	6	I
6515-01-289-5310	SYRINGE, HYPODERMIC, Luer-Slip Tip, Disposable, 30 or 35 ml, 60s	5	I
6515-00-706-3521	SYRINGE, HYPODERMIC, Disposable, 5 or 6 ml, Luer-Lock, 50s	2	II
6515-00-465-1008	SYRINGE, HYPODERMIC, Disposable, 20 ml, Luer-Lock, 40s	4	II
6515-00-929-0677	SYRINGE, HYPODERMIC, Disposable, 30 or 35 ml, Luer-Slip, 50s	5	I
6515-00-494-8117	SYRINGE, HYPODERMIC, Disposable, 35 ml, Luer-Lock, 50s	5	II
6515-00-465-1009	SYRINGE, HYPODERMIC, Disposable, 50 or 60 ml, Luer-Lock, Each	6	II
6515-00-226-7688	SYRINGE, HYPODERMIC, Disposable, 10 or 12 ml, Luer-Lock, 50s	3	II

NOTE (2): As an alternate, a statistically equivalent sampling plan which would ensure a comparable degree of freedom from defects is acceptable.

NOTE (3): ANSI and ISO standards are available from the American National Standards Institute, Inc., West 42nd Street, New York, NY 10036

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-5202
Location: ML Disk #2
(SYRINGE2)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.