A-A-53813
31 January 1989

COMMERCIAL ITEM DESCRIPTION

SIMULATOR, ARRHYTHMIA, ELECTROCARDIOGRAM

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a portable arrhythmia simulator suitable for use in testing electrocardiograph units.

Salient characteristics.

The unit shall be a portable, battery operated, arrhythmia simulator capable of testing functions of cardiac monitoring systems. The unit shall be capable of simulating normal rhythm and 17 common arrhythmias for the purpose of testing recognition capabilities, detection protocol and proper operation of rate alarms.

Output. The unit shall provide the following waveforms:

a.	VPB 1	j.	Ventricular fibrillation
b.	Multifocal run	k.	Atrial fibrillation
c.	Missed beat	l.	Heart block
d.	Fusion beat	m.	Regular rhythm
e.	VPB 2	n.	Paced rhythm
f.	Couplet (pair)	o.	Tachycardia
g.	Run	p.	Bradycardia
h.	R on T wave	q.	Ventricular tachycardia
i.	Asystole	r.	Bigeminy

The simulator shall provide for 12 lead connections. The unit shall be capable of producing 1 mV and 1 volt lead II output signals. Ten universal ECG jacks shall be provided.

Battery. The unit shall operate from and be provided with an internal battery. The battery shall provide for at least 50 hours of operation. Battery replacement shall not require disassembly of the unit.

Weight. The unit weight shall not exceed 1 pound.

AMSC/NA FSC 6910

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Workmanship. The simulator shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One simulator, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:25Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: This Commercial Item Description is covered by the following item:

NSN	ITEM IDENTIFICATION
6910-01-186-8436	SIMULATOR, ARRHYTHMIA, ELECTROCARDIOGRAM

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

A-A-53813

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6910-0019

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.