A-A-53815
31 January 1989

COMMERCIAL ITEM DESCRIPTION

AUDIOMETER

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a portable manually operated audiometer suitable for diagnostic and screening audiometry in hospital and clinical environments.

Salient characteristics.

The audiometer shall comply with all applicable requirements of the following publications:

ANSI S3.6	Specifications for Audiometers
ANSI S3.13	Artificial Head-Bone for Calibration of Audiometer Bone Vibrators
UL 544	Electric Medical and Dental Equipment

The audiometer shall be suitable for operation on 115 V, 50/60 Hz, electric power. The unit shall be capable of pure tone air and bone conduction testing with masking to include narrow-band, speech weighted and white noise masking. The unit shall provide infinite attenuation of sound level output from -10 to 110 dB with detents at 5dB increments. Patient signaling capability shall be provided for determining hearing Threshold Levels (HTL). One pair of TDH-50 cushioned earphones which are color coded for the left and right ear shall be provided with each unit.

The unit shall be capable of producing narrow band frequencies of 250, 500, 750, 1000, 1500, 2000, 3000, 4000, 6000, and 8000 Hz. The unit shall meet or exceed the following tolerances:

Frequency accuracy, 3 percent of indicated value

Noise level, 70dB below test tone

AMSC/NA FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Air conduction distortion, 3 percent total harmonic distortion

Attenuation linearity, 3dB over a 100dB range

Each unit shall be calibrated to the specified requirements. A certificate of calibration shall be provided with each unit.

Each unit shall be furnished with two (2) complete instructions for operation, maintenance and repair. A carrying case shall be supplied for moving and storage of the unit.

Workmanship. The audiometer shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One audiometer, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

First article inspection. First article inspection shall be performed on the audiometer when a first article sample is required. This inspection shall include the examination and all tests. All tests are to be performed using each specified input power voltage and frequency.

Sampling. Sampling for examination and testing shall be conducted in accordance with MIL-STD-105, with an AQL of 4.0 (percent defective) and inspection level for II for examination purposes and an AQL of 1.5 (percent defective) and inspective level of S-2 for testing purposes. The unit of product for examination and testing purposes shall be one audiometer. Examination shall not be limited to defects listed:

Any component part missing, broken or otherwise not usable.
Exposed edge not free of burrs or sharpness.
Finish not free of chips, cracks, scratches and other appearance defects.
Any component not secured in place.

Tests. Tests shall be conducted to determine compliance with this specification. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Verify actual frequencies are within 3 percent of indicated frequencies.
Verify amplitude linearity is within 3dB for 100dB range.
Verify air conduction distortion is less than 3 percent total harmonic distortion.
Verify rise and fall times are as specified in ANSI 3.6.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-160-8369	AUDIOMETER, Manually Operated, Portable 115 Volt, 50-60 Hz, AC, Solid State

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4387

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53815	AUDIOMETER

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (DO NOT STAPLE) and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation is too rigid, restrictive, loose, ambiguous or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4968 WASHINGTON D.C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.