A-A-53817
1 February 1989

COMMERCIAL ITEM DESCRIPTION

DISH, BIOLOGICAL STAINING (With Tray)

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification DD-D-431D which is cancelled.

This commercial item description covers a glass biological staining dish with a removable glass staining tray, in accordance with the following options:

Option I. Consists of glass staining dish with cover, removable glass staining tray, and a metal spring wire tray remover.

Option II. Consists of only a glass staining dish with glass cover.

Option III. Consists of only a replacement glass staining tray with metal spring wire tray remover.

Salient characteristics.

Shall be a rectangular glass staining dish with glass cover, and a removable staining tray with open bottom. A detachable spring wire holder shall also be supplied with the staining tray for lifting it in or out of the staining dish. The staining tray shall have the capacity to hold twenty (20) microscope slides in sizes measuring 73.5 to 77 millimeters (mm) long by 25 to 50 mm wide. The microscope slides shall be held upright in slots of the staining tray with the long dimensions in the horizontal position. The twenty (20) microscopes shall be held in back-to-back pairs.

Glass material. The staining dish, dish cover and staining tray shall be fabricated of clear soda lime glass.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Staining dish. The staining dish shall be heavy-walled, and rectangular in shape with side walls tapered on the inside. The staining dish shall have the following inside dimensions:

Length at top opening	94 ± 2 mm.
Length at bottom	89 ± 2 mm.
Width	74 ± 2 mm.
Depth	64 ± 2 mm.

Dish cover. The dimensions of the dish cover shall allow it to properly mate with the staining dish in closure. The dish cover shall include a heavy shoulder which, in closure, protrudes inside the staining dish.

Staining tray. The staining tray shall be designed to be inserted and removed from the staining dish while holding microscope slides. The staining tray shall be heavy-walled with open-bottom construction. Both ends along the short dimension (width) shall have ten (10) grooves integrally molded on their inner surfaces to accommodate twenty (20) microscope slides in pairs. Each groove shall be able to hold a pair of microscope slides back to back. The grooves shall be constructed in a manner that enables inserted microscope slides to fit snugly and bottom completely. The top location of both end walls shall have a hole for inserting the metal wire tray remover. The staining tray shall have the following outside dimensions:

Length (at top ends)	89 ± 2 mm.
Width (at top ends)	69 ± 2 mm.
Length (at bottom ends)	84 ± 2 mm.
Width (at bottom ends)	65 ± 2 mm.
Height	49 ± 2 mm.

The staining tray shall be capable of being placed inside the staining dish with sufficient clearance from the walls to preclude a snug fit. The inserted staining tray containing microscope slides shall not interfere with proper closure of the covering for the staining dish (i.e. without wire tray remover connected to staining tray).

Metal wire tray remover. A spring wire handle shall be supplied for inserting into staining tray cutouts for purposes of lifting the staining tray in and out of a staining dish. The wire handle shall have a nominal 5/64 inch diameter and shall be fabricated from corrosion-resisting steel or nickel copper alloy.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Workmanship. The staining dish, cover, staining tray and wire tray remover shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One dish with cover, tray, and tray remover, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and test shall be conducted in accordance with commercial practice to determine compliance with specification requirements.

Sampling for inspection and tests. Sampling for inspection and tests shall be conducted in accordance with commercial practice.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensionns and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractor/suppliers, with the requirement of all other applicable Federal, State and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	Item Identification	Specification Designated Option
6640-00-422-3810	DISH, BIOLOGICAL STAINING	I
6640-00-422-3815	DISH, BIOLOGICAL STAINING	II
6640-00-422-3820	TRAY, BIOLOGICAL STAINING	III

Ordering data. (Option required, intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6640-1522

Location: Disk No. 16
CID/DISH/BIOLOGICAL/S31

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.