A-A-53822
28 February 1989

COMMERCIAL ITEM DESCRIPTION

WIRE, NONELECTRICAL (Corrosion-resisting Steel, Soft)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers four sizes of soft corrosion-resisting nonelectrical wire for use in orthognathic surgery and treatment of maxillofacial trauma.

Salient characteristics.

General.

The wire shall be ligature wire suitable for use in orthognathic surgery and treatment of maxillo-facial trauma. The wire shall be round and considered "dead soft" for proper manipulation. The wire shall be of the following sizes:

SIZE	GAGE
1	24 (.020 inch)
2	25 (.018 inch)
3	27 (.014 inch)
4	28 (.012 inch)

Material. The wire shall be fabricated from corrosion resistant stainless steel type 316L in accordance with ANSI/ASTM A580.

Dimensions. The wire shall have the following dimensions.

SIZE	DIAMETER, INCH ± 0.0003	LENGTH, FEET Minimum
1	0.0200	45
2	0.0180	50
3	0.0140	90
4	0.0120	125

ANMC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Mechanical properties. The wire shall have the following mechanical properties when tested in accordance with ASTM A370.

SIZE	TENSILE Strength, PSI Minimum	YIELD PSI Minimum	ELONGATION Percent per 10 inch Minimum
1	100,000	30,000	40
2	109,000	30,000	48
3	105,800	30,000	48
4	98,000	25,000	48

Corrosion resistance. The wire shall show no corrosion when tested in accordance with the boil test of ASTM F1089.

Finish. The wire shall be furnished in the bright annealed or dead soft temper and shall have a bright clean smooth finish.

Marking. Spool shall be marked legibly with item identification.

Workmanship. The wire shall be uniform in quality, free from inclusions, kinks, nicks, splits, scratches, burrs, crimps, sharp edges or any defect that detracts from its appearance or impairs its serviceability.

Unit. Spool (SL).

For size 1, one spool containing .020 inch diameter, as specified, constitutes one unit.

For size 2, one spool containing .018 inch diameter, as specified, constitutes one unit.

For size 3, one spool containing .014 inch diameter, as specified, constitutes one unit.

For size 4, one spool containing .012 inch diameter, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one spool of wire. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5. Sampling for tests shall be conducted at inspection level S-1, AQL 1.0.

Examination. The wire shall be examined to determine compliance with the requirements contained in this specification. The wire shall be examined for defects, including, but not limited to, those listed below.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Defects

Size not as specified.

The wire not free of inclusions, kinks, nicks, or splits.

The wire not free of scratches, burrs, crimps, or sharp edges.

Finish not as specified.

Tests. Tests shall be conducted to include dimensions, to determine compliance with specification requirements.

Boil Test. The boil test shall be in accordance with ASTM F1089.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting of rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	SIZE
6520-01-057-9202	WIRE, NONELECTRICAL, CORROSION-RESISTING STEEL, ROUND, SOFT, 0.014 inch diameter, 27 gage, 90 feet.	3
6520-01-057-9203	WIRE, NONELECTRICAL, CORROSION-RESISTING STEEL, ROUND, SOFT, 0.012 inch diameter, 28 gage, 125 feet.	4

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

NSN	ITEM IDENTIFICATION	SIZE
6520-01-058-2517	WIRE, NONELECTRICAL, CORROSION-RESISTING STEEL, ROUND, SOFT, 0.020 inch diameter, 24 gage, 45 feet.	1
6520-01-058-7428	WIRE, NONELECTRICAL, CORROSION-RESISTING STEEL, ROUND, SOFT, 0.018 inch diameter, 25 gage, 50 feet.	2

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC DF010

Project No. 6520-2004

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4968 WASHINGTON D. C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-.RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53822	WIRE, NONELECTRICAL (Corrosion-resisting Steel, Soft)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.