COMMERCIAL ITEM DESCRIPTION

FORCEPS, TOOTH EXTRACTING, DENTAL, GENERAL SPECIFICATION FOR

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specifications GG-F-001240A and GG-F-1240 which are cancelled.

This Commercial Item Description covers dental tooth extracting forceps.

Salient Characteristics:

General.

Commercial Item Description Sheet. The tooth extracting forceps shall be in accordance with all requirements specified herein and the applicable commercial item description sheet (slash sheet-see Notes).

Material. The forceps shall be made of stainless steel in accordance with ASTM F899 as follows: The forceps forgings shall be made of class 4 martensitic stainless steel. The pins shall conform to the composition requirements of class 6 ferritic stainless steel. The screws shall conform to the composition requirements of class 4 martensitic stainless steel or class 6 ferritic stainless steel.

Hardness. After proper stress relieving, the forceps (beaks and handles) shall have a Rockwell hardness of C40 to C49, when tested in accordance with ASTM E18 or ASTM E92 and E140.

AMSC N/A FSC 6520

Distribution Statement A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Deformation. The forceps shall withstand the minimum loads as specified in the applicable commercial item description sheets for the respective forceps number, when tested as specified in the deformation test in the test section. During or as a result of testing, no chipping, cracking, fracturing, or permanent deformation shall occur on any portion of the forceps, and the beaks, pins, or screws shall not show any signs of weakness.

Style and design. The extracting forceps shall have anatomically shaped beaks and design of handles (for the applicable forceps number) suitable for use in exodontic procedures. The style, design, and dimensions of the forceps shall be as specified in the detail commercial item description sheet for the applicable forceps number and classification.

Construction. The gripping surfaces of the handle shall have male diamond knurling free from rough edges and sharp corners. The handle shall provide a comfortable slip resistant gripping surface. The edges of the joints shall be properly beveled to prevent pinching of tissues. The forceps shall be free from excessive stiffness and shall not have free play or indication of looseness when the forceps are in the open or closed position. The forceps shall comply with the box lock and clearance requirements of ASTM F1026.

Beaks. Concave surfaces (grooves) of beaks shall be properly centered with respect to the long axis of the beaks. All grooves, surfaces, and tips shall be well defined and properly formed.

Passivation. The forceps shall be properly passivated.

Finish. Unless otherwise specified in the commercial item description sheets, the forceps including the knurled gripping portion, shall have a bright finish. The finish of the grooves and grasping edges of the forceps (where applicable) shall be as specified in the commercial item description sheet for the applicable forceps number.

Weight. The weight of the forceps shall be as specified in the commercial item description sheets for the applicable forceps.

Corrosion resistance. The forceps shall show no signs of corrosion when tested in accordance with ASTM F1089.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Identification marking. The forceps shall be marked in accordance with ASTM F1026. In addition, the word "stainless" shall be inscribed into the instrument.

Workmanship. Forceps shall be free from defects which detract from their appearance or impair their serviceability, proper functioning, and intended use.

Unit. Each (EA). One forceps, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one forceps. The forceps shall be examined for defects, including but not limited to, those listed below:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Defects

Forceps not free from cracks, pits, burrs, sharp edges, scale, rough areas or corrosion.

Finish of the instrument not as specified.

Style and design not as specified.

Identification markings not present, complete, correct, or legible.

Edges of the joint of the forceps not properly beveled.

Joint of forceps not free from excessive stiffness.

Joint of forceps has excessive play or is excessively loose.

Contour of beaks not as specified.

Interfaces of lock joint are not regular or have large crevices.

Handle finish not as specified.

Instrument not permanently and legibly marked with the name or trademark of the contractor.

Instrument number not present.

Knurling or grooves on handle not as specified.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified below. The unit of product for test purposes shall be one forceps.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Characteristic	Requirement	Test procedures	Inspec. level	AQL
Composition analysis	As specified herein	As specified herein	*	---
Hardness	As specified herein	ASTM E18 or E92	S-2	1.0
Dimensions	As specified herein	_____	S-1	2.5
Copper sulfate test	ASTM F1089	ASTM F1089	S-2	1.0
Boil test	ASTM F1089	ASTM F1089	S-2	1.0
Weight	As specified herein	As specified herein	S-2	4.0
Deformation	As specified herein	As specified herein	S-2	1.0

*Sufficient amount from each lot of each metal used in the construction of the instrument shall be selected for composition analysis.

Chemical composition. In the absence of certificates of quality for the metallic raw material, chemical analysis shall be conducted in accordance with method 111 or 112 of FED-STD-151. In case of dispute, method 111 shall govern.

Hardness. Hardness shall be determined in accordance with ASTM Designation E18 (Rockwell) or E92 (Vickers or Diamond Pyramid).

Deformation test. A brass plate shall be placed between the beaks of the forceps and the load, as specified in the deformation section, shall be gradually applied over a period of 25 to 35 seconds, at points of greatest convexity of the handles. The test plate shall meet the requirements of class 8, alloy number 270, yellow brass for composition only. The test plate shall have a Rockwell hardness of B70 to B90. For all forceps, the test plate shall be of sufficient thickness so that the handle ends (free ends) and the beak tips shall not touch when the forceps are subjected to full load as specified.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Weight. The forceps shall be weighed on a calibrated scale of suitable sensitivity and accuracy.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling etc., involving shipment from receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTES:

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this specification.

(b) Title, number and date of the applicable specification sheet which shall include this specification letter and number, and the slash number. Examples:

A-A-53825/1	Forceps, Tooth Extracting, No. 22
A-A-53825/2	Forceps, Tooth Extracting, No. 222
A-A-53825/3	Forceps, Tooth Extracting, No. 13
A-A-53825/4	Forceps, Tooth Extracting, No. 150-AS
A-A-53825/5	Forceps, Tooth Extracting, No. 17
A-A-53825/6	Forceps, Tooth Extracting, No. 151-AS
A-A-53825/7	Forceps, Tooth Extracting, No. 210-AS
A-A-53825/8	Forceps, Tooth Extracting, No. 53L
A-A-53825/9	Forceps, Tooth Extracting, No. 53R
A-A-53825/10	Forceps, Tooth Extracting, No. 65
A-A-53825/11	Forceps, Tooth Extracting, No. 150
A-A-53825/12	Forceps, Tooth Extracting, No. 151
A-A-53825/13	Forceps, Tooth Extracting, No. 150S
A-A-53825/14	Forceps, Tooth Extracting, No. 151S

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

A-A-53825

MILITARY INTERESTS:

PREPARING ACTIVITY:

DoD-MB

Custodians:

Army-MD
Navy-MS
Air Force-03

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6520-2249
Location- NCR Disk 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS POP... CUT ALONG THIS LINE)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53825	FORCEPS, TOOTH EXTRACTING

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*: _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-BSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:25Z
Check the source to verify that this is the current version before use.