MIL-T-0012756B(EL)
31 December 1969
USED IN LIEU OF
MIL-T-12756A
20 May 1959

MILITARY SPECIFICATION

TEST SET, BATTERY

(BATTERY TESTER TS-183( )/U)

This limited coordination Military specification has been prepared by the US Army Electronics Command based upon currently available technical information, but it has not been approved for promulgation as a coordinated revision of Military Specification MIL-T-12756A. It is subject to modification. However, pending its promulgation as a coordinated Military Specification, it may be used in procurement.

1. SCOPE

1.1 This specification covers an equipment designated as Battery Tester TS-183( )/U, for use in determining the serviceability of dry batteries ranging in voltage rating from 1.5 to 200 volts.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-B-636	Box, Fiberboard
PPP-B-640	Box, Fiberboard, Corrugated, Triple-Wall
PPP-F-320	Fiberboard, Corrugated and Solid, Sheet Stock Container Grade and Cut Shapes
PPP-S-760	Strapping, nonmetallic, and connectors
PPP-T-97	Tape, Pressure-Sensitive Adhesive, Filament Reinforced

FSC 6625

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

MILITARY

MIL-P-116	Preservation, Methods of
MIL-P-11268	Parts, Materials and Processes used in Electronic Communication Equipment
MIL-M-13231	Marking of Electronic Equipment
MIL-F-14072	Finishes for Ground Signal Equipment
MIL-B-43291	Box, Fiberboard, Corrugated, Double-Wall, Weather Resistant

STANDARDS

MILITARY

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes
MIL-STD-129	Marking for Shipment and Storage
MIL-STD-252	Wired Equipment, Classification of Visual and Mechanical Defects
MIL-STD-810	Military Standard Environmentsl Test Methods for Aerospace and Ground Equipment

DRAWINGS

ELECTRONICS COMMAND

SC-A-46439	List of Accessories for Package Tester
SC-DL-86004	Battery Tester TS-183( )/U

(Copies of specifications, standards and drawings required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer. Both the title and number or symbol should be stipulated when requesting copies.)

3. REQUIREMENTS

3.1 Description. The Battery Tester TS-183( )/U is a portable unit used in measuring the load and open circuit voltages of dry batteries. Voltage range 0 to 2, 10, 50, 200 Vdc. Voltmeter sensitivity 1,000 ohms/volt.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

3.8 Service conditions. The Test Set Battery TS-183( )/U shall be constructed so that no fixed part shall become loose, cracked, flaked, pitted or blistered, no moveable part or control shall be shifted in setting, position or adjustment, and no permanent degradation shall be caused in performance below that specified in the following service conditions.

3.8.1 Low temperature. The test shall be subjected to the low temperature extremes of -4°F operating and -65°F storage and shall meet all Group A requirements at times specified when tested in accordance with 4.8.1.

3.8.2 High temperature The test set shall be subjected to the high temperature extremes of 125°F operating and 160°F storage and shall meet all Group A requirements at times specified when tested in accordance with 4.8.2.

3.8.3 Humidity. The test set shall exhibit no physical damage of parts and material such as corrosion rust blistering, swelling deterioration, etc., or electrical degradation and shall meet all Group requirements at times specified when tested in accordance with 4.8.3.

3.8.4 Vibration. The test shall show no evidence of breakage or loosening of parts and shall meet all Group A requirements when tested in accordance with 4.8.4.

3.8.5 Bounce, loose cargo. The test set shall show no evidence of physical damage except minor surface abrasion and shall meet all Group A requirements when tested in accordance with 4.8.5.

3.8.6 Rain The test set shall show no evidence of water penetration and shall meet all Group A requirements when tested in accordance with 4.8.6.

3.8.7 Fungus. The test set shall show no evidence of viable fungus and shall meet all Group A requirements when tested in accordance with 4.8.7.

3.9 Visual and mechanical The test set shall meet the requirements of MIL-STD-252 as specified in 4.9

3.10 Interchangeability. Like units, assemblies, subassemblies, and replaceable parts shall be physically and functionally interchangeable, without modification of such items or the equipment. Individual items shall not be hand picked for fit or performance Reliance shall not be placed on any unspecified dimension rating, characteristics etc. (see 4.10).

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

3.8 Service conditions. The Test Set Battery TS-183( )/U shall be constructed so that no fixed part shall become loose, cracked, flaked, pitted or blistered, no moveable part or control shall be shifted in setting, position or adjustment, and no permanent degradation shall be caused in performance below that specified in the following service conditions.

3.8.1 Low temperature. The test shall be subjected to the low temperature extremes of -40°F operating and -65°F storage and shall meet all Group A requirements at times specified when tested in accordance with 4.8.1

3.8.2 High temperature The test set shall be subjected to the high temperature extremes of 125°F operating and 160°F storage and shall meet all Group A requirements at times specified when tested in accordance with 4.8.2.

3.8.3 Humidity. The test set shall exhibit no physical damage of parts and material such as corrosion rust blistering, swelling deterioration, etc , or electrical degradation and shall meet all Group requirements at times specified when tested in accordance with 4.8.3.

3.8.4 Vibration. The test shall show no evidence of breakage or loosening of parts and shall meet all Group A requirements when tested in accordance with 4.8.4

3.8.5 Bounce, loose cargo The test set shall show no evidence of physical damage except minor surface abrasion and shall meet all Group A requirements when tested in accordance with 4.8.5.

3.8.6 Rain The test set shall show no evidence of water penetration and shall meet all Group A requirements when tested in accordance with 4.8.6

3.8.7 Fungus. The test set shall show no evidence of viable fungus and shall meet all Group A requirements when tested in accordance with 4.8.7.

3.9 Visual and mechanical The test set shall meet the requirements of MIL-STD-252 as specified in 4.9.

3.10 Interchangeability Like units, assemblies, subassemblies, and replaceable parts shall be physically and functionally interchangeable, without modification of such items or the equipment. Individual items shall not be hand picked for fit or performance. Reliance shall not be placed on any unspecified dimension rating, characteristics etc. (see 4.10).

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

3.11 Bounce preconditioning. Each test set produced shall be subjected to the bounce preconditioning test of 4.6 prior to Group A inspection. After bounce preconditioning, the test set shall not be repaired, aligned, cleaned or otherwise changed prior to performing electrical preconditioning and Group A inspection.

3.12 Workmanship. The test set shall be manufactured and assembled in accordance with the applicable portions of MIL-P-11268 as follows:

General requirements for plastic material and parts

Wiring and cabling including:
- Slack
- Protection
- Clearance
- Splicing
- Connectors, general
- Grounding, general

Flux and cleaning agents for soldering
Process for soldering
Cleaning of equipment
General process for securing of parts
Process of welding
Connectors, general
Mechanical operation of controls
Tropicalization of material
Lubricants

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the contract and specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

4.2 Classification of inspection. Inspection shall be classified as follows:

a. First article inspection (does not include preparation for delivery) (see 4.3).

b. Inspection covered by subsidiary documents (see 4.4).

c. Quality conformance inspection.

(1) Quality conformance inspection of equipment before preparation for delivery (see 4.5).

(2) Quality conformance inspection of preparation for delivery (see 4.11)

4.3 First article inspection. This inspection shall be performed by the contractor and witnessed by a Government representative unless otherwise specified in the contract. It shall consist of Group A, B and C inspections (see Tables I, II and III respectively) on the three (3) samples of the completely assembled Test Set Battery TS-183( )/U The first article samples shall be inspected in this order Group A and B for all units; and designated units for specific Group C inspections. Designation of units to specific Group C inspections shall be determined by the Government production engineer. After completion of all Group C tests, conforming units must be reinspected and pass all Group A inspections.

4.4 Inspection covered by subsidiary documents. The following shall be inspected under the applicable subsidiary documents as part of the inspection of equipment before preparation for delivery.

Item	Where required
Finish	3.5
Marking	3.6
Workmanship	3.12

4.5 Quality conformance inspection of equipment before preparation for delivery. The contractor shall perform the inspection specified in 4.4 and 4.5.1 through 4.5.4. This does not relieve the contractor of his responsibility for performing any additional inspection which is necessary to control the quality of the product and to assure compliance with all specification requirements The Government will review and evaluate the contractor's inspection procedures and examine the contractor's inspection records. In addition the Government-at its discretion-may perform all or any part of the specified inspection to verify the contractor's compliance with specified requirements. Test Equipment for Government verification inspection shall be made available by the contractor.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

4.5.1 Group A inspection. Each unit produced shall be inspected by the contractor for conformance to all the Group A requirements and tests in Table I. Each lot shall be subject to Government verification and this inspection, including sampling shall conform to Table I and the ordinary inspection proceaures of MIL-STD-105.

4.5.1 1 Order of inspection within Group A. Group A inspection shall be performed in any order which is satisfactory to the Government.

TABLE I - Group A Inspection

Inspection	Req Para	Insp Para	AQL Major	AQL Minor
Electrical	3.7	4.7	1.5dphu*	**
Visual and Mechanical (Battery Tester)	3.9	4.9	1.5dphu	6.5dphu
Visual and Mechanical (Wood Case)	3.9	4 9	1.0dphu	4.0dphu

*Defects per nundred units
**All electrical tests are major

4.5.2 Group B inspection. This inspection, including sampling, shall conform to Table II and to the special procedures for small sample inspection of MIL-STD-105. Group B inspection shall be performed on inspection lots that have passed Group A inspection and on samples selected from units that have been subjected to and meet the Group A inspection.

4.5.2.1 Order of inspection within Group B. Group B inspection snall be performed in any order which is satisfactory to the Government.

TABLE II - Group B Inspection

Inspection	Req Para	Test Para	AQL
Interchangeability	3.10	4.10	6 5aphu*

*Defects per hundred units

4.5.3 Group C inspection This inspection shall consist of the inspection specified in Table II anc snall be performed or sample units that have been subjected to anc met Group A and B inspection. Sample units shall be selected in accordance with 4.5.3.1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

4.5.3.1 Sampling for inspection.

4.5.3.1.1 Three (3) samples of complete equipment shall be selected at random. One (1) sample for each Group 1 through Group 3 in Table III. The samples selected shall be at the start of the contract from the first quality conformance inspection lot. These samples shall constitute Group C requirement for fifty (50) units produced.

4.5.3.1.2 Thereafter two (2) samples of complete equipment shall be selected at random; one (1) sample for each Group 1 and Group 2 in Table III from every fifty (50) units and fraction thereof produced.

4.5.3.2 Evaluation of samples submitted to individual Group C inspection. After referenced samples in 4.5.3.1 have been subjected to an individual Group C inspection test and prior to subjection to the next Group C test (unless otherwise specified in the individual test) the sample shall meet all Group A requirements.

4.5.3.3 Noncompliance. The contractor shall immediately report in writing each Group C failure occurrence, including details of the failure and characteristics affected. The contractor shall immediately investigate the cause of failure and further report the results of investigation and details of the proposed corrective action on (i) the process and materials, as applicable and (ii) all units of product which were manufactured under the same conditions and which the Government considers subject to the same failure. Reports shall be forwarded to the responsible technical activity designated in the contract through the Quality Assurance Representative. After corrective action has been taken additional sample units shall be subjected to Group C inspection (all inspections, or the inspections which the sample failed at the option of the Government). The final acceptance and shipment will be withheld until the Group C reinspection results have shown that the corrective action was effective.

4.5.3.4 Order of inspection within Group C. Group C inspection shall be performed in the order shown in Table III.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

TABLE III - Group C Inspection

Inspection	Req Para	Test Para
Group 1		
Low Temperature	3.8.1	4.8.1
High Temperature	3.8.2	4.8.2
Humidity	3.8.3	4.8.3
Group 2		
Vibration	3.8.4	4.8.4
Bounce, loose cargo	3.8.5	4.8.5
Rain	3.8.6	4.8.6
Group 3		
Fungus	3.8.7	4.8.7

4.5.4 Reinspection of conforming Group B and C sample units. Unless otherwise specified, sample units which have been subjected to and passed both Group B and Group C inspection may be accepted on contract or order provided all damage is repaired and the sample units are subjected to and pass Group A inspection.

4.5.5 Disposition of Group C sample units. Unless otherwise specified, sample units which have been subjected to and passed Group C inspection may be delivered on the contract or order provided units meet requirements of 4.5.4.

4.6 Bounce preconditioning. Each test set produced shall be placed in its normal operating position on the table of a package tester of suitable capacity as made by the L.A.E. Corporation, Skaneateles, New York, with accessories as per Model 1000SC and SC-A-46439, or equal. The package tester, shafts in phase, shall have speed such that it is just possible to insert a 1/32 inch thick strip of material under one corner or edge of the equipment under test to a distance of three (3) inches as the equipment bounces. The equipment shall be subjected to this preconditioning for one (1) minute and shall meet the requirements of 3.11.

4.7 Electrical.

4.7.1 Electrical tests. The tester shall be calibrated with the "no Load-test" switch in the "No Load" position. The multiplier resistors shall be adjusted on all ranges to a tolerance of ±1/2 percent of the voltage indicated at Division 26 on the upper scale and at Division 35 on the lower scale. The indicated reading after each test shall meet the applicable requirements specified in 3.7.1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

4.7.2 Voltage test 2V range. Apply 0.4, 1.4 and 2.0 volt to this range and read error from Table IV.

4.7.3 Voltage test 10V range. Apply 2, 7 and 10 volt to this range and read error from Table IV.

4.7.4 Voltage test 50V range. Apply 10, 35 and 50 volt to this range and read error from Table IV.

4.7.5 Voltage test 200V range. Apply 40, 140 and 200 volt to this range and read error from Table IV.

4.7.6 Load resistor test. Using Wheatstone Bridge measure the resistance of the load resistors and read error from Table V.

TABLE IV

Meter Range	Test Value	Limits Max.	Limits Min.	Test Value	Limits Max.	Limits Min.	Test Value	Limits Max.	Limits Min.
2	0.4	0.42	0.38	1.4	1.428	1.37	2	2.10	1.90
10	2	2.10	1.90	7	7.14	6.86	10	10.5	9.5
50	10	10.5	9.50	35	35.7	34.30	50	52.5	47.5
200	40	42.0	38.0	140	142.8	137.2	200	210	190

TABLE V

Resistor Number	Nom Value	Limits Max.	Limits Min.	Resistor Number	Nom Value	Limits Max.	Limits Min.
R-1	2	2.02	1.98	R-14	40K	42K	38K
R-2	4	4.04	3.96	R-15	860	903	817
R-3	6	6.06	5.94	R-16	1K	1050	950
R-4	8	8.08	7.92	R-17	1.3K	1365	1235
R-5	12	12.12	11.88	R-18	2K	2100	1900
R-6	20	20.2	19.8	R-19	3.1K	3255	2945
R-7	35	35.35	34.65	R-20	3.5K	3675	3325

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

TABLE V (Cont'd)

Resistor Number	Nom. Value	Limits Max.	Limits Min.	Resistor Number	Nom Value	Limits Max.	Limits Min.
R-8	40	40.4	39.6	R-21	5K	5250	4750
R-9	50	50.5	49.5	R-22	7.5K	7875	7125
R-10	100	101	99	R-23	9K	9450	8550
R-11	300	303	297	R-24	12K	12600	11400
R-12	550	555.5	544.5	R-25	15K	15750	14250
R-13	30K	31500	28500	R-26	50K	52500	47500

4.8 Service conditions. The Test Set Battery TS-183( )/U shall be tested in accordance with the following conditions.

4.8.1 Low temperature. The test set in its normal operating configuration shall be subjected to the low temperature test specified in Procedure I, Method 502 of MIL-STD-810, with temperatures in Steps 2 and 4 held for one (1) hour each, and shall meet the requirements of 3.8.1.

4.8.2 High temperature. The test set in its normal operating configuration shall be subjected to the high temperature test specified in Procedure I, Method 501 of MIL-STD-810, with temperature in Step 4 held for one (1) hour and shall meet the requirements of 3.8.2.

4.8.3 Humidity. The test set in its normal operating configuration shall be subjected to the humidity test specified in Procedure II, Method 507 of MIL-STD-810, and shall meet the requirements of 3.8.3.

4.8.4 Vibration. The test set shall be subjected to the vibration test specified in Part 1, Procedure IX, Method 514 of MIL-STD-810, and shall meet the requirements of 3.8.4.

4.8.5 Bounce, loose cargo. The test set as prepared for field transportation shall be subjected to the bounce, loose cargo test specified in Part 2, Procedure XI, Method 514 of MIL-STD-810, and shall meet the requirements of 3.8.5.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

4.8.6 Rain. The test set in its normal operating configuration shall be subjected to the rain test specified in Procedure I, Method 506 of MIL-STD-810, and shall meet the requirements of 3.8.6.

4.8.7 Fungus. The test set in its normal operating configuration shall be subjected to twenty-eight (28) days of the fungus test specified in Procedure I, Method 508 of MIL-STD-810 and shall meet the requirements of 3.8.7.

4.9 Visual and mechanical. The test set shall be examined for the defects listed in MIL-STD-252 and shall meet the requirements of 3.9

4.10 Interchangeability The dimensions listed below shall be measured to determine conformance to the physical interchangeability requirement of 3.10. When a listed dimension is not within specified or design limits, it shall be considered a major defect.

a. External and internal dimensions of case, cover and insertable assemblies, when such dimensions affect mating of parts.

b. Location of hinges and fasteners on separable parts or assemblies which must mate, such as case, covers and mountings.

d. Location of terminals.

e. Location of switches, sockets and fasteners.

f. Location of cover and fasteners on carrying case which must mate.

g. Location of panel mounting screw holes on basic unit for mating with case.

4.11 Quality conformance inspection of preparation for delivery. Preparation for delivery shall be inspected in accordance with MIL-P-116 to determine conformance to the requirements of Section 5.

5. PREPARATION FOR DELIVERY

5.1 Preservation and Packaging. Preservation and packaging shall be level A or C as specified (see 6.1).

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

5.1.1 Level A.

5.1.1.1 Cleaning. Each Test Set, Battery TS-183( )/U shall be cleaned in accordance with process C-1 of MIL-P-116.

5.1.1.2 Drying. Each Test Set, Battery TS-183( )/U shall be dried in accordance with the applicable procedure of MIL-P-116.

5.1.1.3 Preservation application. None required.

5.1.1.4 Unit packaging. Unit packaging shall be in accordance with the method prescribed in MIL-P-116 as specified herein.

5.1.1.4.1 Technical literature Each technical literature shall be packaged per Method 1C-1.

5.1.1.4.2 Test Set, Battery TS-183( )/U. Each Test Set, Battery TS-183( )/U shall be packaged Method III as follows: Secure the test cable in the clips provided on the cover, close cover of transit case and secure fastenings provided. Cushion each transit case on all surfaces with cells or pads or both fabricated of fiberboard conforming to PPP-F-320, W6c, weather-resistant, designed and fabricated to protect all projections and absorb the shock of impact in handling and transit. Place the cushioned item within a close-fitting fiberboard box conforming to PPP-B-636, W5c, weather-resistant. Place the technical literature packaged as specified in 5.1.1.4.1 in the box directly under the lid of the box. Box closure shall be as specified in the appendix of the box specification.

5.1.2 Level C. Test Set, Battery TS-183( )/U shall be packaged in a manner that will afford adequate protection against corrosion, deterioration and damage during shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2(b)).

5.2.1 Level A. A quantity of Test Sets, Battery TS-183( )/U packaged as specified in 5.1 shall be packed within a close-fitting fiberboard box conforming to MIL-B-43291 or PPP-B-640, class weather-resistant, not to exceed the gross weight limitation of two hundred (200) pounds. Closure shall be in accordance with the appendix of the applicable box specification. The shipping container shall be reinforced by pressure-sensitive filament type or non-metallic banding conforming to PPP-T-97 Type IV or PPP-S-760, Type II, respectively, as specified in the appendix of the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

5.2.2 Level B. A quantity of Test Sets, Battery TS-183( )/U packaged as specified in 5.1 shall be packed in accordance with 5.2.1.

5.2.3 Level C. A quantity of Test Sets, Battery TS-183( )/U packaged as specified in 5.1 shall be packed in shipping containers in a manner that will afford adequate protection against damage to the package and its contents during shipment from the supply source to the first receiving activity, shipping containers shall comply with the rules and regulations of the command carrier as applicable to the mode of transportation.

5.3 Marking. In addition to any special marking required by the contract or order, interior packages and exterior shipping containers shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The Test Set, Battery TS-183( )/U is particularly adapted for use with smaller installations of communication equipments, where simplicity of operation and portability are important considerations.

6.2 Ordering data. Procurement documents should specify the following:

a. Title, number and date of this specification and any amendment thereto.

b. Level of preservation and packaging and packing (see section 5).

c. When rough handling and functional tests are required.

d. Quantity of first articles required.

e. Place of final inspection.

f. Technical literature required.

g. Quantity of tools and running spare parts required.

h. Marking for shipment and shipping containers.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

6.3 Nomenclature and type designation. The contractor should apply for nomenclature and type designation in accordance with the applicable clause in the contract. The parentheses in the type designation will be deleted or replaced by a letter identifying the particular design for example TS-12F/U.

6.4 Group C inspection. Approval to ship may be withheld, at the discretion of the Government, pending the decision from the contracting officer on the adequacy of corrective action. (See 4.5.3.3)

6.5 Verification inspection. Verification inspection by the Government will be limited to the amount deemed necessary to determine compliance with the contract and will be limited in severity to the definitive quality assurance provisions established in this specification and the contract. The amount of verification inspection by the Government will be adjusted to make maximum utilization of the contractor's quality control system and the quality history of the product

Custodian:
Army - EL

Preparing Activity:
Army - EL

Project No. 6625-A435

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No 22-R255

INSTRUCTIONS This sheet is to be filled out by personnel, either Government or contractor involved in the use of the specification in procurement of products for ultimate use by the Department of Defense This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost Comments and the return of this form will be appreciated Fold on lines or reverse side staple in corner and send to preparing activity Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements

SPECIFICATION
MIL-T-0012756B(EL), TEST SET, BATTERY (BATTERY TESTER TS-183( )/U)

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1 HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A GIVE PARAGRAPH NUMBER AND WORDING

B RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2 COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3 IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes" in what way?)*

4 REMARKS *(Attach any pertinent data which may be of use in improving this specification If there are additional papers attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 66 1426

REPLACES EDITION OF 1 OCT 64 WHICH MAY BE USED

ESC-FM 1068-68

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

FOLD

Department of the Army
Headquarters
U S.Army Electronics Command
Fort Monmouth, New Jersey 07703

POSTAGE AND FEES PAID

OFFICIAL BUSINESS

Commanding General
U.S.Army Electronics Command
ATTN AMSEL-TD-SS
Fort Monmouth, New Jersey 07703

FOLD

To detach this form, cut along this line

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.