MIL-T-12756A
20 MAY 1969
SUPERSEDING
MIL-T-12756(SigC)
22 MAY 1963

MILITARY SPECIFICATION

TEST SET, BATTERY (BATTERY TESTER TS-183( )/U)

This specification has been approved by the Department of Defense, and is mandatory for use by the Departments of the Army, the Navy, and the Air Force.

1. SCOPE

1.1 This specification covers a battery test set designated as Battery Tester TS-183( )/U (see 6.3). It is used to determine the serviceability of dry batteries ranging in voltage rating from 1.5 volts to 200 volts

2. APPLICABLE SPECIFICATIONS, DRAWINGS, AND STANDARDS

2.1 The following specifications, drawings and standards, of the issue in effect on date of invitation for bids, form a part of this specification:

SPECIFICATION

FEDERAL

NN-P-515 — Plywood, Container grade

PPP-T-76 — Tape, Pressure-Sensitive Adhesive, Paper, Water-Resistant

PPP-B-585 — Boxes; Wood, Wire bound

PPP-B-591 — Boxes, Fiberboard, Wood-Cleated

PPP-B-601 — Boxes, Wood, Cleated Plywood

PPP-B-621 — Boxes, Wood, Nailed and Lock-Corner.

PPP-B-636 — Boxes, Fiber.

PPP-B-640 — Boxes, Folding, Fiberboard, Corrugated, Triple-wall

MILITARY

MIL-P-116 — Preservation, Methods of

MIL-B-117 — Bags, Interior Packaging

MIL-T-945 — Test Equipment, for use with Electronic Equipment General Specification

MIL-M-10304 — Meters, Ruggedized, Electrical

MIL-W-13169 — Wire, Electrical (WS-16/U, WS-17/U, and WS-18/U) For Instrument Test Lead

STANDARDS

MILITARY

MIL-STD-105 — Sampling Procedures and Tables for Inspection by Attributes

MIL-STD-129 — Marking of Shipments

MIL-STD-252 — Wired Equipment, Classification of Visual and Mechanical Defects.

FSC 6625

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

DRAWINGS

SIGNAL CORPS

SC-D-15914 — Moisture - Resistance Test Cycle for Ground Signal Equipment.

SC-DL-86004 — Battery Tester TS-183( )/U Drawing List.

2.2 Other publications. The following documents form a part of this specification. Unless otherwise indicated, the issue in effect on date of invitation for bids shall apply.

AMERICAN TRUCKING ASSOCIATION, INC

National Motor Freight Classification Rules.

(Application for copies should be addressed to the American Trucking Association, Inc., 1424 16th Street, N.W., Washington 6, D.C.)

CONSOLIDATED CLASSIFICATION COMMITTEE

Uniform Freight Classification

(Copies of Uniform Freight Classification may be obtained from Consolidated Classification Committee, Room 202, Union Station, Chicago 6, Ill.)

U.S. POST OFFICE DEPARTMENT

Parcel Post Regulations

(Application for copies of the above publication should be obtained from any first class U.S. Post Office.)

3 REQUIREMENTS

3.1 Construction. Battery Tester TS-183 ( )/U shall be constructed in accordance with Drawing SC-DL-86004.

3.2 Test leads. All test leads shall conform to the applicable requirements of Specification MIL-T-9160.

3.3 Preproduction samples. The contractor shall furnish preproduction samples for approval as required by the invitation for bids and contract (see 6.2(c)).

3.4 Parts, materials, and processes. Parts, materials, and processes shall conform to Specification MIL-T-945 whenever applicable unless otherwise specified herein.

3.5 Meters and multipliers. The meter shall be a round flange, flush-mounting, hermetically-sealed voltmeter, 1000 ohms per volt, conforming to Specification MIL-M-10304. It shall be equipped with a knife-edge pointer, or equal, and capable of operation over all voltage ranges by the use of external multipliers. The meter shall be mounted using three, 9/64"-inch bolts spaced 120° apart and shall have the following dimensions:

Diameter of flange	4½ inches.
Depth behind barrel	1¼ inches.
Diameter of barrel	3⅝ inches.
Radius mounting bolt circle	3 1/32 inches.

3.6 Electrical requirements. The tester shall be calibrated with the "no load-test" switch in the "No Load" position. The multiplier resistors shall be adjusted on all ranges to a tolerance of ± ½% of the voltage indicated at Division 26 on the upper scale and at Division 35 on the lower scale. The meter reading from the remaining positions on all ranges shall be accurate within a tolerance of ± 2%. The test load resistors shall be within the tolerances specified in Drawing SC-DL-86004.

3.7 Service conditions.

3.7.1 *Equipment operating.* The equipment shall meet the requirements of Specification MIL-T-945 except as follows:

(a) *Shock test.* Shock test shall not be performed.

3.7.2 *Equipment nonoperating.* The equipment shall meet the requirements of Specification MIL-T-945 except as follows:

(a) *Fungus.* Does not apply (MIL-T-945).

(b) *Rain.* See 3.7.2.1.

(c) *Bounce.* See 3.7.2.2.

3.7.2.1. *Rain.* There shall be no visible evidence of leakage when subjected to the test of 4.7.

3.7.2.2. *Bounce.* The equipment shall not suffer any physical damage except surface abrasion when subjected to the bounce test 4.6.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

3.6 **Electrical requirements**. The tester shall be calibrated with the "no load-test" switch in the "No Load" position. The multiplier resistors shall be adjusted on all ranges to a tolerance of ± ½% of the voltage indicated at Division 26 on the upper scale and at Division 35 on the lower scale. The meter reading from the remaining positions on all ranges shall be accurate within a tolerance of ±2%. The test load resistors shall be within the tolerances specified in Drawing SC-DL-86004.

3.7 **Service conditions**.

3.7.1 **Equipment operating**. The equipment shall meet the requirements of Specification MIL-T-945 except as follows:

(a) **Shock test**. Shock test shall not be performed.

3.7.2 **Equipment nonoperating**. The equipment shall meet the requirements of Specification MIL-T-945 except as follows:

(a) **Fungus**. Does not apply (MIL-T-945).

(b) **Rain**. See 3.7.2.1.

(c) **Bounce**. See 3.7.2.2.

3.7.2.1. **Rain**. There shall be no visible evidence of leakage when subjected to the test of 4.7.

3.7.2.2. **Bounce**. The equipment shall not suffer any physical damage except surface abrasion when subjected to the bounce test 4.6.

3.7.3 **Test requirements**. After being subjected to the above tests, the battery tester shall meet the requirements of 3.6 except that the tolerances "± 1/2 of 1 percent" and "± 2 percent" stated in 3.6 should be replaced with the tolerances of "± 2 percent" and "± 5 percent," respectively, for these tests only.

3.8 **Instruction and data booklet**. The instructions and operation data shall be furnished in looseleaf booklet form, fastened securely to the inside of the cover of each battery tester. The instruction and data shall be printed on a good grade of white book paper laminated between two sheets of clear, transparent, polyvinyl - chloride - acetate plastic sheets. The two sheets shall be bonded together to seal against moisture.

3.9 **Preconditioning**. Items shall be capable of meeting the inspection of section 4, without subsequent processing, after subjection to the bounce preconditioning of 4.4.

3.10 **Technical literature**. Technical literature shall be furnished as specified by the contract.

3.11 **Workmanship**. Workmanship on components and equipments covered by this specification shall be such as to meet all the requirements of this specification and any referenced subsidiary specification, drawing, or other document, when

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

inspected in accordance with 4.5.

4. QUALITY ASSURANCE PROVISIONS

4.1 Contractor responsibility. Unless otherwise specified herein the supplier is responsible for the performance of all inspection requirements prior to submission for Government inspection and acceptance. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. Inspection records of the examinations and tests shall be kept complete and available to the Government as specified in the contract or order.

4.1 **Classification of inspection.** Inspection shall be classified as follows:

(a) Preproduction inspection.

(b) Procurement inspection.

(1) Procurement inspection of items before preparation for delivery (see 4.3)
(2) Procurement inspection of preparation for delivery (see 4.9)

4.2 **Preproduction inspection.** This inspection shall consist of the preproduction inspection specified in table I and the Group A inspection specified in table II respectively. Other nondestructive tests on preproduction samples may be performed to determine compliance with specified requirements. The preproduction inspection will normally be performed in this order: (1) vibration, (2) bounce, (3) extreme temperature, (4) rain; other preproduction inspection may precede, follow, or be interspersed between the foregoing.

Table I.

Preproduction Inspection

Inspection	Requirement par.	Test par.
Vibration	MIL-T-945	MIL-T-945
Bounce	3.7.2.2	4.16
Extreme Temperature	MIL-T-945	MIL-T-945
Rain	3.7.2.1	4.7
Humidity	MIL-T-945	MIL-T-945

4.3 **Procurement inspection of items before preparation for delivery.** The minimum inspection performed by the contractor shall be as specified by (a) through (c) below to demonstrate compliance with specified requirements. The Government will review and evaluate the contractor's inspection procedures and examine the contractor's records. In addition, the Government, at its discretion, will verify product inspection, the contractor's compliance with specified requirements (see 6.5). Such verification shall not exceed the amount

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

of inspection required herein of the contractor.

(a) Procurement inspection shall consist of Group A inspection as specified in 4.3.1

(b) When Standard MIL-STD-105 specifies action by the Government, the Government may authorize the contractor to perform any of such actions except that responsibility for acceptance rests with the Government.

(c) Each unit which will be subjected to Group A inspection shall be preconditioned after final assembly (see 3.8).

4.3.1 **Group A inspection.** This inspection (including sampling) shall conform to table II and Standard MIL-STD-105. Unless otherwise specified, normal inspection shall be used at the start of the contract Group A inspection shall be performed in any order which is satisfactory to the Government.

Table II
Government Group A inspection

Inspection	Requirement par.	Test par.	AQL	
			Major	Minor
Electrical Tests.	3.6	4.8	1.5 dphu*	. ..
Visual and Mechanical, (Battery Tester)..	.....	4.5	1 5 dphu*	6 5 dphu*
Visual and Mechanical, (Wood Case)...	.....	4.5	1.0%	4.0%

*Defects per hundred units.

4 4 **Bounce preconditioning.** The unit, with shock mounts (if any) removed or blocked, shall be placed in its normal operation position on the table of the Package Tester as made by the L. A. B. Corporation, Skaneateles, N.Y., or equal. The package tester, shafts in phase, shall have a speed such that it is just possible to insert a 1/32-inch-thick strip of material under one corner or edge of the unit to a distance of 3 inches as the unit bounces The unit shall be subjected to this preconditioning for 1 minute After bounce preconditioning, the unit shall not be repaired, alined, cleaned. or otherwise changed prior to subjection to procurement inspection.

4.5 **Visual and mechanical inspection.** Visual and mechanical inspection shall be in accordance with Standard MIL-STD--252.

4.6 **Bounce test.** The equipment, in its transit case, shall be placed on the table of the Package Tester as made by the L.A B. Corporation, Skaneateles, N Y., or equal, and shall be constrained from horizontal motion of more than 2 inches by suitable wooden fences. The package tester, shafts in phase, shall

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

be operated at a speed of 285 r.p.m + 1 percent, for a total of 3 hours. During each 1/2-hour period of the test, the transit case (with contained equipment) shall rest on a different face

4 7 **Rain test** The battery tester shall be tested, step by step, as follows

4.7 1 Dry at 150°F ± 5°F. for 48 hours

4 7.2 Condition for 4 hours at 77° ± 5°F. and 40 to 60 percent relative humidity

4 7 3 Take measurements to insure that the battery tester meets performance requirements.

4.7.4 Place the battery tester, in its normal operating position, in a test chamber maintained at 77° ± 5°F. Allow simulated rainfall to fall on exposed surfaces of the closed battery tester The water shall be at a temperature of 65° ± 5°F. and shall fall freely at a rate of 2 inches per hour Maintain this condition for 1 hour

4.7 5 With the rainfall continuing at the 2-inch rate, introduce a simulated wind blowing horizontally at 20 ± 5 miles per hour against one side of the battery tester The water shall fall or exposed surfaces and be blown by the wind against one side of the battery tester Maintain this condition for 1 hour.

4 7 6 Rotate the battery tester 90° horizontally to expose a second side of the battery tester to the wind and rain Maintain this condition for 1 hour

4 7 7 Continue as 4 7.6 until all four sides of the battery tester have been exposed to the wind and rain, each side for 1 hour.

4 7.8 Immediately after completion of the foregoing, repeat the measurement specified in 4 7 3 without removal from the chamber but with rain and wind discontinued Complete the measurements within 1 hour The measurements shall meet performance requirements.

4 8 **Electrical test** The accuracy shall be determined by using a d.c -voltage source variable over the range from 1 to 200 volts The low range for example 1 to 10 volts, shall be variable in increments of 0 1 volts A secondary meter such as, or equal to, a sensitive research polyranger with an accuracy of 1/4 of 1 percent, or better, shall be connected in parallel with the instrument under test Production units shall meet the accuracy requirements of 3.6.

4.9 **Procurement inspection of preparation for delivery.** The minimum inspection performed by the contractor shall be as specified in 4.11.1 and 4.11.2. The Government will review and evaluate the contractor's inspection procedures and examine the contractor's records. In addition, the Government - at its discretion - will verify, by product inspection, the contractor's compliance with specified requirements (see 6.5). Such verification shall not exceed

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

the amount of inspection required herein of the contractor. When Standard MIL-STD-105 specifies actions by the Government, the Government may authorize the contractor to perform any of such actions except that responsibility for acceptance rests with the Government.

4 9.1 **Preservation and packaging** Inspection of preservation and packaging shall be as specified in Specification MIL-P-116 Classification of defects shall be as shown on table III.

Table III. **Preservation, and packaging and marking thereof**

Major	**Minor**
1 Use of improper or defective material.	1. Item not properly blocked or braced within the unit package to prevent movement
2. Quantity in unit package not as specified.	2 Packaging material damaged
3. Incorrect packaging method applied.	3. Any item of marking information other than 7 through 9 omitted, incorrect, or illegible.
4 Cushioning or padding omitted.	
5 Cushioning inadequate for the physical and mechanical protection of the item.	
6. Unsealed, punctured, or improperly sealed barrier bag, wrap, or envelope.	
7. Stock number omitted, incorrect, or illegible.	
8 Nomenclature omitted, incorrect, or illegible.	
9 Marking of quantity of items in package omitted, incorrect or illegible	
10 Different stock numbered items in the same unit package	

4 9.2 **Packing and marking of exterior containers.** Packing and marking of exterior containers shall be given visual inspection for the defects listed below and to determine conformance with the approved process sheet furnished by the contractor as required by the contract. This inspection shall conform to the appendix to Standard MIL-STD-105. Inspection level L-8 shall be used for normal inspection and L-6 for reduced inspection. Unless otherwise specified herein, normal inspection shall be used at the start of the contract The reduced inspection procedure shall be R-1. The AQL for major defects shall be 4 percent and the AQL for minor defects shall be 10 percent Classification of defects shall be as shown on table IV.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

Table IV. **Packing and marking thereof**

Major	Minor
1. Use of improper or defective material.	1. Unsealed carton.
2 Quantity in pack not as specified	2. Defective taping or sealing of carton.
3 Gross weight in excess of specified amount.	3 Any other box defect which may be considered minor by definition of Standard MIL-STD-105.
4 Box closure not as specified.	4 Any item of required marking information other than 10 through 16 listed under major defects omitted, incorrect or illegible
5 Type, grade, class, and style of the shipping container not as specified.	
6 Strapping omitted.	
7 Strapping inadequate or incorrectly applied	
8 Items not adequately blocked, braced, or cushioned within the shipping container to prevent movement or damage.	
9 Shipping documents or packing list omitted.	
10 Stock number omitted, incorrect, or illegible	
11 Nomenclature omitted, incorrect, or illegible.	
12. Marking of quantity of items in pack omitted, incorrect, or illegible.	
13 Destination marking omitted, incorrect, or illegible	
14 Service designation (color marking) omitted	
15 Specified special marking and labeling such as NAP labels, shipment digit markings, etc omitted, incorrect, or illegible	
16. Overseas code marking omitted, incorrect, or illegible.	

4.9.3. **Inspection lot.** A lot of visual inspection of the pack shall be all completed packs which are identical and will be submitted for acceptance at one time.

4.9.4 **Procedure in case of failure** If an inspection lot fails, the contractor shall immediately investigate the cause of failure and shall report to the Government inspector the results thereof and details of the corrective action taken. If the contractor and Government inspector cannot agree on the effectiveness of the corrective action, the matter will be referred to the contracting officer for resolution.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

4.9.5 **Disposition of nonconforming product.** Disposition of nonconforming product (sample units and inspection lots) shall be in accordance with the requirements of Standard MIL-STD-105 for disposition of rejected product. When submitted for acceptance such product shall be suitably tagged or identified by equivalent means to indicate the cause of failure and means employed to correct the fault. The required information shall be presented to the Government when the product is submitted and shall become the property of the Government

5. **PREPARATION FOR DELIVERY**

5.1 **Preservation and packaging.**

5.1.1 Level A

5.1.1.1 Unit packaging.

5.1.1.1.1 **Test set** The set shall be packaged method 15-5 in accordance with Specification MIL-P-116 and as follows Each cord shall be coiled to the minimum safe diameter and secured with test plug and probes in the clips provided in the case. Cushion each case on all surfaces with cells or pads fabricated from fiberboard meeting the requirements of class 2, W5c of Specification PPP-B-636. Each set shall be packaged in a close-fitting fiberboard box conforming to class 2 or 3 of Specification PPP-B-636 The box shall be sealed in accordance with the appendix of the box specification using tape conforming to Specification PPP-T-76. The technical manual shall be secured to the lid of the box with tape as specified above.

5.1.1.1.2 **Technical manual.** Each technical manual shall be packaged method 10-3 in accordance with Specification MIL-P-116 using a close-fitting heat-sealed bag conforming to type II, class b of Specification MIL-B-117

5 1.2 **Level C** The test set shall be packaged in accordance with commercial practice and in a manner that will afford protection against corrosion, deterioration, and physical damage during direct shipment to the first receiving activity.

5 2 **Packing** Packing shall be level A, B, or C as specified (see 6 2)

5.2.1 **Level A** The sets shall be packed in a close-fitting wirebound wood box, wood-cleated fiberboard box, wood cleated-plywood box or a fiberboard box conforming to Specification PPP-B-585, class 3, PPP-B-591 (overseas-type), PPP-B-601, style A or B, using plywood conforming to type II or III, class 2 of Specification NN-P-515, PPP-B-621, class 2 and PPP-B-636, class 2 or 3, respectively Closure and strapping requirements shall be as specified in the applicable box specification or appendix thereto. The gross weight of wood or wood-cleated boxes shall not exceed approximately 200 pounds. The gross weight of fiberboard boxes shall not exceed the weight limitations of the applicable box specification.

5.2.2. **Level B.** The sets shall be packed in a close-fitting wirebound woodbox, wood-cleated fiberboard box, wood cleated-plywood box, nailed woodbox or fiberboard boxes conforming to Specification PPP-B-585, class 1, PPP-P-591 (domestic-type), PPP-B-601 (domestic-type), PPP-B-621, class 1, PPP-B-636,

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

class 1 and PPP-B-640, respectively. Closure shall be as specified in the applicable box specification or appendix thereto. The gross weight of wood or wood-cleated boxes shall not exceed approximately 200 pounds. The gross weight of fiberboard boxes shall not exceed the weight limitations of the applicable box specification

5.2.3. **Level C.** The test set shall be packed for shipment in a manner, conforming to the requirements of Uniform Freight Classification No 4 for rail shipment National Motor Freight Classification No A-3 and No. 14 for truck shipment; Parcel Post Regulations, and the regulations of other carriers as applicable to the mode of transportation employed, at the lowest transportation rate.

5.3 **Marking.** Interior packages and exterior shipping containers shall be marked in accordance with the applicable provisions of Standard MIL-STD-129

6 **NOTES**

6 1 **Intended use.** Battery Tester TS-183( )/U is used to determine the serviceability of dry batteries ranging in voltage rating from 1 5 volts to 200 volts.

6.2 **Ordering data.** Procurement documents should specify the following

(a) Title, number, and date of this specification and any amendment thereto

(b) Type required.

(c) Level of packaging and level of packing for shipment (level A, level B, or level C).

(d) The specific paragraphs of section 5 which are applicable to the particular procurement.

(e) Preproduction inspection

(1) Two samples of each item cited in section 1 are generally required so that lengthy environmental tests can be completed on one sample while complete performance measurements can be made on the second sample

(2) Preproduction pack(s) as follows

a Makeup of pack(s)

b. Number of each kind of pack to be submitted.

c. Inspection to be performed thereon-including rough handling, which will not be performed as procurement inspection.

d. Five copies of process sheet, using format shown on figure 1 and accompanied by such drawings, sketches, and other data as may be necessary to completely describe the

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

procedure followed in fabrication of the preproduction pack(s) and to identify the items therein These data should be submitted to the contracting officer with the preproduction pack.

e. A packaging technician from the procuring agency will evaluate the preproduction pack(s)

(f) Marking and shipping of samples

(g) Place of final inspection

(h) Technical literature required.

6.3 **Nomenclature.** The parentheses in the nomenclature will be deleted or replaced by a letter identifying the particular design; for example: TS-183X/U. The contractor should apply for nomenclature in accordance with the applicable clause in the contract.

6 4 **Definition of level B.** Level B packaging requirements detailed herein are designed to furnish protection against corrosion, deterioration, and physical damage during domestic shipments; covered storage for a period of 3 years or less, and distribution within the continental limits of the United States

6.5 **Verification inspection.** The amount of verification inspection (by the Government) will be adjusted to make maximum utilization of the contractor's quality control system and the quality history of the product, and will normally be identified by the category listed below:

(a) Type A - The total of that inspection set forth in the Quality Assurance Provisions of this specification or the contract. Included in this category is that amount of inspection referred to as normal and tightened inspection by Military Standard 105.

Notice When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever, and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

Custodians:
Army - Signal Corps
Navy - Marine Corps
USAF - Air Force

Other interests:
Army - CE.Ord.
Navy

Preparing activity:
Signal Corps

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

PACKAGING, PACKING, AND MARKING PILOT PACK PROCESS SHEET

CONTRACTOR ORDER NO

STREET ADDRESS EQUIPMENT

CITY AND STATE LEVEL

DATE PILOT PACK COMPLETED FOR EVALUATION

ITEM NAME								
PACKAGING METHOD								
CLEANING	SPECIFICATION							
	MATERIAL DESCRIPTION	T	G	C	T	G		
PRESERVATION	SPECIFICATION							
	MATERIAL DESCRIPTION	T	G	C	T	G		
CUSHIONING	SPECIFICATION							
	MATERIAL DESCRIPTION	T	G	C	T	G		
	SIZE							
SECURING	SPECIFICATION							
	MATERIAL DESCRIPTION	T	G	C	T	G		
	SIZE							
BARRIER	SPECIFICATION							
	MATERIAL DESCRIPTION	T	G	C	T	G		
	SIZE							
DESICCANT	SPECIFICATION							
	MATERIAL DESCRIPTION	T	G	C	T	G		
	SIZE							
UNIT CARTON	SPECIFICATION							
	MATERIAL DESCRIPTION	T	G	C	T	G		
	SIZE							
INTERMEDIATE CARTON	SPECIFICATION							
	MATERIAL DESCRIPTION	T	G	C	T	G		
	SIZE							
OUTER CARTON (CONSOLIDATED OR IIB)	SPECIFICATION							
	MATERIAL DESCRIPTION	T	G	C	T	G		
	SIZE							
SECURE CLOSURE	SPECIFICATION							
	MATERIAL DESCRIPTION	T	G	C	T	G		
	SIZE							

CASE LINER	SPECIFICATION	
	SEALING MATERIAL	
	SIZE	
SHIPPING CONTAINER	SPECIFICATION	
	STYLE	
	SIZE	
STRAPPING	SPECIFICATION	
	MATERIAL DESCRIPTION	
	SIZE	

LEGEND FOR MATERIAL T—TYPE, G—GRADE, C—CLASS

FIGURE 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Bureau No 119 R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor involved in the use of the specification in procurement of products for ultimate use by the Department of the Navy. This sheet is provided for obtaining information on the use of this specification for which will insure that such products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner and send to preparing activity (as indicated on reverse hereof).

SPECIFICATION

ORGANIZATION (Of submitter) | CITY AND STATE

CONTRACT NO | QUANTITY OF ITEM PROCURED | DOLLAR AMOUNT $

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1 HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT?

A GIVE PARAGRAPH NUMBER AND WORDING

B RECOMMENDATION FOR CORRECTION OF THE DEFICIENCIES

2 COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3 IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO IF "YES" IN WHAT WAY

4 REMARKS (Attach any pertinent data which may be of use in improving this specification. If there are additional papers attach to form and place both in an envelope addressed to preparing activity)

SUBMITTED BY (Printed or typed name and activity) | DATE

DD FORM 1426 REPLACES NAVSHIPS FORM 4863 WHICH IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.

FOLD

DEPARTMENT OF THE NAVY

POSTAGE AND FEES PAID
NAVY DEPARTMENT

OFFICIAL BUSINESS

COMMANDING GENERAL
U S ARMY ELECTRONICS COMMAND
ATTN: AMSEL-PP-PTS
FORT MONMOUTH, N.J. 07703

FOLD

PLATE NO 19419 (DAC

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T23:17Z
Check the source to verify that this is the current version before use.