MIL-B-42017/12
22 September 1986

MILITARY SPECIFICATION SHEET

BUR, BONE SURGERY (TUNGSTEN CARBIDE, 5.5 MM HEAD DIAMETER BY 10 MM CUTTING LENGTH, OVAL CUTTING-SHAPE D, LONG SHANK-LENGTH 3)

The complete requirements for acquiring the bur described herein shall consist of this specification sheet and the latest issue of MIL-B-42017.

REQUIREMENTS:

Shall be an oval cutting, long shank, surgical buffing bur for use with the long bur guard in National Stock Number (NSN) 6515-01-116-8197, DRILL, BONE, PNEUMATIC.

Shall be of two-piece construction. The head shall be made of C1 or C2 grade tungsten carbide with a grain size of 1 to 2 microns. The shank shall be made of C13 or C14 grade tungsten carbide.

Shall be examined at 7X magnification for bur defects not readily perceptible to the naked eye.

Shall be of the design and dimensions shown in Figure 1 (7755).

NOTES:

This specification sheet covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6515-01-139-7755	BUR, BONE SURGERY, Oval Cutting, Tungsten Carbide, 0.21 inch (5.5 mm) Head Diameter,0.39 inch(10 mm) Cutting Length: Non-notched shank.

Agent:
DLA-DM

Preparing activity:
DoD-MB

(Project No. 6515-3806)

AMSC N/A

FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:12Z
Check the source to verify that this is the current version before use.




Figure 1 (7755). Bur, Bone Surgery, 5.5 mm head diameter by 10 mm cutting length, Shape D, Length 3
NSN: 6515-01-139-7755

Source: https://assist.dla.mil -- Downloaded: 2026/10-21T20:12Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
MIL-B-42017 (12)	BUR, BONE SURGERY (Tungsten Carbide, Shape D, Length 3)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* - Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* - Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* - Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:12Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in bmitting comments and suggestions for improvements. All users of military standardization documents are invited to provide ggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

Fold along this line)

DEPARTMENT OF THE ARMY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 12062 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEPARTMENT OF THE ARMY

Commander
Defense Personnel Support Center
ATTN: DPSC-RST
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:12Z
Check the source to verify that this is the current version before use.