Inch-Pound

MIL-C-42020A
29 March 1989
SUPERSEDING
MIL-C-42020(DM)
1 October 1986

MILITARY SPECIFICATION

# CURETTE SET, ANTRUM, COAKLEY

This specification is approved for by all Departments and Agencies of the Department of Defense.

## 1. SCOPE

1.1 Scope. This specification covers a set of six antrum curettes of the Coakley design.

## 2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications, standards and handbooks. The following specifications, standards and handbooks form part of this document to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DODISS) and supplement thereto, cited in the solicitation (see 6.2).

SPECIFICATIONS

FEDERAL

PPP-B-601 - Boxes, Wood, Cleated-Plywood.
PPP-B-621 - Boxes, Wood, Nailed and Locked-covered.
PPP-B-636 - Box, Fiberboard.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, by using self-addressed Standardization Document Improvement Proposal, (DD Form 1426) appearing at the end of this document or by letter.

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

MILITARY

MIL-P-116 - Preservation-Packaging, Methods of.
MIL-L-10547 - Liner, Case and Sheet, Overwrap, Water Vaporproof or Waterproof, Flexible.

STANDARDS

FEDERAL

Federal Test Method Standard No. 151 - Metals; Test Methods.

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-147 - Palletized Unit Loads.
MIL-STD-129 - Marking for Shipment and Storage.
MIL-STD-794 - Parts and Equipment, Procedures for Packaging of.

(Unless otherwise indicated, copies of federal and military specifications, standards and hardbooks are available from the Naval Publications and Forms Center, (ATTN: NPODS), 5801 Tabor Avenue, Philadelphia, Pa 19120-5099.)

2.1.2 Other Government documents, drawings, and publications. The following other Government documents, drawings, and publications form a part of this document to the extent specified herein. Unless otherwise specified, the issues are those cited in the solicitation.

DRAWINGS

Defense Personnel Support Center (DPSC)

No. 24538 Curette Set, Antrum, Coakley

(Copies of other Government documents, drawings and publications are available from Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101.)

2.2 Non-Government publications. The following documents form a part of this document to the extent specified herein. Unless otherwise specified, the issues of the documents which are DoD adopted are those listed in the issue of the DODISS specified in the solicitation. Unless otherwise specified, the issues of documents not listed in the DODISS are the issues of documents cited in the solicitation (see 6.2).



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM) STANDARDS

ASTM E 18 - Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials, Tests for.

ASTM E 92 - Vickers Hardness of Metallic Materials, Tests for.

ASTM E 140 - Hardness Conversion Tables for Metals (Relationship Between Brinell Hardness, Vickers Hardness, Rockwell Hardness, Rockwell Superficial Hardness and Knoop Hardness).

ASTM F 899 - Stainless Steel Billet, Bar and Wire for Surgical Instruments.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103-1187).

(Nongovernment standards and other publications are normally available from the organizations that prepare or which distribute the documents. These documents also may be available in or through libraries or other informational services.)

2.3 Order of precedence. In the event of a conflict between the text of this document and the references cited herein, the text of this document shall take precedence. Nothing in this document, however, shall supersedes applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 Material. Curettes shall be manufactured from class 4 martensitic corrosion-resistant steel. As an alternate, handles may be manufactured from Class 3 austenitic corrosion-resistant steel. Steels shall be as specified in American Society for Testing and Materials (ASTM) Designation F899. Composition shall be tested in accordance in 4.5.1.

3.2 Hardness. The curette set shall be properly passivated and the loops shall have a Rockwell Hardness C scale (HRC) of 40 minimum. Hardness shall be tested as specified in 4.5.2.

3.3 Style, design and dimensions. Shall be a set of 6 antrum curettes of the Coakley design, used in ENT procedures. The set shall consist of the following curettes with the designated style:



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

| | | |
|---|---|---|
| Style 1 | - | Large, light curve |
| Style 2 | - | Large, strong curve |
| Style 3 | - | Large, acute angle |
| Style 4 | - | Small, light curve |
| Style 5 | - | Small, strong curve |
| Style 6 | - | Triangular |

The curette set shall be in accordance with Defense Personnel Support Center (DPSC) Drawing 24538.

3.4 Construction. Construction shall be as specified herein.

3.4.1 Handles and shanks. The handles and shanks shall be fabricated as a single component or may be designed as a two piece assembly, with solid or hollow handles. When constructed of more than one piece, the shanks shall be securely attached to the handles.

3.4.2 Loops. The loops shall be integral to or securely and firmly attached to the shanks. Where the loops enter the shanks and for all curettes, there shall be no crevices or spaces where foreign material may collect or which may be subject to corrosion.

3.5 Finish. The shanks and loops shall have a mirror finish. Handles shall be knurled to ease in the handling of the curettes. The finish shall be uniform and free from grind or tool marks.

3.6 Corrosion. Curettes shall show no signs of corrosion when tested for the boil and copper sulfate tests specified in ASTM Designation F1089.

3.7 Markings. The curettes shall bear the manufacturer's name or registered trademark and the word "stainless." In addition, retractors of foreign manufacture shall bear the name of the country of manufacture. Each curette shall be marked with its style designation number.

3.8 Workmanship. The curettes shall be free from any defect which detract from its appearance or impair its serviceability. The curettes shall be without rough edges, burrs, cracks, pits, dents and corrosion.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

## 4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements (examinations and tests) as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Responsibility for compliance. All items must meet all requirements of sections 3 and 5. The inspection set forth in this specification shall become a part of the contractor's overall inspection system or quality program. The absence of any inspection requirements in the specification shall not relieve the contractor of the responsibility of ensuring that all products or supplies submitted to the Government for acceptance comply with all requirements of the contract. Sampling inspection, as part of manufacturing operations, is an acceptable practice to ascertain conformance to requirements, however, this does not authorize submission of known defective material, either indicated or actual, or does it commit the Government to accept defective material.

4.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc. investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties), of the item.

4.2.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.2.2 Certificate of quality. Certificate of quality supplied by the manufacturer of the steel may be furnished in lieu of actual performance of composition testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number, the item identification, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.3 Sampling.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

4.3.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as indicated in table I. Unit of product for examination purpose shall be one curette set.

TABLE I. Sampling criteria for examination.

| Type of defect | Inspection level | AQL (percent defective) |
|---|---|---|
| Major | II | 1.0 |
| Minor | II | 2.5 |

4.3.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in table II. The unit of product for test purposes shall be one curette set. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

TABLE II. Sampling criteria for tests.

| Characteristic | Requirement | Test procedure | Inspection level | AQL (percent defective) |
|---|---|---|---|---|
| Composition | 3.1 | 4.5.1 | 1/ | 2/ |
| Hardness | 3.2 | 4.5.2 | S-2 | 1.0 |
| Dimensions | 3.3 | --- | S-2 | 2.5 |
| Corrosion Resistance | ASTM F1089 | ASTM F1089 | S-2 | 1.0 |

1/ A sufficient amount of raw material shall be tested.

2/ The lot shall be rejected if test requirements are not met.

4.4 Examination.

4.4.1 Examination of curette set. The curette sets shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in tables I and III. Examination, however, shall not be restricted to defects listed in table III. Any curette set in the sample containing one or more defects shall be rejected and if the number of defective curette sets in any sample exceeds the acceptance number for the sample, the lot represented by that sample shall be rejected.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects.

| Categories | Defects |
|---|---|
| Critical | None defined |
| Major | |
| 101 | Curette set not of the specified design for appropriate size, stock number and item identification. |
| 102 | Curette set not free from cracks, pits, burrs, sharp edges, scale, rough areas or corrosion. |
| 103 | Finish of curette set not uniform and not as specified. |
| 104 | Curette set has grind or tool marks. |
| 105 | Identification markings not present, not complete or not legible. |
| Minor | |
| 201 | Curette set not free from dirt, polishing material or foreign matter. |

4.5 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.5.1 Composition. Raw material shall be tested for composition in accordance with Federal Test Method 111 or 112 of Federal Test Method Standard Number 151 (see 4.2.2).

4.5.2 Hardness test. Hardness shall be determined in accordance with ASTM Standards E 18 or E 92. Conversion of Vickers hardness values to Rockwell values shall be in accordance with ASTM Standard E 140.

4.5.3 Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

4.6 Inspection of packaging. The sampling and inspection of the preservation and interior package marking shall be in accordance with groups A and B quality conformance inspection requirements of MIL-P-116. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit package. Curettes comprising each set shall be clean, dry, individually wrapped in neutral tissue and packaged in accordance with MIL-P-116 method IC-2 except that a preservative shall not be used.

5.1.1.2 Intermediate package. Twelve unit packages shall be packaged in a box conforming to PPP-B-636, class domestic. Closure shall be as specified in the box specification.

5.1.2 Level C.

5.1.2.1 Unit package. Preservation for each curette set shall conform to the requirements of MIL-STD-794 for level C.

5.1.2.2 Intermediate package. Twelve unit packages shall be packaged in accordance with the requirements of MIL-STD-794 for level C.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Four intermediate packages (48 units) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type, PPP-B-621, class 2, or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. Four intermediate packages (48 units) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type, PPP-B-621, class 1 or PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

5.2.3 Level C. Four intermediate packages (48 units) shall be packaged in accordance with the requirements of MIL-STD-794 for level C.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination exceeds 250 pounds (excluding the pallet) or 20 cubic feet. Loads shall be unitized as specified in MIL-STD-147, on Type IV or Type V 4-way entry pallets. Pallet shall have a length of 40 inches and a width of 48 inches. Pallet loads, including the pallet, shall not exceed 54 inches in height, 43 inches in length and 52 inches in width. Quantity for shipment to one designation of less than 250 lbs. or 20 cubic feet need not be palletized.

5.2.5.1 Levels A and B. Curette sets packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets as specified in 5.2.5.

5.2.5.2. Level C. Curette sets packed as specified in 5.2.3 shall be unitized as specified in MIL-STD-794.

5.3 Marking.

5.3.1 Levels A, B and C. Each unit pack, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.6.

6. NOTES
(This section contains information of general or explanatory nature that may be helpful, but is not mandatory.)

6.1 Intended use. The curette set shall be used for curettage of the antrum in ENT surgical procedures.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

6.2 Acquisition requirements. Acquisition documents must specify the following:

a. Title, number, and date of the specification.
b. Issue of DODISS to be cited in the solicitation, and if required, the specific issue of individual documents referenced (see 2.1.1 and 2.1.2).
c. Levels of packaging and packing (see 5.1 and 5.2).
d. Applicable National Stock Number.

6.3 National Stock Number Coverage. This specification covers the following National Stock Numbers (NSN):

| National Stock No. | Item Description |
|---|---|
| 6515-01-110-1382 | CURETTE SET, ANTRUM, COAKLEY<br>Set of 6 |

6.4 Subject term (keyword) listing:

Cleaner
Instrument set
Scooper
Scraper
Surgical set

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4458



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

*(See Instructions – Reverse Side)*

| 1. DOCUMENT NUMBER | 2. DOCUMENT TITLE |
|---|---|
| MIL-C-42020A | CURETTE SET, ANTRUM, COAKLEY |

| 3a. NAME OF SUBMITTING ORGANIZATION | 4. TYPE OF ORGANIZATION *(Mark one)* |
|---|---|
| | ☐ VENDOR |
| b. ADDRESS *(Street, City, State, ZIP Code)* | ☐ USER |
| | ☐ MANUFACTURER |
| | ☐ OTHER *(Specify)*: |

**5. PROBLEM AREAS**

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

**6. REMARKS**

| 7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional | b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional |
|---|---|
| c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional | 8. DATE OF SUBMISSION *(YYMMDD)* |

*(TO DETACH THIS FORM, CUT ALONG THIS LINE.)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.