MILITARY SPECIFICATION

CATHETER, CARDIOVASCULAR

This specification is approved for use within the Defense Personnel Support Center and is available for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers various cardiovascular catheters for use with contrast media.

1.2 Classification.

	Type	Catheter Size (French)	Length (cm)	Balloon Size Uninflated: Diameter, Fr	Inflated Volume, Ml
1	Arterial Embolectomy	3	80	3-4	0.25 ± 0.05
	Arterial Embolectomy	5	80	5-6	0.65 ± 0.10
2	Venous Thrombectomy	8	80	9-10	1.75 ± 0.50
3	Occlusion	8	80	13-14	18 ± 2

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications and standards. The following specifications and standards form a part of the specification to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DoDISS) and supplement thereto, cited in the solicitation.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-RST, 2800 S. 20th Street, Philadelphia, PA 19101, by using self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

AMSC/NA FSC 6515
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:16Z
Check the source to verify that this is the current version before use.

AMERICAN NATIONAL STANDARDS INSTITUTE (ANSI)

ANSI/HIMA MD70.1 - American National Standard for Medical Materiel, Luer Taper Fittings, Performance.

(Applications for copies shall be addressed to Publication Sales, Dept. STD, American Institute of Physics, 335 East 54th St., NY, NY 10017).

UNITED STATES PHARMACOPEIAL CONVENTION, INC.

Pharmacopeia of the United States (USP).

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042).

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence. Nothing in this specification, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 Material. All rubber and plastic components shall be sterile, pyrogen-free, and non-toxic when tested as specified in section 4. Plastics shall meet the requirements of the USP for Class IV plastics. Components which contact patient's blood shall contain not more than 5 total parts per million (ppm) of lead, barium and cadium.

3.1.1 Plastic tubing. The plastic tubing shall be polyvinyl chloride suitably pliable for atraumatic navigation through blood vessels, but sufficiently rigid to prevent collapse during normal usage.

3.1.2 Balloon. The balloon shall be made from non-oxidizing, low-sulfur, wax-free latex rubber.

3.2 Style and design. The cardiovascular catheter shall be of double-lumen construction comprising one each for the catheter lumen and the balloon. Catheter French size shall be as specified in 6.3. The distal end shall have a soft, smooth tip to minimize vessel trauma. A balloon shall be permanently attached by epoxied sutures to the distal end. The catheter shall have markings each ten centimeters from the tip to the connector.

3.3 Security of balloon attachment. When tested as specified in 4.4.2, the inflated balloon shall not separate from the catheter.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:16Z
Check the source to verify that this is the current version before use.

4.1.3 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.4 Certificate of quality. Certificate of quality supplied by the manufacturer of the materials of components may be furnished in lieu of actual performance of composition testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number, the item identification, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in Table I. The unit of product for examination shall be one complete catheter.

TABLE I. Sampling for examination.

Type of defect	Inspection level	AQL (percent defective)
Major	S-2	1.5
Minor	S-2	4.0

4.2.2 For tests. Sampling for tests shall be in accordance with MIL-STD-105 and as indicated in table II. The unit of product for test purposes shall be one complete catheter.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:16Z
Check the source to verify that this is the current version before use.

4.4 Tests.

4.4.1 Class IV plastic. The catheter shall be tested in accordance with the USP for Class IV plastics. Sampling shall be as shown in Table II.

4.4.2 Security of balloon attachment. The balloon shall be suspended vertically through a funnel with the balloon fully inflated as the only means of support. The following forces shall then be suspended from the proximal side of the balloon for 30 seconds.

Size (French)	Force (lbs.)
3	2
4	3
5	4
6	5
7	5
8	5

4.4.3 Balloon leakage. The balloon shall be fully inflated with air, immersed in water and examined for leaks over a period of 10 seconds.

4.4.4 Luer hub leakage. The Luer hub of the catheter shall be tested for leakage in accordance with ANSI/HIMA MD70.1.

4.4.5 Saline resistance. The balloon shall be placed in USP isotonic saline solution at 37 ± 1°C for 72 hours. Afterward, the balloon shall be fully inflated and deflated with air once every 20 minutes for 8 hours.

4.4.6 Sterility test. The catheter shall be tested for sterility in accordance with the method specified in the USP using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates it from product intended for distribution. The incubation medium for the paper spore strip (carrier) shall be soybean-casein digest. The incubation media for the inoculated product, or inoculated simulated product shall be fluid thioglycollate or alternative thioglycollate at 30° to 35°C for one half of the specimens and soybean-casin digest at 20° to 25°C for the remaining half of the specimens. The minimum incubation time shall be ten days. Sampling shall be as follows:

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:16Z
Check the source to verify that this is the current version before use.

4.4.7 Freedom from pyrogens and non-toxicity. Samples of the catheters shall be tested for freedom from pyrogens and non-toxicity in accordance with the USP. Sampling shall be shown in Table II.

4.5 Inspection of packaging. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit package. Each catheter shall be packaged in a see through rigid plastic tray having a rigid plastic lid so designed to avoid crushing of the catheter. The balloon shall be provided with a protector for protection from the atmosphere. The tray shall be further packaged in a sealed peel-open type bag capable of maintaining sterility of the contents unless the bag is damaged or opened. The bag shall be constructed of tyvek having a clear plastic backing.

5.1.1.2 Intermediate package. Five unit packages shall be packaged in a box conforming to PPP-B-566, PPP-B-636, class domestic or PPP-B-676. Closure shall be as specified in the applicable box specification.

5.1.2 Level C. The level C preservation for catheters shall conform to MIL-STD-794 requirements for this level.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Four intermediate packages (20 units) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be liineend with a waterproof case linerer conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. Four intermediate packages (20 units) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 202610-21T20:16Z
Check the source to verify that this is the currentversion before use.

6.1.3 Type 3. Type 3 catheters shall be used in vessel tamponade procedures.

6.2 Ordering data.

6.2.1 Acquisition requirements. Acquisition documents should specify the following:

(a) Title, number, and date of this specification.
(b) Type and size required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(d) Applicable National Stock Number.

6.3 This specification covers the following National Stock Numbers:

National Stock No.	Item Identification
6515-01-115-1726	CATHETER, CARDIOVASCULAR, Fogarty, Arterial Embolectomy, Plastic, 3 Fr
6515-01-115-4710	CATHETER, CARDIOVASCULAR, Fogarty, Arterial Embolectomy, Plastic, 5 Fr
6515-01-115-1725	CATHETER, CARDIOVASCULAR, Fogarty, Occlusion, Plastic, 8/10 Fr
6515-01-115-1724	CATHETER, CARDIOVASCULAR, Fogarty, Venous Thrombectomy, Plastic, 8/14 Fr

6.4 Subject term (key word) listing.

Cardiovascular catheter
Arterial-Embolectomy
Venous Thrombectomy
Occlusion

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:16Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 202640-21T20:16Z
Check the source to verify that this is the current version before use.