MILITARY SPECIFICATION

INTRAVENOUS INJECTION SET, PEDIATRIC, DISPOSABLE

This specification is approved for use within Defense Personnel Support Center and is availabe for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers an intravenous injection set for use in administration of fluid or blood into the scalp vein of a premature infant, infant, young child or into any small vein.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications and standards. The following specifications and standards form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DoDISS) and supplement thereto, cited in the solicitation.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-RST, 2800 S. 20th Street, Philadelphia, PA 19101, by using self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

AMSC-N/A FSC 6515
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

FEDERAL

PPP-B-566 - Boxes, Folding, Paperboard.
PPP-B-601 - Boxes, Wood, Cleated-plywood.
PPP-B-621 - Boxes, Wood, Nailed and Lock-corner.
PPP-B-636 - Boxes, Shipping, Fiberboard.
PPP-B-676 - Boxes, Set-up.
GG-N-196 - Needle, Hypodermic for Luer Syringes.

MILITARY

MIL-L-10547 - Liner, Case and Sheet, Overwrap, Water Vaporproof or Waterproof, Flexible.
MIL-S-36586 - Sterilization Test Strip Set Bacterial Spore.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.
MIL-STD-147 - Palletized Unit Loads.
MIL-STD-794 - Parts and Equipment, Procedures for Packaging of.

(Copies of specifications and standards required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of the documents which are DoD adopted shall be those listed in the issue of the DoDISS specified in the solicitation. Unless otherwise specified, the issues of documents not listed in the DoDISS shall be the issue of the nongovernment documents which is current on the date of the solicitation.

AMERICAN NATIONAL STANDARDS INSTITUTE (ANSI)

ANSI/HIMA MD70.1 - American National Standard for Medial Materiel, Luer Taper Fittings, Performance.

(Applications for copies shall be addressed to Publication Sales, Dept. STD, American Institute of Physics, 333 East 54th St., NY, NY 10017).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM) STANDARDS.

ASTM, D2240 - Standard Test Method for Rubber Property - Durometer Hardness

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race St., Philadelphia, PA 19103).

UNITED STATES PHARMACOPEIAL CONVENTION, INC.

Pharmacopeia of the United States (USP).

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042).

(Nongovernment standards and other publications are normally available from the organizations which prepare or which distribute the documents. These documents also may be available in or through libraries or other informational services).

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence. Nothing in this specification, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 Material.

3.1.1 Plastic components. All plastic components shall be of the best quality and suitable in all respects for the intended purpose. All components shall be sterile, free from pyrogens, compatible and safe when tested as specified in section 4. Those plastic components coming in contact with solutions being administered and end protectors shall contain not more than 5 parts per million total lead, barium and cadmium when tested as specified in 4.5.1 or 4.5.1.1. The plastics shall meet the requirements of the Pharmacopeia of the United States (USP) for Class IV plastics and for transfusion and infusion assemblies when tested as specified in 4.5.2.

3.1.1.1 Plastic tubing. Plastic tubing shall be clear, transparent or translucent, flexible polyvinyl chloride having a hardness of 68 + 6, -5 Shore A durometer, when tested as specified in 4.5.3.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

3.1.2 Finger grip device. The finger grip device shall be fabricated from flexible polyvinyl chloride.

3.1.3 Needles. The needle shall be fabricated from AISI 304 stainless steel. Each needle shall meet the corrosion-resistance, elasticity and resistance-to-breakage tests as specified in GG-N-196. The cannulae shall be fabricated from either seamless drawn or welded and drawn tubing. The needle shall be treated with a non-toxic, non-pyrogenic silicone.

3.1.4 Other material. All other material specified shall be of the best quality and suitable in all respects for the purpose intended.

3.2 Intravenous injection set. The intravenous injection set shall consist of a length of tubing with a vein needle cannula. The set shall have a finger grip device attached behind the needle and a female Luer fitting attached to the other end of the tubing. The assembly of the set shall be structurally sound and shall not leak fluid during use or when tested in accordance with 4.5.7.

3.2.1 Tubing. Each set shall be supplied with 12 ± 1/2 inches of plastic tubing between the finger grip and the female Luer fitting. The tubing shall have an inside diameter of 0.040 inch + .006, - .000 inch with a wall thickness of 0.015 inch + 0.006, -0.000 inch.

3.2.2 Intravenous needle. The needle shall be a 25 gage by 3/4 inch needle in accordance with the requirements of GG-N-196 for Type I, point style b, except for stiffness. The needle marking required by GG-N-196 may be omitted. The non-pointed end of the cannula shall be permanently secured in a leakproof manner to the plastic tubing when tested as described in 4.5.4. The cannula shall be a straight tube with a short bevel point on the distal end. The point shall be ground concave or flat and may have a center ridge at the distal end to strengthen the tip. The point shall be so ground at the outer end as to enhance the keeness of the tip and cutting edges. Both outer and inner edges of the point shall be free from burrs and slivers. The heel of the point shall be dull to prevent coring. The tip of the point shall be straight and strong, and of maximum keeness. The proximal end of the cannula shall extend well into the hub sufficiently to allow for secure staking. The cannula shall be centered in the distal end of the hub in such a manner as to provide a straight continuous fluid path from the hub cavity. The cannula shall not deviate from the central longitudinal axis of the needle by more than 3 degrees. The exterior of the cannula shall be clean, smooth and free from pits, deep furrows and die marks.

Source: https://assist.dla.mil -- Downloaded: 202610-21T20:14Z
Check the source to verify that this is the currentversion before use.

3.2.3 Finger grip device. A plastic finger grip shall be securely attached to the cannula. As an alternate, the finger grip may be permanently attached to the needle cannula. The finger grip shall have flexible, flat wings which can be bent to facilitate insertion of the needle into the vein. The wings shall lie flat to permit taping to the skin in such a manner that the needle shall not roll in the vein or be dislodged during infusion. The grip shall not have a stiff, upright middle wing. The plane of the wings shall be parallel to the lumen of the tubing and to the short axis of the needle bevel plus or minus 30 degrees. The length of the wings shall not be less than 0.35 inch nor more than 0.55 inch. The width of the wings shall not be less than 1.18 inches nor more than 1.50 inches. As an alternate, the width shall be 1-1/8 inches plus or minus 1/16 inch. The length shall be 7/16 inch plus or minus 1/16 inch. The wings shall be straight or swept back at an angle of 20 plus or minus 2 degrees. At the area of attachment to the cannula and tubing, the wing thickness shall not exceed 1/8 inch. The wing surfaces may be texture for ease of gripping, with the size and manufacturer identification molded in.

3.2.4 Female Luer fitting. The female Luer fitting shall be fabricated of a suitable rigid plastic. The female Luer slip fitting shall be in accordance with ANSI/HIMA MD70.1 as stated in 4.5.5. The Luer fitting shall not leak or separate when connected to a male Luer tip.

3.2.5 End protectors. Two protectors shall be supplied, one for each end of the injection set. The protector shall be attached to the needle in such a manner which will prevent contact of the needle point with the inside of the protector. End protectors shall be readily removable, plastic and constructed in such a manner that when placed over each open end of the injection set it shall, after sterilization, maintain sterility of the needle and the entire fluid path of the device.

3.3 Manufacturer's record. The manufacturer shall place a lot control number on the paper label or on the printed carton which can be verified against the manufacturer's records of all tests made available for inspection by any duly qualified medical officer of the Armed Forces, any official of the Food and Drug Administration, or any other duly authorized Government representative. No shipment shall be made until all tests have been completed, verified and recorded in the permanent record of the manufacturer.

3.4 Sterility, safety and freedom from pyrogens. All equipment shall be safe. Fluid paths, end connectors and needles shall be sterile and pyrogen free. The fluid path, end connectors and needles shall be pyrogen free when tested as specified in 4.5.8 and sterile when tested as specified in 4.5.9. The sets shall be safe when tested for safety as specified in 4.5.10.

Source: https://assist.dla.mil -- Downloaded: 202610-21T20:14Z
Check the source to verify that this is the currentversion before use.

3.5 Certificates. Certificate of sterility, freedom from pyrogens and safety shall be furnished to the procuring activity via the government inspector.

3.5.1 Certificate of sterility. A certificate of sterility, listing all sterility lot or control numbers shall state that samples from each sterility lot or control have been tested for sterility as required in 3.4 and were found to be sterile. The certificate shall reference the method used in determining sterility.

3.5.2 Certificate of freedom from pyrogens. A certificate of freedom, listing all lot or control numbers, shall state that samples from each lot or control have been tested for pyrogens as specified in 3.4 and were found to comply with such requirements. The certificate shall reference the method used in determining freedom from pyrogens.

3.5.3 Certificate of Safety. A certificate of safety, listing all lot or control numbers, shall state that samples from each lot or control have been tested for safety as specified in 3.4 and were found to comply with such requirements. The certificate shall reference the method used in determining safety.

3.6 Compatibility. The materials of construction of each intravenous injection set shall not react with, or be acted upon by, the liquids with which they will be in contact during normal use and operation. There shall not be any materials incorporated into the set which will cause any deleterious effect when the set is used for the purpose intended.

3.7 Leakage. Each set shall be tested for leakage and separation of the set as specified in 4.5.6. Sets which show leakage or separation shall be rejected. A leak is defined as the formation of an air (gas) bubble or liquid drop as appropriate, which breaks away from the set within the specified test time.

3.8 Flow path. Fluid path of the set shall be clear, clean and free from particles. Sets which are sealed or obstructed shall be rejected.

3.9 Workmanship. Workmanship shall be first class throughout. The intravenous injection set shall be free from defects which detract from their appearance or impair their serviceability.

3.10 Instructions. Each individual set shall have detailed instructions for the proper use of the set for single use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.2 Inspection of packaging for delivery. The sample unit shall be one exterior container. The sampling unit shall be in accordance with MIL-STD-105. The inspection level shall be S-4 with AQL of 2.5 expressed in terms of percent defective. The inspection of the packaging, packing and marking for shipment and storage shall conform to the quality assurance provisions of the applicable container specification and marking requirements of MIL-STD-129.

4.3 Sampling.

4.3.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in Table I. The unit of product for examination shall be one intravenous injection set.

TABLE I. Sampling for examination.

Type of defect	Inspection level	AQL (percent defective)
Major	S-4	1.0
Minor	S-4	2.5

4.3.2 For tests. Sampling for tests shall be in accordance with MIL-STD-105 and as indicated in table II. The unit of product for test purpose shall be one intravenous injection set.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

TABLE II. Sampling for other tests

Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Heavy Metal ** (extractable)	3.1.1	4.5.1	S-1	*
Class IV Plastic **	3.1.1	4.5.2	S-1	*
Tubing Hardness	3.1.1.1	4.5.3	S-2	*
Leakage	3.2	4.5.7	S-2	*
Intravenous Needle (security of attachment)	3.2.2	4.5.4	S-2	*
Female Luer fitting (gaging test)	3.2.4	4.5.5	S-2	*
Freedom from pyrogens	3.4	4.5.8	S-1	*
Sterility	3.4	4.5.9	See 4.5.9	*
Safety	3.4	4.5.10	S-1	*
Leakage and separation of set	3.7	4.5.6	S-3	*
Fluid tightness	3.2.4	4.5.7	S-2	*

* Acceptance number shall be zero.

** Each lot of product need not be tested provided evidence is furnished that samples of the formulation and identical method of processing have previously been tested, that, they comply with USP requirements for Class IV plastics and heavy metals and that they are the same as being offered on this contract.

4.3.3. Sample unit usage. Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.4 Examination. The sets shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted in accordance with MIL-STD-105 and the classification of defects specified in Table III. Examination, however, shall not be restricted to the classified possible defects listed in Table III.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects

Categories	Defects
Critical	None defined
Major	
101	End protectors or other components missing or dislodged.
102	End protectors not properly designed to permit flow of sterilizing gases, if used.
103	Not free of bloom or tackiness.
104	End protectors not properly designed to maintain sterility.
105	Style and design not as specified.
106	Presence of foreign matter.
107	Set not free of holes or cuts.
108	Dimensions not as specified.
109	Fluid path not clean.
Minor	
201	Packaging, packing and marking not as specified.
202	Instructions missing, incomplete or illegible.
203	Contractor's name or registered trademark, missing, incomplete, or illegible.

4.5 Tests.

4.5.1 Heavy metals extractable test. A 2-gram composite sample shall be obtained from the rubber and plastic components. The material shall be cut into small slivers, so that no thickness exceeds 3 mm, and eluted for 72 hours at 50°C, 24 hours at 70°C, 60 minutes at 121°C in one menstruum of 0.9% sodium chloride. Each gram of the material shall be covered with 5 milliliters (ml) of the menstruum. The menstrua are collected, agitated after elution and then submitted to spectographic absorption, spectrophotometric or specific wet analysis for lead, barium, and cadmium. The metals are reported in parts per million. Thus 1.0 ppm in elute is equivalent to extracting 1.0 mg from 200 g of test material dispersed in 1,000 ml of elution menstruum.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

4.5.1.1 Alternate heavy metals extractable test. Accurately weigh a 5-gram sample for each of the metals (barium, cadmium or lead) into a 1-oz. Vycor crucible. Reduce to ash in a muffle furnace at 650°C for eight hours. Let cool to room temperature and add 15 ml of 1:1 hydrochloric acid. Bring to a boil on a hot plate, allow to boil to 2 to 3 minutes, remove, let cool, and filter through Whatman #2 filter paper into a 50-ml volumetric flask. Rinse the residue on the filter with deionised water and dilute to 50 ml. Prepare standard solutions of 1, 2 and 5 ppm for each of the metals, namely barium, cadmium, and lead. Determine the absorbances of these solutions at appropriate wave lengths using a Hollow Cathode Ba, Cd, or Pb lamp for each of the respective metals and 10X scale expansion on a spectrophotometer, such as the Perkin-Elmer Model 303, Atomic Absorption Spectrophotometer. Determine the absorbance of the sample solutions using the same conditions as with the standards.

Calculation:

$$\frac{A \times F \times \text{dilution}}{W} = \text{ppm barium, cadium, or lead}$$

A = Absorbance
F = Factor = $\left(\frac{\text{Concentration standard in ppm Ba, Cd, or PB}}{\text{absorbance}}\right)$*
W = Sample weight in grams

*The three standards are averaged to give ppm/unit A.

4.5.2 Biological tests. The plastics shall be tested in accordance with the USP for Class IV plastics. Sampling shall be as shown in Table II.

4.5.3 Tubing hardness. Plastic tubing hardness shall be determined in accordance with ASTM D2240 Standard Test Method for Rubber Property - Durometer Hardness. The test shall be conducted at a room temperature of 21 to 24°C on samples which had been conditioned at this temperature for not less than 3 hours prior to testing. The test sample shall be laminated to a thickness of not less than 1/4 inch. The gage reading shall be made 15 seconds after application of the instrument.

4.5.4 Security of assembly test. The tubing of the set shall be severed at the midpoint of the distance between the needle and female Luer connector. The security of assembly of the needle and the female Luer connector shall be tested by gradually applying a one pound weight to the severed end of the tubing while the cannula of the needle and the female connector respectively are held in a clamping device so that the tubing hangs vertically downward. This force shall be maintained for 10 minutes. The union of the component and tubing shall not separate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

4.5.5 Gaging test. The female Luer cavity shall be gaged with a plug gage conforming to Figure 5 of ANSI/HIMA MD70.1 with the feeler gage removed. The fit of the female cavity shall be such as not to bottom out on the gage but shall not exceed the maximum of the plug gage in the female Luer cavity.

4.5.6 Leakage and separation test. The female Luer cavity shall be tested for leakage and separation with 8 psig internal air pressure by slipping the female Luer cavity of the completed assembly onto a plug gage of the configuration specified in Figure 6 of ANSI/HIMA MD70.1. An axial force of 5 Newtons (1.1 lbs) shall be used to slip the set onto the gage. The needle tip shall be plugged and the assembly submerged in water for 10 seconds and observed for leakage or separation.

4.5.7 Fluid tightness. With the needle in horizontal position, the hub shall be attached by means of a plug gage, as shown in Figure 6 of the ANSI/HIMA MD70.1 to a water pressure source having a pressure gage. The water shall be run through the needle so as to fill any hub cavity space and the lumen of the cannula; the lumen shall then be sealed by a piece of solid rubber slipped onto the point, and a pressure of 45 pounds per square inch applied for 10 seconds.

4.5.8 Freedom from pyrogens. Samples of the finished sets shall be tested for freedom from pyrogens, in accordance with the USP. Sampling shall be as shown in Table II.

4.5.9 Sterility test. The intravenous injection sets shall be tested for sterility in accordance with the method specified in the USP using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates it from product intended for distribution. The incubation medium for the paper spore strip (carrier) shall be soybean-casein digest. The incubation media for the inoculated product, or inoculated simulated product shall be fluid thioglycollate or alternative thioglycollate at 30° to 35°C for one half of the specimens and soybean-casin digest at 20° to 25°C for the remaining half of the specimens. The minimum incubation time shall be ten days. Sampling shall be as follows:

Medium	Paper Spore Strip	Inoc. Prod. or Inoc. Simulated Prod.
Soybean-casein digest	S-3*	S-3*
Fluid thioglycollate or Alternative thioglycollate		S-3*

* Use a minimum of 20 specimens. Acceptance number is zero.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

4.5.9.1 Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism
Steam	B. stearothermophilus
Dry heat	B. stearothermophilus
	B. subtilis
Ethylene oxide	B. subtilis
Gamma radiation	B. pumilus

The number of viable spores shall be sufficient to provide assurance that the product has been sterilized.

4.5.9.2 Inoculated carrier. Inoculated carriers (paper spore strips) if used, shall be in accordance with MIL-S-36586, Sterilization Test Strip Set, Bacterial Spore. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.5.9 Sterility test and 4.5.9.1 Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in 4.5.9.1. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP and paragraphs 4.5.9 and 4.5.9.1.

4.5.9.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer and approval obtained before employment of the alternate method on a government contract.

4.5.9.4 Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with USP. The inspection level shall be as stated in paragraph 4.5.9 above. Acceptance number shall be zero.

4.5.10 Safety test. The finished sets shall be tested for safety in accordance with the USP. Sampling shall be as shown in Table II.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C as specified (see 6.2).

5.1.1 Level A.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

5.1.1.1 Individual package. Each intravenous injection set shall have the tubing coiled in such a manner to prevent kinking. Each set shall be further packaged sterile in a sealed peel-open type container capable of maintaining sterility of contents unless the package is damaged or opened.

5.1.1.2 Unit package. Forty individually packaged intravenous injection sets shall be packaged in a snug fitting box conforming to PPP-B-566, PPP-B-636, class domestic, or PPP-B-676. Closure shall be as specified in the appendix of the applicable box specification.

5.1.2 Level C. The level C preservation for intravenous injection sets shall conform to MIL-STD-794 requirements for this level.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Three unit packages (120 IV sets) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure and strappings shall be as specified in the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof liner conforming to MIL-L-10547. Closure shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. Three unit packages (120 IV sets) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

5.2.3 Level C. Three unit packages (120 IV sets) shall be packed in accordance with the requirements of MIL-STD-794 for level C.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in a exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination equal 40 cubic feet or more. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 40 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 43 inches in height, 48 inches in width, and 40 inches in length, and shall conform to Type IV or Type V.

Source: https://assist.dla.mil -- Downloaded: 202610-21T20:14Z
Check the source to verify that this is the current version before use.

5.2.5.1 Levels A and B. Intravenous injection sets packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets as specified in 5.2.5.

5.2.5.2 Level C. Intravenous injection sets packed as specified in 5.2.3 shall be unitized as specified in MIL-STD-794.

5.3 Marking.

5.3.1 Levels A, B and C.

5.3.1.1 Individual package. Each individual package shall be marked in accordance with commercial pratice and in addition, the marking shall include the lot number and legend:

"STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED."

5.3.1.2 Unit package, exterior container and unitized load. Each unit package, eexterior container and unitized load shall be marked as specified in MIL-STD-129. Marking shall be also be include the lot number.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.4.

6. NOTES

6.1 Intended use. The intravenous injection set is intended for use in administration of fluid or blood into the scalp vein of a premature infant, infant, young child or into any small vein.

6.2 Ordering data.

6.2.1 Acquisition requirements. Acquisition documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(c) Applicable NSN.

Source: https://assist.dla.mil -- Downloaded: 202610-21T20:14Z
Check the source to verify that this is the currentversion before use.

6.3 National Stock Number (NSN) coverage. This specification covers the following National Stock Number:

National Stock No.	Item Identification
6515-01-107-1200	INTRAVENOUS INJECTION SET, 25 Gage Needle, Pediatric, Disposable, 40s

6.4 Subject term (key word) listing.

Injection set
Intravenous injection set
Pediatric intravenous injection set

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-3878

Source: https://assist.dla.mil -- Downloaded: 202610-21T20:14Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4086 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T20:14Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
MIL-I-42021 (DM)	INTRAVENOUS INJECTION SET, PEDIATRIC, DISPOSABLE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 202610-21T20:14Z
Check the source to verify that this is the current version before use.